



Kit Quality Control Report

RPH CENTRAL PHARMA

CNPJ: 19.315.658/0001-10

R: Maestro Cardim, 769 - São Paulo/SP

Farm. RT: Priscila Larcher CRF/SP-74.417

Date: From 25/06/2021 To 25/06/2021

| Product: | Manuf: | Calibration | D/T | QC By | Solution | Media | Origin | Middle | Top | Background | Particle Size/ OK? | % Tag | Pass Fail |
|----------------------|--------|-------------|-------|-------|-----------------------------|-------------------|--------|--------|-------|----------------------|--------------------|----------------------|-----------|
| DMSA-99mTc | | 25/06/2021 | 07:34 | PP | Butanona | 3MM Whatman | 431 | | 2 | 0 | | 96,005 % | Pass |
| Lot # K-20210625-025 | | | | | Normal Saline | Silica gel | 13 | | 355 | pH (Low/Actual/High) | | 2,00 <= 3,00 <= 3,00 | |
| | | | | | | Límpido/Incolor | | | | | | | |
| DMSA-99mTc | | 25/06/2021 | 08:02 | PP | Butanona | 3MM Whatman | 456 | | 2 | 0 | | 96,065 % | Pass |
| Lot # K-20210625-030 | | | | | Normal Saline | Silica gel | 12 | | 331 | pH (Low/Actual/High) | | 2,00 <= 3,00 <= 3,00 | |
| | | | | | | Límpido/Incolor | | | | | | | |
| DMSA-99mTc | | 25/06/2021 | 08:03 | PP | Butanona | 3MM Whatman | 399 | | 2 | 0 | | 96,29 % | Pass |
| Lot # K-20210625-031 | | | | | Normal Saline | Silica gel | 7 | | 211 | pH (Low/Actual/High) | | 2,00 <= 3,00 <= 3,00 | |
| | | | | | | Límpido/Incolor | | | | | | | |
| DMSA-99mTc | | 25/06/2021 | 10:49 | WG | Butanona | 3MM Whatman | 189,7 | | 0,6 | 0 | | 94,537 % | Pass |
| Lot # K-20210625-043 | | | | | Normal Saline | Silica gel | 7,3 | | 134,5 | pH (Low/Actual/High) | | 2,00 <= 3,00 <= 3,00 | |
| | | | | | | Límpido/Incolor | | | | | | | |
| DOTATATE-68Ga | | 25/06/2021 | 12:35 | JS | Citrato de Sódio 0,1M | Silica gel | 106 | | 9 | 0 | | 79,13 % | Fail |
| Lot # K-20210625-050 | | | | | Metanol/Acetato de Amonio 1 | Silica gel | 12 | | 80 | pH (Low/Actual/High) | | 4,00 <= 4,50 <= 8,00 | |
| | | | | | | Límpido e Incolor | | | | | | | |
| DOTATOC-68Ga | | 25/06/2021 | 13:37 | JS | Citrato de Sódio 0,1M | Silica gel | 49,2 | | 0,3 | 0 | | 99,394 % | Pass |
| Lot # K-20210625-051 | | | | | | | | | | pH (Low/Actual/High) | | 4,00 <= 5,00 <= 8,00 | |
| | | | | | | Límpido e Incolor | | | | | | | |
| DTPA-99mTc | RPH | 25/06/2021 | 04:21 | PP | Acetona | 3MM Whatman | 387 | | 1 | 0 | | 95,325 % | Pass |
| Lot # K-20210625-012 | | | | | Normal Saline | 3MM Whatman | 22 | | 476 | pH (Low/Actual/High) | | 3,80 <= 6,00 <= 7,50 | |
| | | | | | | Límpido/Incolor | | | | | | | |
| DTPA-99mTc | RPH | 25/06/2021 | 04:42 | PP | Acetona | 3MM Whatman | 632 | | 1 | 0 | | 97,585 % | Pass |
| Lot # K-20210625-017 | | | | | Normal Saline | 3MM Whatman | 10 | | 433 | pH (Low/Actual/High) | | 3,80 <= 6,00 <= 7,50 | |
| | | | | | | Límpido/Incolor | | | | | | | |
| DTPA-99mTc | RPH | 25/06/2021 | 07:14 | PP | Acetona | 3MM Whatman | 565 | | 5 | 0 | | 96,054 % | Pass |
| Lot # K-20210625-023 | | | | | Normal Saline | 3MM Whatman | 12 | | 379 | pH (Low/Actual/High) | | 3,80 <= 6,00 <= 7,50 | |
| | | | | | | Límpido/Incolor | | | | | | | |
| DTPA-99mTc | RPH | 25/06/2021 | 10:50 | WG | Acetona | 3MM Whatman | 180,7 | | 0,3 | 0 | | 98,302 % | Pass |
| Lot # K-20210625-044 | | | | | Normal Saline | 3MM Whatman | 2,8 | | 179,9 | pH (Low/Actual/High) | | 3,80 <= 6,00 <= 7,50 | |
| | | | | | | Límpido/Incolor | | | | | | | |
| FITATO-99mTc | | 25/06/2021 | 03:20 | PP | Butanona | 3MM Whatman | 251 | | 1 | 0 | | 99,603 % | Pass |
| Lot # K-20210625-003 | | | | | | | | | | pH (Low/Actual/High) | | 5,00 <= 6,00 <= 7,00 | |
| | | | | | | Límpido/Incolor | | | | | | | |

| Product: | Manuf: | Calibration D/T | Solution | Media | Origin | Middle | Top | Solution 2 | Media 2 | Origin 2 | Top 2 | % Tag | QC By | Pass Fail |
|--------------------------------------|--------|------------------|----------|----------|--------|-------------|-----------------|------------|---------|----------|-------|----------------------|--------------|-----------|
| FITATO-99mTc Lot # K-20210625-026 | | 25/06/2021 07:41 | PP | Butanona | | 3MM Whatman | | 393 | | 2 | 0 | 99,494 % | Pass | |
| | | | | | | | Límpido/Incolor | | | | | pH (Low/Actual/High) | 5,00 <= 6,00 | <= 7,00 |
| MAA-99mTc Lot # K-20210625-008 | | 25/06/2021 04:06 | PP | Butanona | | 3MM Whatman | | 108 | | 2 | 0 | 98,182 % | Pass | |
| | | | | | | | Turvo/Incolor | | | | | pH (Low/Actual/High) | 3,00 <= 6,00 | <= 6,00 |
| MAA-99mTc Lot # K-20210625-009 | | 25/06/2021 04:08 | PP | Butanona | | 3MM Whatman | | 314 | | 3 | 0 | 99,054 % | Pass | |
| | | | | | | | Turvo/Incolor | | | | | pH (Low/Actual/High) | 3,00 <= 6,00 | <= 6,00 |
| MAA-99mTc Lot # K-20210625-010 | | 25/06/2021 04:09 | PP | Butanona | | 3MM Whatman | | 130 | | 1 | 0 | 99,237 % | Pass | |
| | | | | | | | Turvo/Incolor | | | | | pH (Low/Actual/High) | 3,00 <= 6,00 | <= 6,00 |
| MAA-99mTc Lot # K-20210625-011 | | 25/06/2021 04:11 | PP | Butanona | | 3MM Whatman | | 304 | | 1 | 0 | 99,672 % | Pass | |
| | | | | | | | Turvo/Incolor | | | | | pH (Low/Actual/High) | 3,00 <= 6,00 | <= 6,00 |
| MAA-99mTc Lot # K-20210625-015 | | 25/06/2021 04:38 | PP | Butanona | | 3MM Whatman | | 230 | | 2 | 0 | 99,138 % | Pass | |
| | | | | | | | Turvo/Incolor | | | | | pH (Low/Actual/High) | 3,00 <= 6,00 | <= 6,00 |
| MAA-99mTc Lot # K-20210625-016 | | 25/06/2021 04:40 | PP | Butanona | | 3MM Whatman | | 230 | | 1 | 0 | 99,567 % | Pass | |
| | | | | | | | Turvo/Incolor | | | | | pH (Low/Actual/High) | 3,00 <= 6,00 | <= 6,00 |
| MAA-99mTc Lot # K-20210625-028 | | 25/06/2021 07:49 | PP | Butanona | | 3MM Whatman | | 314 | | 4 | 0 | 98,742 % | Pass | |
| | | | | | | | Turvo/Incolor | | | | | pH (Low/Actual/High) | 3,00 <= 6,00 | <= 6,00 |
| MAA-99mTc Lot # K-20210625-029 | | 25/06/2021 07:52 | PP | Butanona | | 3MM Whatman | | 238 | | 2 | 0 | 99,167 % | Pass | |
| | | | | | | | Turvo/Incolor | | | | | pH (Low/Actual/High) | 3,00 <= 6,00 | <= 6,00 |
| MAA-99mTc Lot # K-20210625-039 | | 25/06/2021 10:08 | WG | Butanona | | 3MM Whatman | | 103,1 | | 0,4 | 0 | 99,614 % | Pass | |
| | | | | | | | Turvo/Incolor | | | | | pH (Low/Actual/High) | 3,00 <= 6,00 | <= 6,00 |
| MAA-99mTc Lot # K-20210625-040 | | 25/06/2021 10:11 | WG | Butanona | | 3MM Whatman | | 74,4 | | 1,8 | 0 | 97,638 % | Pass | |
| | | | | | | | Turvo/Incolor | | | | | pH (Low/Actual/High) | 3,00 <= 6,00 | <= 6,00 |
| MAA-99mTc Lot # K-20210625-049 | | 25/06/2021 11:34 | WG | Butanona | | 3MM Whatman | | 164,6 | | 1,8 | 0 | 98,918 % | Pass | |
| | | | | | | | Turvo/Incolor | | | | | pH (Low/Actual/High) | 3,00 <= 6,00 | <= 6,00 |
| MAA-99mTc Lot # K-20210625-052 | | 25/06/2021 15:16 | JS | Butanona | | 3MM Whatman | | 112,5 | | 0,6 | 0 | 99,469 % | Pass | |
| | | | | | | | Turvo/Incolor | | | | | pH (Low/Actual/High) | 3,00 <= 6,00 | <= 6,00 |

| Product: | Manuf: | Calibration D/T | Solution | Media | Origin | Middle | Top | Solution 2 | Media 2 | Origin 2 | Top 2 | % Tag | QC By | Pass Fail |
|------------------------------------|------------|-----------------|----------|--------------------------|-------------------|---------|-----|------------|---------|-------------|-------|----------------------|----------------------|-----------|
| MDP-99mTc Lot # K-20210625-018 | 25/06/2021 | 04:45 | PP | Acetona Normal Saline | 3MM | Whatman | | 1313 19 | | 55 1452 | 0 | | 94,688 % | Pass |
| | | | | | | | | | | | | pH (Low/Actual/High) | 4,00 <= 6,00 <= 7,80 | |
| | | | | | | | | | | | | Límpido/Incolor | | |
| MDP-99mTc Lot # K-20210625-032 | 25/06/2021 | 08:10 | PP | Acetona Normal Saline | 3MM | Whatman | | 902 7 | | 3 641 | 0 | | 98,588 % | Pass |
| | | | | | | | | | | | | pH (Low/Actual/High) | 4,00 <= 6,00 <= 7,80 | |
| | | | | | | | | | | | | Límpido/Incolor | | |
| MDP-99mTc Lot # K-20210625-035 | 25/06/2021 | 08:33 | WG | Acetona Normal Saline | 3MM | Whatman | | 375 3,4 | | 12 163,7 | 0 | | 94,865 % | Pass |
| | | | | | | | | | | | | pH (Low/Actual/High) | 4,00 <= 6,00 <= 7,80 | |
| | | | | | | | | | | | | Límpido/Incolor | | |
| MDP-99mTc Lot # K-20210625-036 | 25/06/2021 | 08:44 | BB | Acetona Normal Saline | 3MM | Whatman | | 436 3 | | 1 311 | 0 | | 98,816 % | Pass |
| | | | | | | | | | | | | pH (Low/Actual/High) | 4,00 <= 5,00 <= 7,80 | |
| | | | | | | | | | | | | Límpido/Incolor | | |
| MDP-99mTc Lot # K-20210625-042 | 25/06/2021 | 10:39 | BB | Acetona Normal Saline | 3MM | Whatman | | 381 4 | | 1 373 | 0 | | 98,677 % | Pass |
| | | | | | | | | | | | | pH (Low/Actual/High) | 4,00 <= 6,00 <= 7,80 | |
| | | | | | | | | | | | | Límpido/Incolor | | |
| MDP-99mTc Lot # K-20210625-046 | 25/06/2021 | 11:02 | WG | Acetona Normal Saline | 3MM | Whatman | | 599 1 | | 3 368 | 0 | | 99,231 % | Pass |
| | | | | | | | | | | | | pH (Low/Actual/High) | 4,00 <= 6,00 <= 7,80 | |
| | | | | | | | | | | | | Límpido/Incolor | | |
| MIBI-99mTc Lot # K-20210625-001 | 25/06/2021 | 03:03 | PP | Normal Saline Metanol | Sílica gel 3MM | Whatman | | 421 2 | | 9 895 | 0 | | 97,684 % | Pass |
| | | | | | | | | | | | | pH (Low/Actual/High) | 5,00 <= 6,00 <= 6,00 | |
| | | | | | | | | | | | | Límpido/Incolor | | |
| MIBI-99mTc Lot # K-20210625-002 | 25/06/2021 | 03:04 | PP | Normal Saline Metanol | Sílica gel 3MM | Whatman | | 646 14 | | 12 912 | 0 | | 96,664 % | Pass |
| | | | | | | | | | | | | pH (Low/Actual/High) | 5,00 <= 6,00 <= 6,00 | |
| | | | | | | | | | | | | Límpido/Incolor | | |
| MIBI-99mTc Lot # K-20210625-004 | 25/06/2021 | 03:38 | PP | Normal Saline Metanol | Sílica gel 3MM | Whatman | | 769 41 | | 10 1865 | 0 | | 96,565 % | Pass |
| | | | | | | | | | | | | pH (Low/Actual/High) | 5,00 <= 6,00 <= 6,00 | |
| | | | | | | | | | | | | Límpido/Incolor | | |
| MIBI-99mTc Lot # K-20210625-005 | 25/06/2021 | 03:41 | PP | Normal Saline Metanol | Sílica gel 3MM | Whatman | | 1291 20 | | 19 1547 | 0 | | 97,273 % | Pass |
| | | | | | | | | | | | | pH (Low/Actual/High) | 5,00 <= 6,00 <= 6,00 | |
| | | | | | | | | | | | | Límpido/Incolor | | |
| MIBI-99mTc Lot # K-20210625-006 | 25/06/2021 | 03:58 | PP | Normal Saline Metanol | Sílica gel 3MM | Whatman | | 454 1 | | 10 400 | 0 | | 97,595 % | Pass |
| | | | | | | | | | | | | pH (Low/Actual/High) | 5,00 <= 6,00 <= 6,00 | |
| | | | | | | | | | | | | Límpido/Incolor | | |
| MIBI-99mTc Lot # K-20210625-007 | 25/06/2021 | 04:01 | PP | Normal Saline Metanol | Sílica gel 3MM | Whatman | | 1599 10 | | 48 1653 | 0 | | 96,484 % | Pass |
| | | | | | | | | | | | | pH (Low/Actual/High) | 5,00 <= 6,00 <= 6,00 | |
| | | | | | | | | | | | | Límpido/Incolor | | |
| MIBI-99mTc Lot # K-20210625-013 | 25/06/2021 | 04:35 | PP | Normal Saline Metanol | Sílica gel 3MM | Whatman | | 1271 48 | | 20 1398 | 0 | | 95,131 % | Pass |
| | | | | | | | | | | | | pH (Low/Actual/High) | 5,00 <= 6,00 <= 6,00 | |
| | | | | | | | | | | | | Límpido/Incolor | | |

| Product: | Manuf: | Calibration D/T | Solution | Media | Origin | Middle | Top | Solution 2 | Media 2 | Origin 2 | Top 2 | % Tag | QC By | Pass Fail |
|------------------------------------|------------|-----------------|----------|--------------------------|---------------------------|--------|-----|------------|---------|-----------|-------|----------------------|----------------------|-------------------|
| MIBI-99mTc Lot # K-20210625-019 | 25/06/2021 | 04:53 | PP | Normal Saline Metanol | Sílica gel 3MM Whatman | | | 1167 17 | | 9 869 | 0 | pH (Low/Actual/High) | 97,316 % | Pass |
| | | | | | | | | | | | | | 5,00 <= 6,00 <= 6,00 | |
| | | | | | | | | | | | | | | Límpido/Incolor |
| MIBI-99mTc Lot # K-20210625-021 | 25/06/2021 | 07:09 | PP | Normal Saline Metanol | Sílica gel 3MM Whatman | | | 561 9 | | 14 769 | 0 | pH (Low/Actual/High) | 96,408 % | Pass |
| | | | | | | | | | | | | | 5,00 <= 6,00 <= 6,00 | |
| | | | | | | | | | | | | | | Límpido/Incolor |
| MIBI-99mTc Lot # K-20210625-022 | 25/06/2021 | 07:12 | PP | Normal Saline Metanol | Sílica gel 3MM Whatman | | | 361 1 | | 6 568 | 0 | pH (Low/Actual/High) | 98,189 % | Pass |
| | | | | | | | | | | | | | 5,00 <= 6,00 <= 6,00 | |
| | | | | | | | | | | | | | | Límpido/Incolor |
| MIBI-99mTc Lot # K-20210625-024 | 25/06/2021 | 07:31 | PP | Normal Saline Metanol | Sílica gel 3MM Whatman | | | 657 1 | | 11 988 | 0 | pH (Low/Actual/High) | 98,252 % | Pass |
| | | | | | | | | | | | | | 5,00 <= 6,00 <= 6,00 | |
| | | | | | | | | | | | | | | Límpido/Incolor |
| MIBI-99mTc Lot # K-20210625-027 | 25/06/2021 | 07:45 | PP | Normal Saline Metanol | Sílica gel 3MM Whatman | | | 548 5 | | 12 633 | 0 | pH (Low/Actual/High) | 97,073 % | Pass |
| | | | | | | | | | | | | | 5,00 <= 6,00 <= 6,00 | |
| | | | | | | | | | | | | | | Límpido/Incolor |
| MIBI-99mTc Lot # K-20210625-033 | 25/06/2021 | 08:25 | WG | Normal Saline Metanol | Sílica gel 3MM Whatman | | | 586 5 | | 3 550 | 0 | pH (Low/Actual/High) | 98,59 % | Pass |
| | | | | | | | | | | | | | 5,00 <= 6,00 <= 6,00 | |
| | | | | | | | | | | | | | | Límpido/Incolor |
| MIBI-99mTc Lot # K-20210625-034 | 25/06/2021 | 08:30 | WG | Normal Saline Metanol | Sílica gel 3MM Whatman | | | 545 13 | | 2 860 | 0 | pH (Low/Actual/High) | 98,145 % | Pass |
| | | | | | | | | | | | | | 5,00 <= 6,00 <= 6,00 | |
| | | | | | | | | | | | | | | Límpido/Incolor |
| MIBI-99mTc Lot # K-20210625-038 | 25/06/2021 | 10:00 | WG | Normal Saline Metanol | Sílica gel 3MM Whatman | | | 218 1 | | 1 440 | 0 | pH (Low/Actual/High) | 99,317 % | Pass |
| | | | | | | | | | | | | | 5,00 <= 6,00 <= 6,00 | |
| | | | | | | | | | | | | | | Límpido/Incolor |
| MIBI-99mTc Lot # K-20210625-041 | 25/06/2021 | 10:14 | WG | Normal Saline Metanol | Sílica gel 3MM Whatman | | | 461 11 | | 2 597 | 0 | pH (Low/Actual/High) | 97,759 % | Pass |
| | | | | | | | | | | | | | 5,00 <= 6,00 <= 6,00 | |
| | | | | | | | | | | | | | | Límpido/Incolor |
| MIBI-99mTc Lot # K-20210625-045 | 25/06/2021 | 10:56 | WG | Normal Saline Metanol | Sílica gel 3MM Whatman | | | 283 1 | | 6 399 | 0 | pH (Low/Actual/High) | 97,674 % | Pass |
| | | | | | | | | | | | | | 5,00 <= 6,00 <= 6,00 | |
| | | | | | | | | | | | | | | Límpido/Incolor |
| MIBI-99mTc Lot # K-20210625-047 | 25/06/2021 | 11:19 | WG | Normal Saline Metanol | Sílica gel 3MM Whatman | | | 316 3 | | 1 501 | 0 | pH (Low/Actual/High) | 99,089 % | Pass |
| | | | | | | | | | | | | | 5,00 <= 6,00 <= 6,00 | |
| | | | | | | | | | | | | | | Límpido/Incolor |
| MIBI-99mTc Lot # K-20210625-048 | 25/06/2021 | 11:32 | WG | Normal Saline Metanol | Sílica gel 3MM Whatman | | | 372 8 | | 1 557 | 0 | pH (Low/Actual/High) | 98,316 % | Pass |
| | | | | | | | | | | | | | 5,00 <= 6,00 <= 6,00 | |
| | | | | | | | | | | | | | | Límpido/Incolor |
| PSMA-68Ga Lot # K-20210625-020 | 25/06/2021 | 06:55 | JS | Acetato de amônio | Sílica gel | | | 0,2 | | 29,6 | 0 | pH (Low/Actual/High) | 99,329 % | Pass |
| | | | | | | | | | | | | | 4,00 <= 5,00 <= 8,00 | |
| | | | | | | | | | | | | | | Límpido e Incolor |

| Product: | Manuf: | Calibration D/T | Solution | Media | Origin | Middle | Top | Solution 2 | Media 2 | Origin 2 | Top 2 | % Tag | QC By | Pass Fail |
|--------------------------------------|--------|------------------|----------|-------------------|-------------|--------|-----|------------|---------|----------|-------|----------------------|----------------------|-------------------|
| PSMA-68Ga Lot # K-20210625-037 | | 25/06/2021 09:54 | BB | Acetato de amônio | Sílica gel | | | 0,2 | 42,2 | | 0 | 99,528 | % Pass | |
| | | | | | | | | | | | | pH (Low/Actual/High) | 4,00 <= 5,00 <= 8,00 | |
| | | | | | | | | | | | | | | Límpido e Incolor |
| Tecnécio-99m Lot # E-20210625-001 | | 25/06/2021 02:55 | PP | Butanona | 3MM Whatman | | | 3 | 678 | | 0 | 99,559 | % Pass | |
| | | | | | | | | | | | | pH (Low/Actual/High) | 5,00 <= 6,00 <= 6,00 | |
| | | | | | | | | | | | | | | Límpido e Incolor |
| Tecnécio-99m Lot # E-20210625-002 | | 25/06/2021 03:01 | PP | Butanona | 3MM Whatman | | | 2 | 117 | | 0 | 98,319 | % Pass | |
| | | | | | | | | | | | | pH (Low/Actual/High) | 5,00 <= 6,00 <= 6,00 | |
| | | | | | | | | | | | | | | Límpido e Incolor |
| Tecnécio-99m Lot # E-20210625-003 | | 25/06/2021 03:17 | PP | Butanona | 3MM Whatman | | | 6 | 284 | | 0 | 97,931 | % Pass | |
| | | | | | | | | | | | | pH (Low/Actual/High) | 5,00 <= 6,00 <= 6,00 | |
| | | | | | | | | | | | | | | Límpido e Incolor |
| Tecnécio-99m Lot # E-20210625-004 | | 25/06/2021 03:35 | PP | Butanona | 3MM Whatman | | | 1 | 606 | | 0 | 99,835 | % Pass | |
| | | | | | | | | | | | | pH (Low/Actual/High) | 5,00 <= 6,00 <= 6,00 | |
| | | | | | | | | | | | | | | Límpido e Incolor |
| Tecnécio-99m Lot # E-20210625-005 | | 25/06/2021 03:35 | PP | Butanona | 3MM Whatman | | | 1 | 234 | | 0 | 99,574 | % Pass | |
| | | | | | | | | | | | | pH (Low/Actual/High) | 5,00 <= 6,00 <= 6,00 | |
| | | | | | | | | | | | | | | Límpido e Incolor |
| Tecnécio-99m Lot # E-20210625-006 | | 25/06/2021 03:54 | PP | Butanona | 3MM Whatman | | | 2 | 261 | | 0 | 99,24 | % Pass | |
| | | | | | | | | | | | | pH (Low/Actual/High) | 5,00 <= 6,00 <= 6,00 | |
| | | | | | | | | | | | | | | Límpido e Incolor |
| Tecnécio-99m Lot # E-20210625-007 | | 25/06/2021 03:54 | PP | Butanona | 3MM Whatman | | | 1 | 143 | | 0 | 99,306 | % Pass | |
| | | | | | | | | | | | | pH (Low/Actual/High) | 5,00 <= 6,00 <= 6,00 | |
| | | | | | | | | | | | | | | Límpido e Incolor |
| Tecnécio-99m Lot # E-20210625-008 | | 25/06/2021 04:04 | PP | Butanona | 3MM Whatman | | | 2 | 103 | | 0 | 98,095 | % Pass | |
| | | | | | | | | | | | | pH (Low/Actual/High) | 5,00 <= 6,00 <= 6,00 | |
| | | | | | | | | | | | | | | Límpido e Incolor |
| Tecnécio-99m Lot # E-20210625-009 | | 25/06/2021 04:04 | PP | Butanona | 3MM Whatman | | | 4 | 134 | | 0 | 97,101 | % Pass | |
| | | | | | | | | | | | | pH (Low/Actual/High) | 5,00 <= 6,00 <= 6,00 | |
| | | | | | | | | | | | | | | Límpido e Incolor |
| Tecnécio-99m Lot # E-20210625-010 | | 25/06/2021 04:32 | PP | Butanona | 3MM Whatman | | | 5 | 541 | | 0 | 99,084 | % Pass | |
| | | | | | | | | | | | | pH (Low/Actual/High) | 5,00 <= 6,00 <= 6,00 | |
| | | | | | | | | | | | | | | Límpido e Incolor |
| Tecnécio-99m Lot # E-20210625-011 | | 25/06/2021 04:36 | PP | Butanona | 3MM Whatman | | | 9 | 641 | | 0 | 98,615 | % Pass | |
| | | | | | | | | | | | | pH (Low/Actual/High) | 5,00 <= 6,00 <= 6,00 | |
| | | | | | | | | | | | | | | Límpido e Incolor |
| Tecnécio-99m Lot # E-20210625-012 | | 25/06/2021 04:43 | PP | Butanona | 3MM Whatman | | | 13 | 927 | | 0 | 98,617 | % Pass | |
| | | | | | | | | | | | | pH (Low/Actual/High) | 5,00 <= 6,00 <= 6,00 | |
| | | | | | | | | | | | | | | Límpido e Incolor |

| Product: | Manuf: | Calibration | D/T | Solution | Media | Origin | Middle | Top | Solution 2 | Media 2 | Origin 2 | Top 2 | % Tag | QC By | Pass Fail |
|--------------------------------------|------------|-------------|-----|----------|-------|--------|---------|-------------------|------------|---------|----------|----------------------|--------------|---------|-----------|
| Tecnécio-99m Lot # E-20210625-014 | 25/06/2021 | 07:02 | PP | Butanona | | 3MM | Whatman | | 2 | | 451 | 0 | 99,558 % | Pass | |
| | | | | | | | | Límpido e Incolor | | | | pH (Low/Actual/High) | 5,00 <= 6,00 | <= 6,00 | |
| Tecnécio-99m Lot # E-20210625-015 | 25/06/2021 | 07:07 | PP | Butanona | | 3MM | Whatman | | 2 | | 264 | 0 | 99,248 % | Pass | |
| | | | | | | | | Límpido e Incolor | | | | pH (Low/Actual/High) | 5,00 <= 6,00 | <= 6,00 | |
| Tecnécio-99m Lot # E-20210625-016 | 25/06/2021 | 07:10 | PP | Butanona | | 3MM | Whatman | | 1 | | 241 | 0 | 99,587 % | Pass | |
| | | | | | | | | Límpido e Incolor | | | | pH (Low/Actual/High) | 5,00 <= 6,00 | <= 6,00 | |
| Tecnécio-99m Lot # E-20210625-017 | 25/06/2021 | 07:25 | PP | Butanona | | 3MM | Whatman | | 2 | | 358 | 0 | 99,444 % | Pass | |
| | | | | | | | | Límpido e Incolor | | | | pH (Low/Actual/High) | 5,00 <= 6,00 | <= 6,00 | |
| Tecnécio-99m Lot # E-20210625-019 | 25/06/2021 | 07:26 | PP | Butanona | | 3MM | Whatman | | 2 | | 512 | 0 | 99,611 % | Pass | |
| | | | | | | | | Límpido e Incolor | | | | pH (Low/Actual/High) | 5,00 <= 6,00 | <= 6,00 | |
| Tecnécio-99m Lot # E-20210625-020 | 25/06/2021 | 07:39 | PP | Butanona | | 3MM | Whatman | | 1 | | 781 | 0 | 99,872 % | Pass | |
| | | | | | | | | Límpido e Incolor | | | | pH (Low/Actual/High) | 5,00 <= 6,00 | <= 6,00 | |
| Tecnécio-99m Lot # E-20210625-021 | 25/06/2021 | 07:42 | PP | Butanona | | 3MM | Whatman | | 2 | | 465 | 0 | 99,572 % | Pass | |
| | | | | | | | | Límpido e Incolor | | | | pH (Low/Actual/High) | 5,00 <= 6,00 | <= 6,00 | |
| Tecnécio-99m Lot # E-20210625-022 | 25/06/2021 | 07:47 | PP | Butanona | | 3MM | Whatman | | 1 | | 158 | 0 | 99,371 % | Pass | |
| | | | | | | | | Límpido e Incolor | | | | pH (Low/Actual/High) | 5,00 <= 6,00 | <= 6,00 | |
| Tecnécio-99m Lot # E-20210625-023 | 25/06/2021 | 08:04 | PP | Butanona | | 3MM | Whatman | | 5 | | 273 | 0 | 98,201 % | Pass | |
| | | | | | | | | Límpido e Incolor | | | | pH (Low/Actual/High) | 5,00 <= 6,00 | <= 6,00 | |
| Tecnécio-99m Lot # E-20210625-024 | 25/06/2021 | 08:21 | PP | Butanona | | 3MM | Whatman | | 2 | | 493 | 0 | 99,596 % | Pass | |
| | | | | | | | | Límpido e Incolor | | | | pH (Low/Actual/High) | 5,00 <= 6,00 | <= 6,00 | |
| Tecnécio-99m Lot # E-20210625-025 | 25/06/2021 | 08:21 | PP | Butanona | | 3MM | Whatman | | 1 | | 342 | 0 | 99,708 % | Pass | |
| | | | | | | | | Límpido e Incolor | | | | pH (Low/Actual/High) | 5,00 <= 6,00 | <= 6,00 | |
| Tecnécio-99m Lot # E-20210625-026 | 25/06/2021 | 08:31 | WG | Butanona | | 3MM | Whatman | | 1 | | 1115,9 | 0 | 99,91 % | Pass | |
| | | | | | | | | Límpido e Incolor | | | | pH (Low/Actual/High) | 5,00 <= 6,00 | <= 6,00 | |
| Tecnécio-99m Lot # E-20210625-027 | 25/06/2021 | 08:42 | WG | Butanona | | 3MM | Whatman | | 1,1 | | 126,2 | 0 | 99,136 % | Pass | |
| | | | | | | | | Límpido e Incolor | | | | pH (Low/Actual/High) | 5,00 <= 6,00 | <= 6,00 | |

| Product: | Manuf: | Calibration D/T | Solution | Media | Origin | Middle | Top | Solution 2 | Media 2 | Origin 2 | Top 2 | % Tag | QC By | Pass Fail |
|--------------------------------------|------------|-----------------|----------|----------|--------|-------------|-------------------|------------|---------|----------|-------|----------------------|--------------|-----------|
| Tecnício-99m Lot # E-20210625-028 | 25/06/2021 | 09:52 | WG | Butanona | | 3MM Whatman | | 0,1 | | 85,7 | 0 | | 99,883 % | Pass |
| | | | | | | | Límpido e Incolor | | | | | pH (Low/Actual/High) | 5,00 <= 6,00 | <= 6,00 |
| Tecnício-99m Lot # E-20210625-030 | 25/06/2021 | 10:06 | WG | Butanona | | 3MM Whatman | | 10 | | 627 | 0 | | 98,43 % | Pass |
| | | | | | | | Límpido e Incolor | | | | | pH (Low/Actual/High) | 5,00 <= 6,00 | <= 6,00 |
| Tecnício-99m Lot # E-20210625-031 | 25/06/2021 | 10:36 | WG | Butanona | | 3MM Whatman | | 0,8 | | 129,8 | 0 | | 99,387 % | Pass |
| | | | | | | | Límpido e Incolor | | | | | pH (Low/Actual/High) | 5,00 <= 6,00 | <= 6,00 |
| Tecnício-99m Lot # E-20210625-032 | 25/06/2021 | 10:42 | WG | Butanona | | 3MM Whatman | | 0,4 | | 180,6 | 0 | | 99,779 % | Pass |
| | | | | | | | Límpido e Incolor | | | | | pH (Low/Actual/High) | 5,00 <= 6,00 | <= 6,00 |
| Tecnício-99m Lot # E-20210625-033 | 25/06/2021 | 10:47 | WG | Butanona | | 3MM Whatman | | 1,3 | | 163,3 | 0 | | 99,21 % | Pass |
| | | | | | | | Límpido e Incolor | | | | | pH (Low/Actual/High) | 5,00 <= 6,00 | <= 6,00 |
| Tecnício-99m Lot # E-20210625-034 | 25/06/2021 | 10:54 | JS | Butanona | | 3MM Whatman | | 5 | | 166 | 0 | | 97,076 % | Pass |
| | | | | | | | Límpido e Incolor | | | | | pH (Low/Actual/High) | 5,00 <= 6,00 | <= 6,00 |
| Tecnício-99m Lot # E-20210625-035 | 25/06/2021 | 11:00 | WG | Butanona | | 3MM Whatman | | 0,3 | | 105,5 | 0 | | 99,716 % | Pass |
| | | | | | | | Límpido e Incolor | | | | | pH (Low/Actual/High) | 5,00 <= 6,00 | <= 6,00 |
| Tecnício-99m Lot # E-20210625-036 | 25/06/2021 | 11:16 | WG | Butanona | | 3MM Whatman | | 0,3 | | 56,3 | 0 | | 99,47 % | Pass |
| | | | | | | | Límpido e Incolor | | | | | pH (Low/Actual/High) | 5,00 <= 6,00 | <= 6,00 |
| Tecnício-99m Lot # E-20210625-037 | 25/06/2021 | 11:27 | WG | Butanona | | 3MM Whatman | | 3,8 | | 172,2 | 0 | | 97,841 % | Pass |
| | | | | | | | Límpido e Incolor | | | | | pH (Low/Actual/High) | 5,00 <= 6,00 | <= 6,00 |
| Tecnício-99m Lot # E-20210625-040 | 25/06/2021 | 15:12 | JS | Butanona | | 3MM Whatman | | 21 | | 837 | 0 | | 97,552 % | Pass |
| | | | | | | | Límpido e Incolor | | | | | pH (Low/Actual/High) | 5,00 <= 6,00 | <= 6,00 |