



Kit Quality Control Report

RPH CENTRAL PHARMA

CNPJ: 19.315.658/0001-10

R: Maestro Cardim, 769 - São Paulo/SP

Farm. RT: Priscila Larcher CRF/SP-74.417

Date: From 01/07/2021 To 01/07/2021

Product:	Manuf:	Calibration	D/T	QC By	Solution	Media	Origin	Middle	Top	Background	Particle Size/ OK?	% Tag	Pass Fail
DMSA-99mTc Lot # K-20210701-013		01/07/2021	04:04	PP	Butanona Normal Saline	3MM Whatman Sílica gel	197 9		2 101	0	pH (Low/Actual/High)	90,813 %	Pass
Límpido/Incolor													
DMSA-99mTc Lot # K-20210701-019		01/07/2021	07:21	WG	Butanona Normal Saline	3MM Whatman Sílica gel	291 3,2		1 194,7	0	pH (Low/Actual/High)	98,041 %	Pass
Límpido/Incolor													
DMSA-99mTc Lot # K-20210701-033		01/07/2021	10:53	WG	Butanona Normal Saline	3MM Whatman Sílica gel	87,6 1		2,9 99,9	0	pH (Low/Actual/High)	95,804 %	Pass
Límpido/Incolor													
DOTATOC-68Ga Lot # K-20210701-015		01/07/2021	06:56	GF	Citrato de Sódio 0,1M	Sílica gel	26,7		0,5	0	pH (Low/Actual/High)	98,162 %	Pass
Límpido e Incolor													
DOTATOC-68Ga Lot # K-20210701-039		01/07/2021	12:46	JS	Citrato de Sódio 0,1M	Sílica gel	23,6		0,4	0	pH (Low/Actual/High)	98,333 %	Pass
Límpido e Incolor													
DTPA-99mTc Lot # K-20210701-004	RPH	01/07/2021	03:01	PP	Acetona Normal Saline	3MM Whatman 3MM Whatman	662 20		2 1026	0	pH (Low/Actual/High)	97,787 %	Pass
Límpido/Incolor													
DTPA-99mTc Lot # K-20210701-017	RPH	01/07/2021	07:10	WG	Acetona Normal Saline	3MM Whatman 3MM Whatman	185,3 6,7		1 154,2	0	pH (Low/Actual/High)	95,299 %	Pass
Límpido/Incolor													
DTPA-99mTc Lot # K-20210701-026	RPH	01/07/2021	10:17	JS	Acetona Normal Saline	3MM Whatman 3MM Whatman	295 2		3 173	0	pH (Low/Actual/High)	97,85 %	Pass
Límpido/Incolor													
FITATO-99mTc Lot # K-20210701-005		01/07/2021	03:04	PP	Butanona	3MM Whatman	655		5	0	pH (Low/Actual/High)	99,242 %	Pass
Límpido/Incolor													
FITATO-99mTc Lot # K-20210701-024		01/07/2021	10:14	JS	Butanona	3MM Whatman	414		2	0	pH (Low/Actual/High)	99,519 %	Pass
Límpido/Incolor													
MAA-99mTc Lot # K-20210701-009		01/07/2021	03:51	PP	Butanona	3MM Whatman	307		1	0	pH (Low/Actual/High)	99,675 %	Pass
Turvo/Incolor													

Product:	Manuf:	Calibration D/T	Solution	Media	Origin	Middle	Top	Solution 2	Media 2	Origin 2	Top 2	% Tag	QC By	Pass Fail
MAA-99mTc Lot # K-20210701-010		01/07/2021 03:53	PP	Butanona		3MM Whatman		288		2	0	99,31 %	Pass	
												pH (Low/Actual/High)	3,00 <= 6,00	<= 6,00
												Turvo/Incolor		
MAA-99mTc Lot # K-20210701-011		01/07/2021 03:55	PP	Butanona		3MM Whatman		333		1	0	99,701 %	Pass	
												pH (Low/Actual/High)	3,00 <= 6,00	<= 6,00
												Turvo/Incolor		
MAA-99mTc Lot # K-20210701-034		01/07/2021 10:55	JS	Butanona		3MM Whatman		141		3	0	97,917 %	Pass	
												pH (Low/Actual/High)	3,00 <= 5,00	<= 6,00
												Turvo/Incolor		
MAA-99mTc Lot # K-20210701-035		01/07/2021 10:57	JS	Butanona		3MM Whatman		159		1	0	99,375 %	Pass	
												pH (Low/Actual/High)	3,00 <= 6,00	<= 6,00
												Turvo/Incolor		
MAA-99mTc Lot # K-20210701-036		01/07/2021 11:00	JS	Butanona		3MM Whatman		285		2	0	99,303 %	Pass	
												pH (Low/Actual/High)	3,00 <= 6,00	<= 6,00
												Turvo/Incolor		
MAA-99mTc Lot # K-20210701-038		01/07/2021 12:31	JS	Butanona		3MM Whatman		132		0,9	0	99,323 %	Pass	
												pH (Low/Actual/High)	3,00 <= 6,00	<= 6,00
												Turvo/Incolor		
MDP-99mTc Lot # K-20210701-002		01/07/2021 02:56	PP	Acetona		3MM Whatman		1107		1	0	99,821 %	Pass	
				Normal Saline		3MM Whatman		1		1121		pH (Low/Actual/High)	4,00 <= 6,00	<= 7,80
												Límpido/Incolor		
MDP-99mTc Lot # K-20210701-007		01/07/2021 03:37	PP	Acetona		3MM Whatman		1855		22	0	95,561 %	Pass	
				Normal Saline		3MM Whatman		77		2280		pH (Low/Actual/High)	4,00 <= 6,00	<= 7,80
												Límpido/Incolor		
MDP-99mTc Lot # K-20210701-008		01/07/2021 03:40	PP	Acetona		3MM Whatman		1634		21	0	98,417 %	Pass	
				Normal Saline		3MM Whatman		4		1271		pH (Low/Actual/High)	4,00 <= 6,00	<= 7,80
												Límpido/Incolor		
MDP-99mTc Lot # K-20210701-016		01/07/2021 06:58	WG	Acetona		3MM Whatman		475		4	0	98,308 %	Pass	
				Normal Saline		3MM Whatman		6		694		pH (Low/Actual/High)	4,00 <= 6,00	<= 7,80
												Límpido/Incolor		
MDP-99mTc Lot # K-20210701-020		01/07/2021 07:34	WG	Acetona		3MM Whatman		574		7	0	98,354 %	Pass	
				Normal Saline		3MM Whatman		2		451		pH (Low/Actual/High)	4,00 <= 6,00	<= 7,80
												Límpido/Incolor		
MDP-99mTc Lot # K-20210701-021		01/07/2021 08:23	WG	Acetona		3MM Whatman		1632		13	0	96,988 %	Pass	
				Normal Saline		3MM Whatman		29		1276		pH (Low/Actual/High)	4,00 <= 6,00	<= 7,80
												Límpido/Incolor		
MIBI-99mTc Lot # K-20210701-001		01/07/2021 02:45	PP	Normal Saline		Sílica gel		2751		48	0	98,234 %	Pass	
				Metanol		3MM Whatman		1		1941		pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
												Límpido/Incolor		

Product:	Manuf:	Calibration D/T	Solution	Media	Origin	Middle	Top	Solution 2	Media 2	Origin 2	Top 2	% Tag	QC By	Pass Fail
MIBI-99mTc Lot # K-20210701-003		01/07/2021 02:57 PP	Normal Saline Metanol		Silica gel 3MM Whatman			2230 22		43 2028	0	pH (Low/Actual/High)	97,035 %	Pass
Límpido/Incolor														
MIBI-99mTc Lot # K-20210701-006		01/07/2021 03:32 PP	Normal Saline Metanol		Silica gel 3MM Whatman			1158 39		13 2370	0	pH (Low/Actual/High)	97,271 %	Pass
Límpido/Incolor														
MIBI-99mTc Lot # K-20210701-012		01/07/2021 03:59 PP	Normal Saline Metanol		Silica gel 3MM Whatman			831 20		22 1258	0	pH (Low/Actual/High)	95,856 %	Pass
Límpido/Incolor														
MIBI-99mTc Lot # K-20210701-014		01/07/2021 06:41 GF	Normal Saline Metanol		Silica gel 3MM Whatman			1034 2		21 1133	0	pH (Low/Actual/High)	97,833 %	Pass
Límpido/Incolor														
MIBI-99mTc Lot # K-20210701-018		01/07/2021 07:15 WG	Normal Saline Metanol		Silica gel 3MM Whatman			477 1		4 412	0	pH (Low/Actual/High)	98,926 %	Pass
Límpido/Incolor														
MIBI-99mTc Lot # K-20210701-022		01/07/2021 09:41 WG	Normal Saline Metanol		Silica gel 3MM Whatman			806 18		29 937	0	pH (Low/Actual/High)	94,642 %	Pass
Límpido/Incolor														
MIBI-99mTc Lot # K-20210701-025		01/07/2021 10:15 JS	Normal Saline Metanol		Silica gel 3MM Whatman			339 3		2 594	0	pH (Low/Actual/High)	98,911 %	Pass
Límpido/Incolor														
MIBI-99mTc Lot # K-20210701-037		01/07/2021 11:05 WG	Normal Saline Metanol		Silica gel 3MM Whatman			300 2		3 424	0	pH (Low/Actual/High)	98,54 %	Pass
Límpido/Incolor														
PSMA-68Ga Lot # K-20210701-023		01/07/2021 09:49 BB	Acetato de amônio		Silica gel			0,2		14,3	0	pH (Low/Actual/High)	98,621 %	Pass
Límpido e Incolor														
Tecnécio-99m Lot # E-20210701-002		01/07/2021 02:43 PP	Butanona		3MM Whatman			4		468	0	pH (Low/Actual/High)	99,153 %	Pass
Límpido e Incolor														
Tecnécio-99m Lot # E-20210701-001		01/07/2021 02:43 PP	Butanona		3MM Whatman			15		1207	0	pH (Low/Actual/High)	98,773 %	Pass
Límpido e Incolor														
Tecnécio-99m Lot # E-20210701-003		01/07/2021 02:53 PP	Butanona		3MM Whatman			4		282	0	pH (Low/Actual/High)	98,601 %	Pass
Límpido e Incolor														
Tecnécio-99m Lot # E-20210701-004		01/07/2021 02:53 PP	Butanona		3MM Whatman			25		3440	0	pH (Low/Actual/High)	99,278 %	Pass
Límpido e Incolor														

Product:	Manuf:	Calibration D/T	Solution	Media	Origin	Middle	Top	Solution 2	Media 2	Origin 2	Top 2	% Tag	QC By	Pass Fail
Tecnécio-99m Lot # E-20210701-005		01/07/2021 03:27	PP	Butanona		3MM Whatman		1		2254	0	99,956 %	Pass	
							Límpido e Incolor				pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00	
Tecnécio-99m Lot # E-20210701-006		01/07/2021 03:35	PP	Butanona		3MM Whatman		2		674	0	99,704 %	Pass	
							Límpido e Incolor				pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00	
Tecnécio-99m Lot # E-20210701-007		01/07/2021 03:41	PP	Butanona		3MM Whatman		13		1103	0	98,835 %	Pass	
							Límpido e Incolor				pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00	
Tecnécio-99m Lot # E-20210701-008		01/07/2021 06:37	GF	Butanona		3MM Whatman		8		434	0	98,19 %	Pass	
							Límpido e Incolor				pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00	
Tecnécio-99m Lot # E-20210701-009		01/07/2021 06:53	GF	Butanona		3MM Whatman		5		422	0	98,829 %	Pass	
							Límpido e Incolor				pH (Low/Actual/High)	5,00 <= 5,00	<= 6,00	
Tecnécio-99m Lot # E-20210701-010		01/07/2021 06:54	GF	Butanona		3MM Whatman		2		103	0	98,095 %	Pass	
							Límpido e Incolor				pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00	
Tecnécio-99m Lot # E-20210701-012		01/07/2021 07:12	WG	Butanona		3MM Whatman		0,7		107	0	99,35 %	Pass	
							Límpido e Incolor				pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00	
Tecnécio-99m Lot # E-20210701-013		01/07/2021 07:30	WG	Butanona		3MM Whatman		2,7		79,9	0	96,731 %	Pass	
							Límpido e Incolor				pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00	
Tecnécio-99m Lot # E-20210701-014		01/07/2021 08:21	WG	Butanona		3MM Whatman		3		221	0	98,661 %	Pass	
							Límpido e Incolor				pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00	
Tecnécio-99m Lot # E-20210701-015		01/07/2021 09:39	JS	Butanona		3MM Whatman		1		759	0	99,868 %	Pass	
							Límpido e Incolor				pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00	
Tecnécio-99m Lot # E-20210701-017		01/07/2021 10:06	WG	Butanona		3MM Whatman		2		998	0	99,80 %	Pass	
							Límpido e Incolor				pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00	
Tecnécio-99m Lot # E-20210701-018		01/07/2021 10:12	WG	Butanona		3MM Whatman		12		298	0	96,129 %	Pass	
							Límpido e Incolor				pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00	
Tecnécio-99m Lot # E-20210701-019		01/07/2021 10:32	JS	Butanona		3MM Whatman		2		223	0	99,111 %	Pass	
							Límpido e Incolor				pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00	

Product:	Manuf:	Calibration D/T	Solution	Media	Origin	Middle	Top	Solution 2	Media 2	Origin 2	Top 2	% Tag	QC By	Pass Fail
Tecnécio-99m Lot # E-20210701-020	01/07/2021	10:50	JS	Butanona	3MM	Whatman		10		868	0	98,861 %	Pass	
							Límpido e Incolor				pH (Low/Actual/High)	5,00 <= 6,00 <= 6,00		
Tecnécio-99m Lot # E-20210701-021	01/07/2021	11:01	JS	Butanona	3MM	Whatman		2		687	0	99,71 %	Pass	
							Límpido e Incolor				pH (Low/Actual/High)	5,00 <= 6,00 <= 6,00		
Tecnécio-99m Lot # E-20210701-022	01/07/2021	12:30	JS	Butanona	3MM	Whatman		2		318	0	99,375 %	Pass	
							Límpido e Incolor				pH (Low/Actual/High)	5,00 <= 6,00 <= 6,00		
TRODAT-99mTc Lot # K-20210701-027	01/07/2021	10:40	WG	Normal Saline	Sílica gel			99,9		1	0	98,079 %	Pass	
				Butanona	3MM	Whatman		1,6		170,5	pH (Low/Actual/High)	6,50 <= 7,00 <= 8,50		
							Límpido/Incolor							
TRODAT-99mTc Lot # K-20210701-028	01/07/2021	10:42	WG	Normal Saline	Sílica gel			111,1		1	0	97,602 %	Pass	
				Butanona	3MM	Whatman		2		130,8	pH (Low/Actual/High)	6,50 <= 7,00 <= 8,50		
							Límpido/Incolor							
TRODAT-99mTc Lot # K-20210701-029	01/07/2021	10:43	WG	Normal Saline	Sílica gel			111,9		2	0	97,504 %	Pass	
				Butanona	3MM	Whatman		0,9		120,7	pH (Low/Actual/High)	6,50 <= 7,00 <= 8,50		
							Límpido/Incolor							
TRODAT-99mTc Lot # K-20210701-030	01/07/2021	10:45	WG	Normal Saline	Sílica gel			119,7		3,2	0	96,094 %	Pass	
				Butanona	3MM	Whatman		2,2		166,8	pH (Low/Actual/High)	6,50 <= 7,00 <= 8,50		
							Límpido/Incolor							
TRODAT-99mTc Lot # K-20210701-031	01/07/2021	10:46	WG	Normal Saline	Sílica gel			117,4		2,7	0	97,557 %	Pass	
				Butanona	3MM	Whatman		0,3		153,9	pH (Low/Actual/High)	6,50 <= 7,00 <= 8,50		
							Límpido/Incolor							
TRODAT-99mTc Lot # K-20210701-032	01/07/2021	10:46	WG	Normal Saline	Sílica gel			155,2		2,2	0	96,622 %	Pass	
				Butanona	3MM	Whatman		4		198	pH (Low/Actual/High)	6,50 <= 7,00 <= 8,50		
							Límpido/Incolor							