



Kit Quality Control Report

RPH CENTRAL PHARMA

CNPJ: 19.315.658/0001-10

R: Maestro Cardim, 769 - São Paulo/SP

Farm. RT: Priscila Larcher CRF/SP-74.417

Date: From 03/07/2021 To 03/07/2021

Product:	Manuf:	Calibration	D/T	QC By	Solution	Media	Origin	Middle	Top	Background	Particle Size/ OK?	% Tag	Pass Fail
DTPA-99mTc Lot # K-20210703-010	RPH	03/07/2021	03:18	WG	Acetona Normal Saline	3MM Whatman 3MM Whatman	534 4		12 453	0	pH (Low/Actual/High)	96,927 %	Pass
Límpido/Incolor													
FITATO-99mTc Lot # K-20210703-012		03/07/2021	03:24	WG	Butanona	3MM Whatman	462		2	0	pH (Low/Actual/High)	99,569 %	Pass
Límpido/Incolor													
MAA-99mTc Lot # K-20210703-011		03/07/2021	03:22	WG	Butanona	3MM Whatman	71,4		1,1	0	pH (Low/Actual/High)	98,483 %	Pass
Turvo/Incolor													
MDP-99mTc Lot # K-20210703-007		03/07/2021	03:13	WG	Acetona Normal Saline	3MM Whatman 3MM Whatman	906 8		4 1137	0	pH (Low/Actual/High)	98,862 %	Pass
Límpido/Incolor													
MDP-99mTc Lot # K-20210703-008		03/07/2021	03:15	WG	Acetona Normal Saline	3MM Whatman 3MM Whatman	299 5		3 208	0	pH (Low/Actual/High)	96,659 %	Pass
Límpido/Incolor													
MDP-99mTc Lot # K-20210703-009		03/07/2021	03:17	WG	Acetona Normal Saline	3MM Whatman 3MM Whatman	1010 20		9 975	0	pH (Low/Actual/High)	97,107 %	Pass
Límpido/Incolor													
MIBI-99mTc Lot # K-20210703-001		03/07/2021	02:26	WG	Normal Saline Metanol	Sílica gel 3MM Whatman	1036 4		12 2722	0	pH (Low/Actual/High)	98,708 %	Pass
Límpido/Incolor													
MIBI-99mTc Lot # K-20210703-002		03/07/2021	02:30	WG	Normal Saline Metanol	Sílica gel 3MM Whatman	1675 1		8 2116	0	pH (Low/Actual/High)	99,477 %	Pass
Límpido/Incolor													
MIBI-99mTc Lot # K-20210703-003		03/07/2021	02:31	WG	Normal Saline Metanol	Sílica gel 3MM Whatman	1618 24		7 2577	0	pH (Low/Actual/High)	98,647 %	Pass
Límpido/Incolor													
MIBI-99mTc Lot # K-20210703-004		03/07/2021	02:37	WG	Normal Saline Metanol	Sílica gel 3MM Whatman	945 4		6 1789	0	pH (Low/Actual/High)	99,146 %	Pass
Límpido/Incolor													
MIBI-99mTc Lot # K-20210703-005		03/07/2021	02:38	WG	Normal Saline Metanol	Sílica gel 3MM Whatman	1926 17		36 3772	0	pH (Low/Actual/High)	97,716 %	Pass
Límpido/Incolor													

Product:	Manuf:	Calibration D/T	Solution	Media	Origin	Middle	Top	Solution 2	Media 2	Origin 2	Top 2	% Tag	QC By	Pass Fail
MIBI-99mTc Lot # K-20210703-006		03/07/2021 02:40	WG	Normal Saline Metanol	Sílica gel 3MM Whatman			1908 5		46 2812	0		97,468 %	Pass
							Límpido/Incolor					pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
Tecnécio-99m Lot # E-20210703-001		03/07/2021 02:21	WG	Butanona	3MM Whatman			7		880	0		99,211 %	Pass
							Límpido e Incolor					pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
Tecnécio-99m Lot # E-20210703-002		03/07/2021 02:33	WG	Butanona	3MM Whatman			12		982	0		98,793 %	Pass
							Límpido e Incolor					pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
Tecnécio-99m Lot # E-20210703-003		03/07/2021 03:11	WG	Butanona	3MM Whatman			67		5609	0		98,82 %	Pass
							Límpido e Incolor					pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
Tecnécio-99m Lot # E-20210703-004		03/07/2021 04:15	WG	Butanona	3MM Whatman			11		1667	0		99,344 %	Pass
							Límpido e Incolor					pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
Tecnécio-99m Lot # E-20210703-005		03/07/2021 04:20	WG	Butanona	3MM Whatman			5		1589	0		99,686 %	Pass
							Límpido e Incolor					pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
Tecnécio-99m Lot # E-20210703-006		03/07/2021 04:21	WG	Butanona	3MM Whatman			33		2897	0		98,874 %	Pass
							Límpido e Incolor					pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
Tecnécio-99m Lot # E-20210703-007		03/07/2021 04:27	WG	Butanona	3MM Whatman			8,9		1236	0		99,285 %	Pass
							Límpido e Incolor					pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
Tecnécio-99m Lot # E-20210703-008		03/07/2021 04:28	WG	Butanona	3MM Whatman			12		3687	0		99,676 %	Pass
							Límpido e Incolor					pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
Tecnécio-99m Lot # E-20210703-009		03/07/2021 04:30	WG	Butanona	3MM Whatman			17		1169	0		98,567 %	Pass
							Límpido e Incolor					pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
Tecnécio-99m Lot # E-20210703-010		03/07/2021 04:51	WG	Butanona	3MM Whatman			11		674	0		98,394 %	Pass
							Límpido e Incolor					pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
Tecnécio-99m Lot # E-20210703-011		03/07/2021 04:51	WG	Butanona	3MM Whatman			3		1502	0		99,801 %	Pass
							Límpido e Incolor					pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
Tecnécio-99m Lot # E-20210703-012		03/07/2021 04:52	WG	Butanona	3MM Whatman			3		1187	0		99,748 %	Pass
							Límpido e Incolor					pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00