



# Kit Quality Control Report

## RPH CENTRAL PHARMA

CNPJ: 19.315.658/0001-10

R: Maestro Cardim, 769 - São Paulo/SP

Farm. RT: Priscila Larcher CRF/SP-74.417

Date: From 07/08/2021 To 07/08/2021

Product:	Manuf:	Calibration	D/T	QC By	Solution	Media	Origin	Middle	Top	Background	Particle Size/ OK?	% Tag	Pass Fail
DTPA-99mTc	RPH	07/08/2021	03:20	LG	Acetona	3MM Whatman	337		12	0		92,64 %	Pass
Lot # K-20210807-009					Normal Saline	3MM Whatman	6		147		pH (Low/Actual/High)	3,80 <= 5,00	<= 7,50
						Límpido/Incolor							
FITATO-99mTc		07/08/2021	03:25	LG	Butanona	3MM Whatman	36		0,5	0		98,63 %	Pass
Lot # K-20210807-010											pH (Low/Actual/High)	5,00 <= 6,00	<= 7,00
						Límpido/Incolor							
MAA-99mTc		07/08/2021	03:27	LG	Butanona	3MM Whatman	32,5		0	0		100,00 %	Pass
Lot # K-20210807-011											pH (Low/Actual/High)	3,00 <= 5,00	<= 6,00
						Turvo/Incolor							
MDP-99mTc		07/08/2021	02:51	LG	Acetona	3MM Whatman	623		1	0		95,768 %	Pass
Lot # K-20210807-005					Normal Saline	3MM Whatman	34		801		pH (Low/Actual/High)	4,00 <= 5,00	<= 7,80
						Límpido/Incolor							
MDP-99mTc		07/08/2021	02:53	LG	Acetona	3MM Whatman	480		1	0		97,598 %	Pass
Lot # K-20210807-006					Normal Saline	3MM Whatman	7		312		pH (Low/Actual/High)	4,00 <= 5,00	<= 7,80
						Límpido/Incolor							
MIBI-99mTc		07/08/2021	02:38	LG	Normal Saline	Sílica gel	956		15	0		96,987 %	Pass
Lot # K-20210807-001					Metanol	3MM Whatman	10		671		pH (Low/Actual/High)	5,00 <= 5,00	<= 6,00
						Límpido/Incolor							
MIBI-99mTc		07/08/2021	02:39	LG	Normal Saline	Sílica gel	960		15	0		97,557 %	Pass
Lot # K-20210807-002					Metanol	3MM Whatman	9		986		pH (Low/Actual/High)	5,00 <= 5,00	<= 6,00
						Límpido/Incolor							
MIBI-99mTc		07/08/2021	02:40	LG	Normal Saline	Sílica gel	800		13	0		97,799 %	Pass
Lot # K-20210807-004					Metanol	3MM Whatman	5		826		pH (Low/Actual/High)	5,00 <= 5,00	<= 6,00
						Límpido/Incolor							
MIBI-99mTc		07/08/2021	02:40	LG	Normal Saline	Sílica gel	899		18	0		95,392 %	Pass
Lot # K-20210807-003					Metanol	3MM Whatman	20		736		pH (Low/Actual/High)	5,00 <= 5,00	<= 6,00
						Límpido/Incolor							
MIBI-99mTc		07/08/2021	02:56	LG	Normal Saline	Sílica gel	693		10	0		98,451 %	Pass
Lot # K-20210807-007					Metanol	3MM Whatman	1		790		pH (Low/Actual/High)	5,00 <= 5,00	<= 6,00
						Límpido/Incolor							
MIBI-99mTc		07/08/2021	02:58	LG	Normal Saline	Sílica gel	429		6	0		97,337 %	Pass
Lot # K-20210807-008					Metanol	3MM Whatman	9		692		pH (Low/Actual/High)	5,00 <= 5,00	<= 6,00
						Límpido/Incolor							

Product:	Manuf:	Calibration D/T	Solution	Media	Origin	Middle	Top	Solution 2	Media 2	Origin 2	Top 2	% Tag	QC By	Pass Fail
Tecnécio-99m Lot # E-20210807-002		07/08/2021 02:35	LG	Butanona		3MM Whatman		0,1		5,86	0		98,322 %	Pass
												pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20210807-001		07/08/2021 02:35	LG	Butanona		3MM Whatman		0,03		3,53	0		99,157 %	Pass
												pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20210807-004		07/08/2021 02:43	LG	Butanona		3MM Whatman		0,1		9,5	0		98,958 %	Pass
												pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20210807-005		07/08/2021 02:44	LG	Butanona		3MM Whatman		0,1		5,86	0		98,322 %	Pass
												pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20210807-006		07/08/2021 04:26	JS	Butanona		3MM Whatman		2		921	0		99,783 %	Pass
												pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20210807-007		07/08/2021 04:31	JS	Butanona		3MM Whatman		1		561	0		99,822 %	Pass
												pH (Low/Actual/High)	5,00 <= 5,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20210807-008		07/08/2021 04:35	JS	Butanona		3MM Whatman		4		353	0		98,88 %	Pass
												pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20210807-009		07/08/2021 04:42	JS	Butanona		3MM Whatman		3		562	0		99,469 %	Pass
												pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20210807-010		07/08/2021 04:45	JS	Butanona		3MM Whatman		3		608	0		99,509 %	Pass
												pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20210807-011		07/08/2021 05:07	PP	Butanona		3MM Whatman		6		126512	0		99,995 %	Pass
												pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20210807-012		07/08/2021 05:09	PP	Butanona		3MM Whatman		2		3249	0		99,938 %	Pass
												pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20210807-013		07/08/2021 05:13	PP	Butanona		3MM Whatman		1		1652	0		99,94 %	Pass
												pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20210807-014		07/08/2021 05:16	PP	Butanona		3MM Whatman		7		1476	0		99,528 %	Pass
												pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
														Límpido e Incolor