



Kit Quality Control Report

RPH CENTRAL PHARMA

CNPJ: 19.315.658/0001-10

R: Maestro Cardim, 769 - São Paulo/SP

Farm. RT: Priscila Larcher CRF/SP-74.417

Date: From 13/08/2021 To 13/08/2021

Product:	Manuf:	Calibration	D/T	QC By	Solution	Media	Origin	Middle	Top	Background	Particle Size/ OK?	% Tag	Pass Fail
DMSA-99mTc Lot # K-20210813-032		13/08/2021	07:41	PP	Butanona Normal Saline	3MM Whatman Sílica gel	128 8		2 101	0	pH (Low/Actual/High)	91,122 %	Pass 2,00 <= 3,00 <= 3,00
						Límpido/Incolor							
DMSA-99mTc Lot # K-20210813-033		13/08/2021	07:43	PP	Butanona Normal Saline	3MM Whatman Sílica gel	103 10		1 119	0	pH (Low/Actual/High)	91,287 %	Pass 2,00 <= 3,00 <= 3,00
						Límpido/Incolor							
DMSA-99mTc Lot # K-20210813-034		13/08/2021	07:45	PP	Butanona Normal Saline	3MM Whatman Sílica gel	175 11		1 107	0	pH (Low/Actual/High)	90,11 %	Pass 2,00 <= 3,00 <= 3,00
						Límpido/Incolor							
DMSA-99mTc Lot # K-20210813-035		13/08/2021	08:00	PP	Butanona Normal Saline	3MM Whatman Sílica gel	91,7 11		0,9 111	0	pH (Low/Actual/High)	90,012 %	Pass 2,00 <= 3,00 <= 3,00
						Límpido/Incolor							
DMSA-99mTc Lot # K-20210813-036		13/08/2021	08:00	PP	Butanona Normal Saline	3MM Whatman Sílica gel	128 10		1 131	0	pH (Low/Actual/High)	92,133 %	Pass 2,00 <= 3,00 <= 3,00
						Límpido/Incolor							
DMSA-99mTc Lot # K-20210813-037		13/08/2021	08:01	PP	Butanona Normal Saline	3MM Whatman Sílica gel	146 5,5		2 66,5	0	pH (Low/Actual/High)	91,01 %	Pass 2,00 <= 3,00 <= 3,00
						Límpido/Incolor							
DMSA-99mTc Lot # K-20210813-044		13/08/2021	10:16	YD	Butanona Normal Saline	3MM Whatman Sílica gel	94,9 2		1,8 71,7	0	pH (Low/Actual/High)	95,425 %	Pass 2,00 <= 3,00 <= 3,00
						Límpido/Incolor							
DOTATOC-68Ga Lot # K-20210813-048		13/08/2021	12:45	PL	Citrato de Sódio 0,1M	Sílica gel	35,9		0,4	0	pH (Low/Actual/High)	98,898 %	Pass 4,00 <= 5,00 <= 8,00
						Límpido e Incolor							
DTPA-99mTc Lot # K-20210813-010	RPH	13/08/2021	03:50	FN	Acetona Normal Saline	3MM Whatman 3MM Whatman	128,5 5,6		1,2 132,1	0	pH (Low/Actual/High)	95,008 %	Pass 3,80 <= 5,00 <= 7,50
						Límpido/Incolor							
DTPA-99mTc Lot # K-20210813-019	RPH	13/08/2021	04:04	FN	Acetona Normal Saline	3MM Whatman 3MM Whatman	303 15		1 435	0	pH (Low/Actual/High)	96,338 %	Pass 3,80 <= 6,00 <= 7,50
						Límpido/Incolor							
DTPA-99mTc Lot # K-20210813-043	RPH	13/08/2021	10:13	JS	Acetona Normal Saline	3MM Whatman 3MM Whatman	85,3 1,5		0,9 158,6	0	pH (Low/Actual/High)	98,019 %	Pass 3,80 <= 5,00 <= 7,50
						Límpido/Incolor							

Product:	Manuf:	Calibration D/T	Solution	Media	Origin	Middle	Top	Solution 2	Media 2	Origin 2	Top 2	% Tag	QC By	Pass Fail
FITATO-99mTc Lot # K-20210813-009		13/08/2021 03:37 FN	Butanona			3MM Whatman		155,1		0,8	0		99,487 %	Pass
												pH (Low/Actual/High)	5,00 <= 6,00	<= 7,00
												Límpido/Incolor		
FITATO-99mTc Lot # K-20210813-027		13/08/2021 07:25 PP	Butanona			3MM Whatman		238		1	0		99,582 %	Pass
												pH (Low/Actual/High)	5,00 <= 6,00	<= 7,00
												Límpido/Incolor		
MAA-99mTc Lot # K-20210813-011		13/08/2021 03:52 FN	Butanona			3MM Whatman		89,9		0,8	0		99,118 %	Pass
												pH (Low/Actual/High)	3,00 <= 6,00	<= 6,00
												Turvo/Incolor		
MAA-99mTc Lot # K-20210813-012		13/08/2021 03:53 FN	Butanona			3MM Whatman		79,2		0,1	0		99,874 %	Pass
												pH (Low/Actual/High)	3,00 <= 6,00	<= 6,00
												Turvo/Incolor		
MAA-99mTc Lot # K-20210813-013		13/08/2021 03:56 FN	Butanona			3MM Whatman		137,4		1,5	0		98,92 %	Pass
												pH (Low/Actual/High)	3,00 <= 6,00	<= 6,00
												Turvo/Incolor		
MAA-99mTc Lot # K-20210813-015		13/08/2021 03:58 FN	Butanona			3MM Whatman		83,1		1	0		98,811 %	Pass
												pH (Low/Actual/High)	3,00 <= 6,00	<= 6,00
												Turvo/Incolor		
MAA-99mTc Lot # K-20210813-016		13/08/2021 03:59 FN	Butanona			3MM Whatman		76,6		1	0		98,711 %	Pass
												pH (Low/Actual/High)	3,00 <= 6,00	<= 6,00
												Turvo/Incolor		
MAA-99mTc Lot # K-20210813-017		13/08/2021 04:01 FN	Butanona			3MM Whatman		43,2		1,5	0		96,644 %	Pass
												pH (Low/Actual/High)	3,00 <= 6,00	<= 6,00
												Turvo/Incolor		
MAA-99mTc Lot # K-20210813-018		13/08/2021 04:02 FN	Butanona			3MM Whatman		552		5	0		99,102 %	Pass
												pH (Low/Actual/High)	3,00 <= 6,00	<= 6,00
												Turvo/Incolor		
MAA-99mTc Lot # K-20210813-028		13/08/2021 07:30 PP	Butanona			3MM Whatman		153		1	0		99,351 %	Pass
												pH (Low/Actual/High)	3,00 <= 6,00	<= 6,00
												Turvo/Incolor		
MAA-99mTc Lot # K-20210813-029		13/08/2021 07:33 PP	Butanona			3MM Whatman		218		2	0		99,091 %	Pass
												pH (Low/Actual/High)	3,00 <= 6,00	<= 6,00
												Turvo/Incolor		
MAA-99mTc Lot # K-20210813-030		13/08/2021 07:36 PP	Butanona			3MM Whatman		206		4	0		98,095 %	Pass
												pH (Low/Actual/High)	3,00 <= 6,00	<= 6,00
												Turvo/Incolor		
MAA-99mTc Lot # K-20210813-031		13/08/2021 07:38 PP	Butanona			3MM Whatman		162		1	0		99,387 %	Pass
												pH (Low/Actual/High)	3,00 <= 6,00	<= 6,00
												Turvo/Incolor		

Product:	Manuf:	Calibration D/T	Solution	Media	Origin	Middle	Top	Solution 2	Media 2	Origin 2	Top 2	% Tag	QC By	Pass Fail
MAA-99mTc Lot # K-20210813-045		13/08/2021 10:19 YD	Butanona		3MM	Whatman		217		1	0	99,541 %	Pass	
							Turvo/Incolor					pH (Low/Actual/High)	3,00 <= 6,00	<= 6,00
MAA-99mTc Lot # K-20210813-046		13/08/2021 10:20 YD	Butanona		3MM	Whatman		115,5		1,2	0	98,972 %	Pass	
							Turvo/Incolor					pH (Low/Actual/High)	3,00 <= 5,00	<= 6,00
MAA-99mTc Lot # K-20210813-047		13/08/2021 11:01 JS	Butanona		3MM	Whatman		58,3		1	0	98,314 %	Pass	
							Turvo/Incolor					pH (Low/Actual/High)	3,00 <= 5,00	<= 6,00
MDP-99mTc Lot # K-20210813-008		13/08/2021 03:35 FN	Acetona		3MM	Whatman		640		12	0	97,464 %	Pass	
			Normal Saline		3MM	Whatman		5		714		pH (Low/Actual/High)	4,00 <= 6,00	<= 7,80
							Límpido/Incolor							
MDP-99mTc Lot # K-20210813-020		13/08/2021 04:28 FN	Acetona		3MM	Whatman		403		1	0	99,274 %	Pass	
			Normal Saline		3MM	Whatman		2		416		pH (Low/Actual/High)	4,00 <= 6,00	<= 7,80
							Límpido/Incolor							
MDP-99mTc Lot # K-20210813-026		13/08/2021 07:21 PP	Acetona		3MM	Whatman		1181		3	0	99,494 %	Pass	
			Normal Saline		3MM	Whatman		3		1186		pH (Low/Actual/High)	4,00 <= 6,00	<= 7,80
							Límpido/Incolor							
MDP-99mTc Lot # K-20210813-025		13/08/2021 07:21 PP	Acetona		3MM	Whatman		872		1	0	99,26 %	Pass	
			Normal Saline		3MM	Whatman		3		477		pH (Low/Actual/High)	4,00 <= 6,00	<= 7,80
							Límpido/Incolor							
MDP-99mTc Lot # K-20210813-040		13/08/2021 09:38 YD	Acetona		3MM	Whatman		796		1	0	98,215 %	Pass	
			Normal Saline		3MM	Whatman		12		711		pH (Low/Actual/High)	4,00 <= 6,00	<= 7,80
							Límpido/Incolor							
MIBI-99mTc Lot # K-20210813-001		13/08/2021 02:02 FN	Normal Saline		Sílica gel			868		20	0	97,652 %	Pass	
			Metanol		3MM	Whatman		1		1041		pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
							Límpido/Incolor							
MIBI-99mTc Lot # K-20210813-002		13/08/2021 02:02 FN	Normal Saline		Sílica gel			900		19	0	97,459 %	Pass	
			Metanol		3MM	Whatman		2		420		pH (Low/Actual/High)	5,00 <= 5,00	<= 6,00
							Límpido/Incolor							
MIBI-99mTc Lot # K-20210813-003		13/08/2021 02:03 FN	Normal Saline		Sílica gel			579		10	0	97,615 %	Pass	
			Metanol		3MM	Whatman		5		723		pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
							Límpido/Incolor							
MIBI-99mTc Lot # K-20210813-004		13/08/2021 02:42 FN	Normal Saline		Sílica gel			1128		26	0	97,403 %	Pass	
			Metanol		3MM	Whatman		4		1158		pH (Low/Actual/High)	5,00 <= 5,00	<= 6,00
							Límpido/Incolor							
MIBI-99mTc Lot # K-20210813-005		13/08/2021 02:48 FN	Normal Saline		Sílica gel			691		11	0	98,101 %	Pass	
			Metanol		3MM	Whatman		2		600		pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
							Límpido/Incolor							

Product:	Manuf:	Calibration D/T	Solution	Media	Origin	Middle	Top	Solution 2	Media 2	Origin 2	Top 2	% Tag	QC By	Pass Fail
MIBI-99mTc Lot # K-20210813-006		13/08/2021 02:49 FN	Normal Saline Metanol		Sílica gel 3MM Whatman			803 8		10 676	0	pH (Low/Actual/High)	97,60 %	Pass 5,00 <= 6,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20210813-007		13/08/2021 03:33 FN	Normal Saline Metanol		Sílica gel 3MM Whatman			656 2		13 891	0	pH (Low/Actual/High)	97,833 %	Pass 5,00 <= 6,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20210813-014		13/08/2021 03:56 FN	Normal Saline Metanol		Sílica gel 3MM Whatman			619 3		10 802	0	pH (Low/Actual/High)	98,038 %	Pass 5,00 <= 6,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20210813-021		13/08/2021 04:40 FN	Normal Saline Metanol		Sílica gel 3MM Whatman			884 7		23 903	0	pH (Low/Actual/High)	96,695 %	Pass 5,00 <= 6,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20210813-022		13/08/2021 06:49 PP	Normal Saline Metanol		Sílica gel 3MM Whatman			2140 34		28 1631	0	pH (Low/Actual/High)	96,666 %	Pass 5,00 <= 6,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20210813-038		13/08/2021 08:21 FN	Normal Saline Metanol		Sílica gel 3MM Whatman			685 12		11 738	0	pH (Low/Actual/High)	96,82 %	Pass 5,00 <= 6,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20210813-039		13/08/2021 09:16 YD	Normal Saline Metanol		Sílica gel 3MM Whatman			607 7		10 587	0	pH (Low/Actual/High)	97,201 %	Pass 5,00 <= 5,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20210813-042		13/08/2021 10:08 YD	Normal Saline Metanol		Sílica gel 3MM Whatman			1006 183		12 3472	0	pH (Low/Actual/High)	93,814 %	Pass 5,00 <= 5,00 <= 6,00
Límpido/Incolor														
PSMA-68Ga Lot # K-20210813-023		13/08/2021 06:59 BB	Acetato de amônio 1M:Metanol		Sílica gel			0,3		13,8	0	pH (Low/Actual/High)	97,872 %	Pass 4,00 <= 5,00 <= 8,00
Límpido e Incolor														
PSMA-68Ga Lot # K-20210813-041		13/08/2021 09:54 YD	Acetato de amônio 1M:Metanol		Sílica gel			0,2		20,5	0	pH (Low/Actual/High)	99,034 %	Pass 4,00 <= 6,00 <= 8,00
Límpido e Incolor														
Tecnécio-99m Lot # E-20210813-001		13/08/2021 01:56 FN	Butanona		3MM Whatman			4		386	0	pH (Low/Actual/High)	98,974 %	Pass 5,00 <= 6,00 <= 6,00
Límpido e Incolor														
Tecnécio-99m Lot # E-20210813-002		13/08/2021 02:40 FN	Butanona		3MM Whatman			7		363	0	pH (Low/Actual/High)	98,108 %	Pass 5,00 <= 6,00 <= 6,00
Límpido e Incolor														
Tecnécio-99m Lot # E-20210813-003		13/08/2021 02:44 FN	Butanona		3MM Whatman			3		136,5	0	pH (Low/Actual/High)	97,849 %	Pass 5,00 <= 6,00 <= 6,00
Límpido e Incolor														

Product:	Manuf:	Calibration D/T	Solution	Media	Origin	Middle	Top	Solution 2	Media 2	Origin 2	Top 2	% Tag	QC By	Pass Fail
Tecnécio-99m Lot # E-20210813-004		13/08/2021 02:44	FN	Butanona		3MM Whatman		1		204	0		99,512 %	Pass
												pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20210813-005		13/08/2021 03:30	FN	Butanona		3MM Whatman		1		219	0		99,545 %	Pass
												pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20210813-006		13/08/2021 03:30	FN	Butanona		3MM Whatman		1		321	0		99,689 %	Pass
												pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20210813-007		13/08/2021 03:44	FN	Butanona		3MM Whatman		1,1		151,9	0		99,281 %	Pass
												pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20210813-008		13/08/2021 03:48	FN	Butanona		3MM Whatman		0,2		145,9	0		99,863 %	Pass
												pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20210813-009		13/08/2021 03:54	FN	Butanona		3MM Whatman		1		587	0		99,83 %	Pass
												pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20210813-010		13/08/2021 04:35	FN	Butanona		3MM Whatman		2		421	0		99,527 %	Pass
												pH (Low/Actual/High)	5,00 <= 5,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20210813-011		13/08/2021 06:42	PP	Butanona		3MM Whatman		32		3504	0		99,095 %	Pass
												pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20210813-013		13/08/2021 07:15	PP	Butanona		3MM Whatman		12		615	0		98,086 %	Pass
												pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20210813-014		13/08/2021 07:15	PP	Butanona		3MM Whatman		4		289	0		98,635 %	Pass
												pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20210813-015		13/08/2021 07:39	PP	Butanona		3MM Whatman		6		649	0		99,084 %	Pass
												pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20210813-016		13/08/2021 08:19	PP	Butanona		3MM Whatman		10		221	0		95,671 %	Pass
												pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20210813-017		13/08/2021 09:12	YD	Butanona		3MM Whatman		4		585	0		99,321 %	Pass
												pH (Low/Actual/High)	5,00 <= 5,00	<= 6,00
														Límpido e Incolor

Product:	Manuf:	Calibration D/T	Solution	Media	Origin	Middle	Top	Solution 2	Media 2	Origin 2	Top 2	% Tag	QC By	Pass Fail
Tecnécio-99m Lot # E-20210813-018		13/08/2021 09:32 YD	Butanona		3MM	Whatman		1		578	0			99,827 % Pass
												pH (Low/Actual/High)	5,00 <= 6,00 <= 6,00	
												Límpido e Incolor		
Tecnécio-99m Lot # E-20210813-020		13/08/2021 10:02 YD	Butanona		3MM	Whatman		3		1441	0			99,792 % Pass
												pH (Low/Actual/High)	5,00 <= 6,00 <= 6,00	
												Límpido e Incolor		
Tecnécio-99m Lot # E-20210813-021		13/08/2021 10:10 YD	Butanona		3MM	Whatman		1		1488	0			99,933 % Pass
												pH (Low/Actual/High)	5,00 <= 6,00 <= 6,00	
												Límpido e Incolor		