



Kit Quality Control Report

RPH CENTRAL PHARMA

CNPJ: 19.315.658/0001-10

R: Maestro Cardim, 769 - São Paulo/SP

Farm. RT: Priscila Larcher CRF/SP-74.417

Date: From 01/09/2021 To 01/09/2021

Product:	Manuf:	Calibration	D/T	QC By	Solution	Media	Origin	Middle	Top	Background	Particle Size/ OK?	% Tag	Pass Fail
DMSA-99mTc Lot # K-20210901-015		01/09/2021	04:07	PP	Butanona Normal Saline	3MM Whatman Sílica gel	390 12		8 347	0	pH (Low/Actual/High)	94,647 %	Pass <= 3,00 <= 3,00
						Límpido/Incolor							
DMSA-99mTc Lot # K-20210901-026		01/09/2021	07:34	FN	Butanona Normal Saline	3MM Whatman Sílica gel	171,3 10,7		1,5 131,4	0	pH (Low/Actual/High)	91,602 %	Pass <= 3,00 <= 3,00
						Límpido/Incolor							
DMSA-99mTc Lot # K-20210901-027		01/09/2021	07:37	FN	Butanona Normal Saline	3MM Whatman Sílica gel	150,1 10,2		2,8 133,6	0	pH (Low/Actual/High)	91,076 %	Pass <= 3,00 <= 3,00
						Límpido/Incolor							
DMSA-99mTc Lot # K-20210901-042		01/09/2021	10:38	JS	Butanona Normal Saline	3MM Whatman Sílica gel	128 2,3		1,7 103	0	pH (Low/Actual/High)	96,505 %	Pass <= 3,00 <= 3,00
						Límpido/Incolor							
DMSA-99mTc Lot # K-20210901-043		01/09/2021	10:40	JS	Butanona Normal Saline	3MM Whatman Sílica gel	76 1,3		0,9 98,8	0	pH (Low/Actual/High)	97,531 %	Pass <= 3,00 <= 3,00
						Límpido/Incolor							
DOTATOC-68Ga Lot # K-20210901-036		01/09/2021	09:39	JS	Citrato de Sódio 0,1M	Sílica gel	13,8		0,4	0	pH (Low/Actual/High)	97,183 %	Pass <= 6,00 <= 8,00
						Límpido e Incolor							
DOTATOC-68Ga Lot # K-20210901-039		01/09/2021	10:26	JS	Citrato de Sódio 0,1M	Sílica gel	14,9		0,2	0	pH (Low/Actual/High)	98,675 %	Pass <= 5,00 <= 8,00
						Límpido e Incolor							
DOTATOC-68Ga Lot # K-20210901-056		01/09/2021	15:44	JS	Citrato de Sódio 0,1M	Sílica gel	9,7		0,4	0	pH (Low/Actual/High)	96,04 %	Pass <= 5,00 <= 8,00
						Límpido e Incolor							
DTPA-99mTc Lot # K-20210901-018	RPH	01/09/2021	04:18	PP	Acetona Normal Saline	3MM Whatman 3MM Whatman	798 29		22 875	0	pH (Low/Actual/High)	94,109 %	Pass <= 6,00 <= 7,50
						Límpido/Incolor							
DTPA-99mTc Lot # K-20210901-030	RPH	01/09/2021	07:44	FN	Acetona Normal Saline	3MM Whatman 3MM Whatman	144,1 5		1,4 222	0	pH (Low/Actual/High)	96,835 %	Pass <= 6,00 <= 7,50
						Límpido/Incolor							
DTPA-99mTc Lot # K-20210901-040	RPH	01/09/2021	10:33	JS	Acetona Normal Saline	3MM Whatman 3MM Whatman	516 8		1 382	0	pH (Low/Actual/High)	97,755 %	Pass <= 5,00 <= 7,50
						Límpido/Incolor							

Product:	Manuf:	Calibration D/T	Solution	Media	Origin	Middle	Top	Solution 2	Media 2	Origin 2	Top 2	% Tag	QC By	Pass Fail
FITATO-99mTc Lot # K-20210901-019		01/09/2021 04:20 PP	Butanona		3MM	Whatman		496		2	0		99,598 %	Pass
											pH (Low/Actual/High)	5,00 <= 6,00	<= 7,00	
											Límpido/Incolor			
FITATO-99mTc Lot # K-20210901-038		01/09/2021 10:22 JS	Butanona		3MM	Whatman		282		2	0		99,296 %	Pass
											pH (Low/Actual/High)	5,00 <= 7,00	<= 7,00	
											Límpido/Incolor			
FITATO-99mTc Lot # K-20210901-055		01/09/2021 15:07 JS	Butanona		3MM	Whatman		56		3	0		94,915 %	Pass
											pH (Low/Actual/High)	5,00 <= 7,00	<= 7,00	
											Límpido/Incolor			
MAA-99mTc Lot # K-20210901-011		01/09/2021 04:00 PP	Butanona		3MM	Whatman		305		3	0		99,026 %	Pass
											pH (Low/Actual/High)	3,00 <= 6,00	<= 6,00	
											Turvo/Incolor			
MAA-99mTc Lot # K-20210901-012		01/09/2021 04:01 PP	Butanona		3MM	Whatman		475		4	0		99,165 %	Pass
											pH (Low/Actual/High)	3,00 <= 6,00	<= 6,00	
											Turvo/Incolor			
MAA-99mTc Lot # K-20210901-013		01/09/2021 04:03 PP	Butanona		3MM	Whatman		326		5	0		98,489 %	Pass
											pH (Low/Actual/High)	3,00 <= 6,00	<= 6,00	
											Turvo/Incolor			
MAA-99mTc Lot # K-20210901-014		01/09/2021 04:04 PP	Butanona		3MM	Whatman		347		1	0		99,713 %	Pass
											pH (Low/Actual/High)	3,00 <= 6,00	<= 6,00	
											Turvo/Incolor			
MAA-99mTc Lot # K-20210901-028		01/09/2021 07:39 FN	Butanona		3MM	Whatman		223		1	0		99,554 %	Pass
											pH (Low/Actual/High)	3,00 <= 6,00	<= 6,00	
											Turvo/Incolor			
MAA-99mTc Lot # K-20210901-029		01/09/2021 07:41 FN	Butanona		3MM	Whatman		167,8		3,5	0		97,957 %	Pass
											pH (Low/Actual/High)	3,00 <= 6,00	<= 6,00	
											Turvo/Incolor			
MAA-99mTc Lot # K-20210901-041		01/09/2021 10:35 JS	Butanona		3MM	Whatman		235		2	0		99,156 %	Pass
											pH (Low/Actual/High)	3,00 <= 5,00	<= 6,00	
											Turvo/Incolor			
MAA-99mTc Lot # K-20210901-045		01/09/2021 10:44 JS	Butanona		3MM	Whatman		60		1	0		98,361 %	Pass
											pH (Low/Actual/High)	3,00 <= 5,00	<= 6,00	
											Turvo/Incolor			
MAA-99mTc Lot # K-20210901-046		01/09/2021 10:46 JS	Butanona		3MM	Whatman		126		0,4	0		99,684 %	Pass
											pH (Low/Actual/High)	3,00 <= 5,00	<= 6,00	
											Turvo/Incolor			
MAA-99mTc Lot # K-20210901-047		01/09/2021 10:49 JS	Butanona		3MM	Whatman		132		1,2	0		99,099 %	Pass
											pH (Low/Actual/High)	3,00 <= 5,00	<= 6,00	
											Turvo/Incolor			

Product:	Manuf:	Calibration D/T	Solution	Media	Origin	Middle	Top	Solution 2	Media 2	Origin 2	Top 2	% Tag	QC By	Pass Fail
MAA-99mTc Lot # K-20210901-048		01/09/2021 11:06 JS	Butanona		3MM	Whatman		159		4	0	97,546 %	Pass	
							Turvo/Incolor				pH (Low/Actual/High)	3,00 <= 5,00 <= 6,00		
MAA-99mTc Lot # K-20210901-051		01/09/2021 11:29 JS	Butanona		3MM	Whatman		137		2	0	98,561 %	Pass	
							Turvo/Incolor				pH (Low/Actual/High)	3,00 <= 5,00 <= 6,00		
MDP-99mTc Lot # K-20210901-016		01/09/2021 04:10 PP	Acetona		3MM	Whatman		1000		67	0	90,089 %	Pass	
			Normal Saline		3MM	Whatman		30		796	pH (Low/Actual/High)	4,00 <= 6,00 <= 7,80		
							Límpido/Incolor							
MDP-99mTc Lot # K-20210901-017		01/09/2021 04:15 BB	Acetona		3MM	Whatman		980		3	0	96,547 %	Pass	
			Normal Saline		3MM	Whatman		30		923	pH (Low/Actual/High)	4,00 <= 6,00 <= 7,80		
							Límpido/Incolor							
MDP-99mTc Lot # K-20210901-033		01/09/2021 08:07 PP	Acetona		3MM	Whatman		970		11	0	98,50 %	Pass	
			Normal Saline		3MM	Whatman		3		790	pH (Low/Actual/High)	4,00 <= 6,00 <= 7,80		
							Límpido/Incolor							
MDP-99mTc Lot # K-20210901-034		01/09/2021 08:10 PP	Acetona		3MM	Whatman		752		3	0	98,432 %	Pass	
			Normal Saline		3MM	Whatman		9		760	pH (Low/Actual/High)	4,00 <= 6,00 <= 7,80		
							Límpido/Incolor							
MDP-99mTc Lot # K-20210901-035		01/09/2021 08:14 PP	Acetona		3MM	Whatman		357		7	0	97,112 %	Pass	
			Normal Saline		3MM	Whatman		3		308	pH (Low/Actual/High)	4,00 <= 6,00 <= 7,80		
							Límpido/Incolor							
MDP-99mTc Lot # K-20210901-044		01/09/2021 10:42 JS	Acetona		3MM	Whatman		483		1	0	99,143 %	Pass	
			Normal Saline		3MM	Whatman		3		458	pH (Low/Actual/High)	4,00 <= 5,00 <= 7,80		
							Límpido/Incolor							
MDP-99mTc Lot # K-20210901-052		01/09/2021 11:31 JS	Acetona		3MM	Whatman		1083		7	0	97,953 %	Pass	
			Normal Saline		3MM	Whatman		6		421	pH (Low/Actual/High)	4,00 <= 6,00 <= 7,80		
							Límpido/Incolor							
MIBI-99mTc Lot # K-20210901-003		01/09/2021 02:33 PP	Normal Saline		Sílica gel			1990		25	0	98,474 %	Pass	
			Metanol		3MM	Whatman		5		1748	pH (Low/Actual/High)	5,00 <= 6,00 <= 6,00		
							Límpido/Incolor							
MIBI-99mTc Lot # K-20210901-004		01/09/2021 02:35 PP	Normal Saline		Sílica gel			1563		32	0	96,988 %	Pass	
			Metanol		3MM	Whatman		28		2757	pH (Low/Actual/High)	5,00 <= 6,00 <= 6,00		
							Límpido/Incolor							
MIBI-99mTc Lot # K-20210901-005		01/09/2021 02:56 PP	Normal Saline		Sílica gel			712		12	0	96,188 %	Pass	
			Metanol		3MM	Whatman		68		3088	pH (Low/Actual/High)	5,00 <= 6,00 <= 6,00		
							Límpido/Incolor							
MIBI-99mTc Lot # K-20210901-007		01/09/2021 02:57 PP	Normal Saline		Sílica gel			906		8	0	98,924 %	Pass	
			Metanol		3MM	Whatman		5		2480	pH (Low/Actual/High)	5,00 <= 6,00 <= 6,00		
							Límpido/Incolor							

Product:	Manuf:	Calibration D/T	Solution	Media	Origin	Middle	Top	Solution 2	Media 2	Origin 2	Top 2	% Tag	QC By	Pass Fail
MIBI-99mTc Lot # K-20210901-008		01/09/2021 03:53 PP	Normal Saline Metanol		Sílica gel 3MM Whatman			2863 3		32 2398	0	pH (Low/Actual/High)	98,77 %	Pass 5,00 <= 6,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20210901-009		01/09/2021 03:54 PP	Normal Saline Metanol		Sílica gel 3MM Whatman			2353 35		29 3049	0	pH (Low/Actual/High)	97,648 %	Pass 5,00 <= 6,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20210901-020		01/09/2021 04:23 PP	Normal Saline Metanol		Sílica gel 3MM Whatman			2474 10		23 1755	0	pH (Low/Actual/High)	98,512 %	Pass 5,00 <= 6,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20210901-023		01/09/2021 07:05 FN	Normal Saline Metanol		Sílica gel 3MM Whatman			1228 6		25 1323	0	pH (Low/Actual/High)	97,553 %	Pass 5,00 <= 5,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20210901-024		01/09/2021 07:12 FN	Normal Saline Metanol		Sílica gel 3MM Whatman			769 9		2 773	0	pH (Low/Actual/High)	98,59 %	Pass 5,00 <= 6,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20210901-025		01/09/2021 07:13 FN	Normal Saline Metanol		Sílica gel 3MM Whatman			957 6		7 803	0	pH (Low/Actual/High)	98,532 %	Pass 5,00 <= 6,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20210901-031		01/09/2021 07:51 FN	Normal Saline Metanol		Sílica gel 3MM Whatman			694 2		13 845	0	pH (Low/Actual/High)	97,925 %	Pass 5,00 <= 6,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20210901-032		01/09/2021 07:55 FN	Normal Saline Metanol		Sílica gel 3MM Whatman			834 1		8 1304	0	pH (Low/Actual/High)	98,973 %	Pass 5,00 <= 6,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20210901-037		01/09/2021 10:11 JS	Normal Saline Metanol		Sílica gel 3MM Whatman			1139 1		16 1019	0	pH (Low/Actual/High)	98,517 %	Pass 5,00 <= 5,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20210901-049		01/09/2021 11:08 JS	Normal Saline Metanol		Sílica gel 3MM Whatman			382 16		13 530	0	pH (Low/Actual/High)	93,778 %	Pass 5,00 <= 5,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20210901-050		01/09/2021 11:24 JS	Normal Saline Metanol		Sílica gel 3MM Whatman			567 12		33 376	0	pH (Low/Actual/High)	91,407 %	Pass 5,00 <= 5,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20210901-053		01/09/2021 11:36 JS	Normal Saline Metanol		Sílica gel 3MM Whatman			315 4		9 1154	0	pH (Low/Actual/High)	96,877 %	Pass 5,00 <= 5,00 <= 6,00
Límpido/Incolor														
PIRO-99mTc Lot # K-20210901-021		01/09/2021 04:26 PP	Butanona Normal Saline		3MM Whatman 3MM Whatman			756 7		2 704	0	pH (Low/Actual/High)	98,752 %	Pass 5,00 <= 6,00 <= 7,00
Límpido/Incolor														

Product:	Manuf:	Calibration	D/T	Solution	Media	Origin	Middle	Top	Solution 2	Media 2	Origin 2	Top 2	% Tag	QC By	Pass Fail
PSMA-177Lu Lot # K-20210901-001	01/09/2021	01:59	BB	Citrato de Sódio 0,1M		Sílica gel			417		5	0	98,815 %		Pass
								Limpido				pH (Low/Actual/High)	4,00 <= 4,50	<= 5,00	
QC Notes: HPLC: 98,44%															
PSMA-177Lu Lot # K-20210901-002	01/09/2021	02:10	BB	Citrato de Sódio 0,1M		Sílica gel			681		2	0	99,707 %		Pass
								Limpido				pH (Low/Actual/High)	4,00 <= 4,50	<= 5,00	
QC Notes: HPLC: 96,24%															
PSMA-68Ga Lot # K-20210901-022	01/09/2021	06:49	FN	Acetato de amônio 1M:Metanol		Sílica gel			0,1		15,37	0	99,354 %		Pass
								Límpido e Incolor				pH (Low/Actual/High)	4,00 <= 6,00	<= 8,00	
PSMA-68Ga Lot # K-20210901-054	01/09/2021	12:40	JS	Acetato de amônio 1M:Metanol		Sílica gel			0,2		25,9	0	99,234 %		Pass
								Límpido e Incolor				pH (Low/Actual/High)	4,00 <= 5,00	<= 8,00	
Tecnécio-99m Lot # E-20210901-001	01/09/2021	02:26	PP	Butanona		3MM Whatman			8		876	0	99,095 %		Pass
								Límpido e Incolor				pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00	
Tecnécio-99m Lot # E-20210901-002	01/09/2021	02:27	PP	Butanona		3MM Whatman			24		2937	0	99,189 %		Pass
								Límpido e Incolor				pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00	
Tecnécio-99m Lot # E-20210901-003	01/09/2021	02:51	PP	Butanona		3MM Whatman			5		1174	0	99,576 %		Pass
								Límpido e Incolor				pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00	
Tecnécio-99m Lot # E-20210901-004	01/09/2021	02:52	PP	Butanona		3MM Whatman			4		401	0	99,012 %		Pass
								Límpido e Incolor				pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00	
Tecnécio-99m Lot # E-20210901-005	01/09/2021	03:11	PP	Butanona		3MM Whatman			30		2322	0	98,724 %		Pass
								Límpido e Incolor				pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00	
Tecnécio-99m Lot # E-20210901-006	01/09/2021	03:49	PP	Butanona		3MM Whatman			18		1861	0	99,042 %		Pass
								Límpido e Incolor				pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00	
Tecnécio-99m Lot # E-20210901-007	01/09/2021	03:50	PP	Butanona		3MM Whatman			2		982	0	99,797 %		Pass
								Límpido e Incolor				pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00	
Tecnécio-99m Lot # E-20210901-008	01/09/2021	04:07	PP	Butanona		3MM Whatman			12		1481	0	99,196 %		Pass
								Límpido e Incolor				pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00	

Product:	Manuf:	Calibration D/T	Solution	Media	Origin	Middle	Top	Solution 2	Media 2	Origin 2	Top 2	% Tag	QC By	Pass Fail
Tecnécio-99m Lot # E-20210901-009		01/09/2021 04:13	PP	Butanona		3MM Whatman		13		250	0		95,057 %	Pass
												pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20210901-010		01/09/2021 04:16	PP	Butanona		3MM Whatman		20		726	0		97,319 %	Pass
												pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20210901-011		01/09/2021 04:21	PP	Butanona		3MM Whatman		22		907	0		97,632 %	Pass
												pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20210901-014		01/09/2021 06:56	FN	Butanona		3MM Whatman		3		301	0		99,013 %	Pass
												pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20210901-013		01/09/2021 07:03	FN	Butanona		3MM Whatman		1		1027	0		99,903 %	Pass
												pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20210901-015		01/09/2021 07:08	FN	Butanona		3MM Whatman		6		258	0		97,727 %	Pass
												pH (Low/Actual/High)	5,00 <= 5,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20210901-016		01/09/2021 07:32	FN	Butanona		3MM Whatman		3,4		139,5	0		97,621 %	Pass
												pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20210901-017		01/09/2021 07:48	FN	Butanona		3MM Whatman		2		163,8	0		98,794 %	Pass
												pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20210901-018		01/09/2021 07:53	FN	Butanona		3MM Whatman		4		326	0		98,788 %	Pass
												pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20210901-019		01/09/2021 08:04	FN	Butanona		3MM Whatman		3,2		186,8	0		98,316 %	Pass
												pH (Low/Actual/High)	5,00 <= 5,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20210901-020		01/09/2021 08:08	FN	Butanona		3MM Whatman		1		303	0		99,671 %	Pass
												pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20210901-021		01/09/2021 08:12	PP	Butanona		3MM Whatman		3		99	0		97,059 %	Pass
												pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20210901-023		01/09/2021 10:02	JS	Butanona		3MM Whatman		2		203	0		99,024 %	Pass
												pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
														Límpido e Incolor

Product:	Manuf:	Calibration D/T	Solution	Media	Origin	Middle	Top	Solution 2	Media 2	Origin 2	Top 2	% Tag	QC By	Pass Fail
Tecnécio-99m Lot # E-20210901-024		01/09/2021 10:17 JS	Butanona		3MM Whatman			2		361	0		99,449 %	Pass
											pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00	
											Límpido e Incolor			
Tecnécio-99m Lot # E-20210901-025		01/09/2021 10:27 JS	Butanona		3MM Whatman			1		248	0		99,598 %	Pass
											pH (Low/Actual/High)	5,00 <= 5,00	<= 6,00	
											Límpido e Incolor			
Tecnécio-99m Lot # E-20210901-027		01/09/2021 10:36 JS	Butanona		3MM Whatman			4		223	0		98,238 %	Pass
											pH (Low/Actual/High)	5,00 <= 5,00	<= 6,00	
											Límpido e Incolor			
Tecnécio-99m Lot # E-20210901-028		01/09/2021 11:03 JS	Butanona		3MM Whatman			3		98	0		97,03 %	Pass
											pH (Low/Actual/High)	5,00 <= 5,00	<= 6,00	
											Límpido e Incolor			
Tecnécio-99m Lot # E-20210901-029		01/09/2021 11:20 JS	Butanona		3MM Whatman			3		267	0		98,889 %	Pass
											pH (Low/Actual/High)	5,00 <= 5,00	<= 6,00	
											Límpido e Incolor			
Tecnécio-99m Lot # E-20210901-030		01/09/2021 11:26 JS	Butanona		3MM Whatman			11		271	0		96,099 %	Pass
											pH (Low/Actual/High)	5,00 <= 5,00	<= 6,00	
											Límpido e Incolor			
Tecnécio-99m Lot # E-20210901-031		01/09/2021 11:33 JS	Butanona		3MM Whatman			1		105	0		99,057 %	Pass
											pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00	
											Límpido e Incolor			
Tecnécio-99m Lot # E-20210901-032		01/09/2021 11:45 JS	Butanona		3MM Whatman			4		830	0		99,52 %	Pass
											pH (Low/Actual/High)	5,00 <= 5,00	<= 6,00	
											Límpido e Incolor			