



Kit Quality Control Report

RPH CENTRAL PHARMA

CNPJ: 19.315.658/0001-10

R: Maestro Cardim, 769 - São Paulo/SP

Farm. RT: Priscila Larcher CRF/SP-74.417

Date: From 02/11/2021 To 02/11/2021

Product:	Manuf:	Calibration	D/T	QC By	Solution	Media	Origin	Middle	Top	Background	Particle Size/ OK?	% Tag	Pass Fail
DMSA-99mTc Lot # K-20211102-010		02/11/2021	04:00	MB	Butanona Normal Saline	3MM Whatman Sílica gel	338 9		2 302	0	pH (Low/Actual/High)	96,518 %	Pass 2,00 <= 3,00 <= 3,00
						Límpido/Incolor							
DMSA-99mTc Lot # K-20211102-016		02/11/2021	07:06	MB	Butanona Normal Saline	3MM Whatman Sílica gel	2780 8		145 432	0	pH (Low/Actual/High)	93,225 %	Pass 2,00 <= 3,00 <= 3,00
						Límpido/Incolor							
DOTATOC-68Ga Lot # K-20211102-015		02/11/2021	06:47	JS	Citrato de Sódio 0,1M	Sílica gel	58,7		0,1	0	pH (Low/Actual/High)	99,83 %	Pass 4,00 <= 5,00 <= 8,00
						Límpido e Incolor							
DOTATOC-68Ga Lot # K-20211102-018		02/11/2021	09:40	JS	Citrato de Sódio 0,1M	Sílica gel	31,1		0,4	0	pH (Low/Actual/High)	98,73 %	Pass 4,00 <= 5,00 <= 8,00
						Límpido e Incolor							
DOTATOC-68Ga Lot # K-20211102-019		02/11/2021	12:38	JS	Citrato de Sódio 0,1M	Sílica gel	15,2		0,5	0	pH (Low/Actual/High)	96,815 %	Pass 4,00 <= 5,00 <= 8,00
						Límpido e Incolor							
DOTATOC-68Ga Lot # K-20211102-020		02/11/2021	13:31	JS	Citrato de Sódio 0,1M	Sílica gel	41,8		0,3	0	pH (Low/Actual/High)	99,287 %	Pass 4,00 <= 5,00 <= 8,00
						Límpido e Incolor							
DTPA-99mTc Lot # K-20211102-006	RPH	02/11/2021	02:58	MB	Acetona Normal Saline	3MM Whatman 3MM Whatman	522 17		8 620	0	pH (Low/Actual/High)	95,822 %	Pass 3,80 <= 5,00 <= 7,50
						Límpido/Incolor							
FITATO-99mTc Lot # K-20211102-007		02/11/2021	03:04	MB	Butanona	3MM Whatman	285		1	0	pH (Low/Actual/High)	99,65 %	Pass 5,00 <= 5,00 <= 7,00
						Límpido/Incolor							
MAA-99mTc Lot # K-20211102-012		02/11/2021	04:13	MB	Butanona	3MM Whatman	137,1		0,7	0	pH (Low/Actual/High)	99,492 %	Pass 3,00 <= 5,00 <= 6,00
						Turvo/Incolor							
MAA-99mTc Lot # K-20211102-011		02/11/2021	04:13	MB	Butanona	3MM Whatman	94,2		0,9	0	pH (Low/Actual/High)	99,054 %	Pass 3,00 <= 5,00 <= 6,00
						Turvo/Incolor							
MAA-99mTc Lot # K-20211102-013		02/11/2021	04:25	MB	Butanona	3MM Whatman	345		5	0	pH (Low/Actual/High)	98,571 %	Pass 3,00 <= 5,00 <= 6,00
						Turvo/Incolor							

Product:	Manuf:	Calibration D/T	Solution	Media	Origin	Middle	Top	Solution 2	Media 2	Origin 2	Top 2	% Tag	QC By	Pass Fail
MAA-99mTc Lot # K-20211102-014		02/11/2021 04:39	MB	Butanona		3MM Whatman		508		5	0	99,025 %	Pass	
											pH (Low/Actual/High)	3,00 <= 5,00	<= 6,00	
											Turvo/Incolor			
MAA-99mTc Lot # K-20211102-017		02/11/2021 07:39	MB	Butanona		3MM Whatman		94,3		1	0	98,951 %	Pass	
											pH (Low/Actual/High)	3,00 <= 5,00	<= 6,00	
											Turvo/Incolor			
MDP-99mTc Lot # K-20211102-004		02/11/2021 02:50	MB	Acetona		3MM Whatman		869		2	0	99,049 %	Pass	
				Normal Saline		3MM Whatman		7		963	pH (Low/Actual/High)	4,00 <= 5,00	<= 7,80	
											Límpido/Incolor			
MDP-99mTc Lot # K-20211102-005		02/11/2021 02:51	MB	Acetona		3MM Whatman		1199		1	0	98,608 %	Pass	
				Normal Saline		3MM Whatman		17		1282	pH (Low/Actual/High)	4,00 <= 5,00	<= 7,80	
											Límpido/Incolor			
MIBI-99mTc Lot # K-20211102-001		02/11/2021 02:42	MB	Normal Saline		Sílica gel		228		7	0	96,563 %	Pass	
				Metanol		3MM Whatman		1		217	pH (Low/Actual/High)	5,00 <= 5,00	<= 6,00	
											Límpido/Incolor			
MIBI-99mTc Lot # K-20211102-002		02/11/2021 02:43	MB	Normal Saline		Sílica gel		1126		16	0	96,271 %	Pass	
				Metanol		3MM Whatman		28		1175	pH (Low/Actual/High)	5,00 <= 5,00	<= 6,00	
											Límpido/Incolor			
MIBI-99mTc Lot # K-20211102-003		02/11/2021 02:44	MB	Normal Saline		Sílica gel		1256		24	0	96,915 %	Pass	
				Metanol		3MM Whatman		21		1715	pH (Low/Actual/High)	5,00 <= 5,00	<= 6,00	
											Límpido/Incolor			
MIBI-99mTc Lot # K-20211102-008		02/11/2021 03:54	MB	Normal Saline		Sílica gel		1842		32	0	96,093 %	Pass	
				Metanol		3MM Whatman		32		1423	pH (Low/Actual/High)	5,00 <= 5,00	<= 6,00	
											Límpido/Incolor			
MIBI-99mTc Lot # K-20211102-009		02/11/2021 03:54	MB	Normal Saline		Sílica gel		1586		1	0	99,231 %	Pass	
				Metanol		3MM Whatman		14		1970	pH (Low/Actual/High)	5,00 <= 5,00	<= 6,00	
											Límpido/Incolor			
Tecnécio-99m Lot # E-20211102-001		02/11/2021 02:36	MB	Butanona		3MM Whatman		16		510	0	96,958 %	Pass	
											pH (Low/Actual/High)	5,00 <= 5,00	<= 6,00	
											Límpido e Incolor			
Tecnécio-99m Lot # E-20211102-002		02/11/2021 02:46	MB	Butanona		3MM Whatman		1		417	0	99,761 %	Pass	
											pH (Low/Actual/High)	5,00 <= 5,00	<= 6,00	
											Límpido e Incolor			
Tecnécio-99m Lot # E-20211102-003		02/11/2021 03:51	MB	Butanona		3MM Whatman		6		704	0	99,155 %	Pass	
											pH (Low/Actual/High)	5,00 <= 5,00	<= 6,00	
											Límpido e Incolor			
Tecnécio-99m Lot # E-20211102-004		02/11/2021 05:02	JS	Butanona		3MM Whatman		1		386	0	99,742 %	Pass	
											pH (Low/Actual/High)	5,00 <= 5,00	<= 6,00	
											Límpido e Incolor			