



Kit Quality Control Report

RPH CENTRAL PHARMA

CNPJ: 19.315.658/0001-10

R: Maestro Cardim, 769 - São Paulo/SP

Farm. RT: Priscila Larcher CRF/SP-74.417

Date: From 29/11/2021 To 29/11/2021

Product:	Manuf:	Calibration	D/T	QC By	Solution	Media	Origin	Middle	Top	Background	Particle Size/ OK?	% Tag	Pass Fail
DMSA-99mTc Lot # K-20211129-025		29/11/2021	07:32	PM	Butanona Normal Saline	3MM Whatman Sílica gel	86,8 2,7		0,8 93,8	0	pH (Low/Actual/High)	96,289 %	Pass
Límpido/Incolor													
DMSA-99mTc Lot # K-20211129-030		29/11/2021	08:06	PM	Butanona Normal Saline	3MM Whatman Sílica gel	150,4 7,9		3,6 140,8	0	pH (Low/Actual/High)	92,35 %	Pass
Límpido/Incolor													
DMSA-99mTc Lot # K-20211129-038		29/11/2021	10:51	VS	Butanona Normal Saline	3MM Whatman Sílica gel	858 4		10 1375	0	pH (Low/Actual/High)	98,558 %	Pass
Límpido/Incolor													
DMSA-99mTc Lot # K-20211129-039		29/11/2021	10:55	VS	Butanona Normal Saline	3MM Whatman Sílica gel	504 7		9 223	0	pH (Low/Actual/High)	95,202 %	Pass
Límpido/Incolor													
DOTATOC-68Ga Lot # K-20211129-034		29/11/2021	09:34	PM	Citrato de Sódio 0,1M	Sílica gel	21		0,2	0	pH (Low/Actual/High)	99,057 %	Pass
Límpido e Incolor													
DTPA-99mTc Lot # K-20211129-014	RPH	29/11/2021	04:16	MB	Acetona Normal Saline	3MM Whatman 3MM Whatman	407 2		1 412	0	pH (Low/Actual/High)	99,272 %	Pass
Límpido/Incolor													
DTPA-99mTc Lot # K-20211129-015	RPH	29/11/2021	04:17	MB	Acetona Normal Saline	3MM Whatman 3MM Whatman	409 10		3 343	0	pH (Low/Actual/High)	96,439 %	Pass
Límpido/Incolor													
DTPA-99mTc Lot # K-20211129-037	RPH	29/11/2021	10:46	VS	Acetona Normal Saline	3MM Whatman 3MM Whatman	976 7		1 1068	0	pH (Low/Actual/High)	99,246 %	Pass
Límpido/Incolor													
FITATO-99mTc Lot # K-20211129-012		29/11/2021	04:13	MB	Butanona	3MM Whatman	402		2	0	pH (Low/Actual/High)	99,505 %	Pass
Límpido/Incolor													
FITATO-99mTc Lot # K-20211129-024		29/11/2021	07:29	PM	Butanona	3MM Whatman	279		1	0	pH (Low/Actual/High)	99,643 %	Pass
Límpido/Incolor													
FITATO-99mTc Lot # K-20211129-041		29/11/2021	11:02	YD	Butanona	3MM Whatman	261		3	0	pH (Low/Actual/High)	98,864 %	Pass
Límpido/Incolor													

Product:	Manuf:	Calibration D/T	Solution	Media	Origin	Middle	Top	Solution 2	Media 2	Origin 2	Top 2	% Tag	QC By	Pass Fail
MAA-99mTc Lot # K-20211129-011		29/11/2021 04:05	MB	Butanona		3MM Whatman		100,7		0,1	0	99,901 %	Pass	
							Turvo/Incolor				pH (Low/Actual/High)	3,00 <= 5,00	<= 6,00	
MAA-99mTc Lot # K-20211129-028		29/11/2021 08:03	PM	Butanona		3MM Whatman		132,5		1,7	0	98,733 %	Pass	
							Turvo/Incolor				pH (Low/Actual/High)	3,00 <= 5,00	<= 6,00	
MAA-99mTc Lot # K-20211129-029		29/11/2021 08:05	PM	Butanona		3MM Whatman		89,8		0,7	0	99,227 %	Pass	
							Turvo/Incolor				pH (Low/Actual/High)	3,00 <= 5,00	<= 6,00	
MAA-99mTc Lot # K-20211129-033		29/11/2021 08:20	PM	Butanona		3MM Whatman		134,4		1,3	0	99,042 %	Pass	
							Turvo/Incolor				pH (Low/Actual/High)	3,00 <= 5,00	<= 6,00	
MAA-99mTc Lot # K-20211129-044		29/11/2021 12:39	VS	Butanona		3MM Whatman		262		10	0	96,324 %	Pass	
							Turvo/Incolor				pH (Low/Actual/High)	3,00 <= 6,00	<= 6,00	
MAA-99mTc Lot # K-20211129-045		29/11/2021 12:41	VS	Butanona		3MM Whatman		241		8	0	96,787 %	Pass	
							Turvo/Incolor				pH (Low/Actual/High)	3,00 <= 6,00	<= 6,00	
MAA-99mTc Lot # K-20211129-046		29/11/2021 12:43	VS	Butanona		3MM Whatman		355		12	0	96,73 %	Pass	
							Turvo/Incolor				pH (Low/Actual/High)	3,00 <= 6,00	<= 6,00	
MAA-99mTc Lot # K-20211129-048		29/11/2021 12:59	VS	Butanona		3MM Whatman		370		8	0	97,884 %	Pass	
							Turvo/Incolor				pH (Low/Actual/High)	3,00 <= 5,00	<= 6,00	
MAA-99mTc Lot # K-20211129-049		29/11/2021 13:01	VS	Butanona		3MM Whatman		537		5	0	99,077 %	Pass	
							Turvo/Incolor				pH (Low/Actual/High)	3,00 <= 6,00	<= 6,00	
MAA-99mTc Lot # K-20211129-050		29/11/2021 13:04	VS	Butanona		3MM Whatman		154		2	0	98,718 %	Pass	
							Turvo/Incolor				pH (Low/Actual/High)	3,00 <= 5,00	<= 6,00	
MAA-99mTc Lot # K-20211129-051		29/11/2021 13:06	VS	Butanona		3MM Whatman		325		4	0	98,784 %	Pass	
							Turvo/Incolor				pH (Low/Actual/High)	3,00 <= 6,00	<= 6,00	
MAA-99mTc Lot # K-20211129-052		29/11/2021 13:19	VS	Butanona		3MM Whatman		218		3	0	98,643 %	Pass	
							Turvo/Incolor				pH (Low/Actual/High)	3,00 <= 6,00	<= 6,00	
MDP-99mTc Lot # K-20211129-008		29/11/2021 03:39	MB	Acetona		3MM Whatman		779		3	0	99,014 %	Pass	
				Normal Saline		3MM Whatman		5		825	pH (Low/Actual/High)	4,00 <= 5,00	<= 7,80	
							Límpido/Incolor							

Product:	Manuf:	Calibration D/T	Solution	Media	Origin	Middle	Top	Solution 2	Media 2	Origin 2	Top 2	% Tag	QC By	Pass Fail
MDP-99mTc Lot # K-20211129-009		29/11/2021 03:42	MB	Acetona Normal Saline	3MM Whatman 3MM Whatman			627 1		1 813	0	pH (Low/Actual/High)	99,718 %	Pass 4,00 <= 5,00 <= 7,80
Límpido/Incolor														
MDP-99mTc Lot # K-20211129-010		29/11/2021 04:01	MB	Acetona Normal Saline	3MM Whatman 3MM Whatman			975 3		5 1066	0	pH (Low/Actual/High)	99,209 %	Pass 4,00 <= 5,00 <= 7,80
Límpido/Incolor														
MDP-99mTc Lot # K-20211129-026		29/11/2021 07:38	PM	Acetona Normal Saline	3MM Whatman 3MM Whatman			629 20		12 739	0	pH (Low/Actual/High)	95,493 %	Pass 4,00 <= 6,00 <= 7,80
Límpido/Incolor														
MDP-99mTc Lot # K-20211129-031		29/11/2021 08:13	PM	Acetona Normal Saline	3MM Whatman 3MM Whatman			903 13		3 835	0	pH (Low/Actual/High)	98,136 %	Pass 4,00 <= 5,00 <= 7,80
Límpido/Incolor														
MDP-99mTc Lot # K-20211129-032		29/11/2021 08:13	PM	Acetona Normal Saline	3MM Whatman 3MM Whatman			1267 14		10 1111	0	pH (Low/Actual/High)	97,972 %	Pass 4,00 <= 5,00 <= 7,80
Límpido/Incolor														
MDP-99mTc Lot # K-20211129-036		29/11/2021 10:01	PM	Acetona Normal Saline	3MM Whatman 3MM Whatman			1505 11		2 1431	0	pH (Low/Actual/High)	99,104 %	Pass 4,00 <= 5,00 <= 7,80
Límpido/Incolor														
MDP-99mTc Lot # K-20211129-040		29/11/2021 10:59	VS	Acetona Normal Saline	3MM Whatman 3MM Whatman			220 3		9 1052	0	pH (Low/Actual/High)	95,786 %	Pass 4,00 <= 7,00 <= 7,80
Límpido/Incolor														
MIBI-99mTc Lot # K-20211129-001		29/11/2021 02:11	MB	Normal Saline Metanol	Sílica gel 3MM Whatman			530 9		7 757	0	pH (Low/Actual/High)	97,522 %	Pass 5,00 <= 5,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20211129-002		29/11/2021 02:15	MB	Normal Saline Metanol	Sílica gel 3MM Whatman			961 17		15 1283	0	pH (Low/Actual/High)	97,155 %	Pass 5,00 <= 5,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20211129-003		29/11/2021 02:21	MB	Normal Saline Metanol	Sílica gel 3MM Whatman			478 48		5 1040	0	pH (Low/Actual/High)	94,553 %	Pass 5,00 <= 5,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20211129-004		29/11/2021 02:41	MB	Normal Saline Metanol	Sílica gel 3MM Whatman			1050 15		16 1464	0	pH (Low/Actual/High)	97,485 %	Pass 5,00 <= 5,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20211129-005		29/11/2021 02:42	MB	Normal Saline Metanol	Sílica gel 3MM Whatman			1130 13		22 1943	0	pH (Low/Actual/High)	97,426 %	Pass 5,00 <= 5,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20211129-006		29/11/2021 03:26	MB	Normal Saline Metanol	Sílica gel 3MM Whatman			895 10		11 701	0	pH (Low/Actual/High)	97,379 %	Pass 5,00 <= 5,00 <= 6,00
Límpido/Incolor														

Product:	Manuf:	Calibration D/T	Solution	Media	Origin	Middle	Top	Solution 2	Media 2	Origin 2	Top 2	% Tag	QC By	Pass Fail
MIBI-99mTc Lot # K-20211129-007	29/11/2021	03:29	MB	Normal Saline Metanol	Sílica gel 3MM Whatman			1228 38		14 1563	0	pH (Low/Actual/High)	96,499 %	Pass 5,00 <= 5,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20211129-016	29/11/2021	04:19	MB	Normal Saline Metanol	Sílica gel 3MM Whatman			451 4		4 100	0	pH (Low/Actual/High)	95,275 %	Pass 5,00 <= 5,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20211129-017	29/11/2021	06:42	PM	Normal Saline Metanol	Sílica gel 3MM Whatman			204 29		3 996	0	pH (Low/Actual/High)	95,721 %	Pass 5,00 <= 6,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20211129-018	29/11/2021	06:45	PM	Normal Saline Metanol	Sílica gel 3MM Whatman			1059 25		13 1631	0	pH (Low/Actual/High)	97,278 %	Pass 5,00 <= 5,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20211129-020	29/11/2021	07:08	PM	Normal Saline Metanol	Sílica gel 3MM Whatman			1135 3		5 230	0	pH (Low/Actual/High)	98,274 %	Pass 5,00 <= 5,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20211129-021	29/11/2021	07:09	PM	Normal Saline Metanol	Sílica gel 3MM Whatman			895 43		7 1364	0	pH (Low/Actual/High)	96,168 %	Pass 5,00 <= 6,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20211129-022	29/11/2021	07:22	JS	Normal Saline Metanol	Sílica gel 3MM Whatman			377 1		6 329	0	pH (Low/Actual/High)	98,13 %	Pass 5,00 <= 5,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20211129-023	29/11/2021	07:23	JS	Normal Saline Metanol	Sílica gel 3MM Whatman			385 6		5 526	0	pH (Low/Actual/High)	97,59 %	Pass 5,00 <= 5,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20211129-027	29/11/2021	07:51	PM	Normal Saline Metanol	Sílica gel 3MM Whatman			1058 5		12 250	0	pH (Low/Actual/High)	96,918 %	Pass 5,00 <= 6,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20211129-035	29/11/2021	09:57	MB	Normal Saline Metanol	Sílica gel 3MM Whatman			1848 5		20 1965	0	pH (Low/Actual/High)	98,676 %	Pass 5,00 <= 5,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20211129-042	29/11/2021	11:11	YD	Normal Saline Metanol	Sílica gel 3MM Whatman			2829 17		59 1971	0	pH (Low/Actual/High)	97,102 %	Pass 5,00 <= 6,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20211129-043	29/11/2021	11:21	YD	Normal Saline Metanol	Sílica gel 3MM Whatman			1829 2		41 1152	0	pH (Low/Actual/High)	97,634 %	Pass 5,00 <= 6,00 <= 6,00
Límpido/Incolor														
PSMA-68Ga Lot # K-20211129-019	29/11/2021	06:48	PM	Acetato de amônio 1M:Metanol	Sílica gel			0,1		28,3	0	pH (Low/Actual/High)	99,648 %	Pass 4,00 <= 5,00 <= 8,00
Límpido e Incolor														

Product:	Manuf:	Calibration D/T	Solution	Media	Origin	Middle	Top	Solution 2	Media 2	Origin 2	Top 2	% Tag	QC By	Pass Fail
PSMA-68Ga Lot # K-20211129-047	29/11/2021	12:52	VS	Acetato de amônio 1M:Metanol	Sílica gel			5		1039	0		99,521 %	Pass
												pH (Low/Actual/High)	4,00 <= 7,00	<= 8,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20211129-001	29/11/2021	01:56	MB	Butanona		3MM Whatman		52		2960	0		98,274 %	Pass
												pH (Low/Actual/High)	5,00 <= 5,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20211129-002	29/11/2021	02:08	MB	Butanona		3MM Whatman		53		3432	0		98,479 %	Pass
												pH (Low/Actual/High)	5,00 <= 5,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20211129-003	29/11/2021	02:17	MB	Butanona		3MM Whatman		0,2		48,6	0		99,59 %	Pass
												pH (Low/Actual/High)	5,00 <= 5,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20211129-004	29/11/2021	02:36	MB	Butanona		3MM Whatman		2		942	0		99,788 %	Pass
												pH (Low/Actual/High)	5,00 <= 5,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20211129-005	29/11/2021	03:24	MB	Butanona		3MM Whatman		8		204	0		96,226 %	Pass
												pH (Low/Actual/High)	5,00 <= 5,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20211129-006	29/11/2021	03:27	MB	Butanona		3MM Whatman		0,3		12,3	0		97,619 %	Pass
												pH (Low/Actual/High)	5,00 <= 5,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20211129-007	29/11/2021	03:37	MB	Butanona		3MM Whatman		1		111	0		99,107 %	Pass
												pH (Low/Actual/High)	5,00 <= 5,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20211129-008	29/11/2021	03:58	MB	Butanona		3MM Whatman		4		556	0		99,286 %	Pass
												pH (Low/Actual/High)	5,00 <= 5,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20211129-009	29/11/2021	04:11	MB	Butanona		3MM Whatman		2		365	0		99,455 %	Pass
												pH (Low/Actual/High)	5,00 <= 5,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20211129-010	29/11/2021	06:35	PM	Butanona		3MM Whatman		11		1453	0		99,249 %	Pass
												pH (Low/Actual/High)	5,00 <= 5,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20211129-012	29/11/2021	07:05	GS	Butanona		3MM Whatman		3		575	0		99,481 %	Pass
												pH (Low/Actual/High)	5,00 <= 5,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20211129-013	29/11/2021	07:18	PM	Butanona		3MM Whatman		1		819	0		99,878 %	Pass
												pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
														Límpido e Incolor

Product:	Manuf:	Calibration D/T	Solution	Media	Origin	Middle	Top	Solution 2	Media 2	Origin 2	Top 2	% Tag	QC By	Pass Fail
Tecnécio-99m Lot # E-20211129-014		29/11/2021 07:45 PM	Butanona		3MM	Whatman		3		300	0		99,01 %	Pass
							Límpido e Incolor				pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00	
Tecnécio-99m Lot # E-20211129-015		29/11/2021 08:09 PM	Butanona		3MM	Whatman		2		776	0		99,743 %	Pass
							Límpido e Incolor				pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00	
Tecnécio-99m Lot # E-20211129-017		29/11/2021 09:35 PM	Butanona		3MM	Whatman		1		422	0		99,764 %	Pass
							Límpido e Incolor				pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00	
Tecnécio-99m Lot # E-20211129-018		29/11/2021 10:41 PM	Butanona		3MM	Whatman		3		348	0		99,145 %	Pass
							Límpido e Incolor				pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00	
Tecnécio-99m Lot # E-20211129-019		29/11/2021 11:07 VS	Butanona		3MM	Whatman		8		651	0		98,786 %	Pass
							Límpido e Incolor				pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00	
Tecnécio-99m Lot # E-20211129-020		29/11/2021 11:18 VS	Butanona		3MM	Whatman		2		249	0		99,203 %	Pass
							Límpido e Incolor				pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00	