



# Kit Quality Control Report

## RPH CENTRAL PHARMA

CNPJ: 19.315.658/0001-10  
 R: Maestro Cardim, 769 - São Paulo/SP  
 Farm. RT: Priscila Larcher CRF/SP-74.417

Date: From 06/11/2021 To 06/11/2021

Product:	Manuf:	Calibration	D/T	QC By	Solution	Media	Origin	Middle	Top	Background	Particle Size/ OK?	% Tag	Pass Fail
DMSA-99mTc Lot # K-20211106-017		06/11/2021	04:30	MB	Butanona	3MM Whatman	174,8		4	0		95,199 %	Pass
					Normal Saline	Silica gel	3	114	pH (Low/Actual/High)	2,00 <= 3,00	<= 3,00	Límpido/Incolor	
DTPA-99mTc Lot # K-20211106-010	RPH	06/11/2021	03:26	MB	Acetona	3MM Whatman	277		4	0		97,478 %	Pass
					Normal Saline	3MM Whatman	3	270	pH (Low/Actual/High)	3,80 <= 5,00	<= 7,50	Límpido/Incolor	
FITATO-99mTc Lot # K-20211106-011		06/11/2021	03:38	MB	Butanona	3MM Whatman	351		10	0		97,23 %	Pass
									pH (Low/Actual/High)	5,00 <= 6,00	<= 7,00	Límpido/Incolor	
FITATO-99mTc Lot # K-20211106-012		06/11/2021	03:39	MB	Butanona	3MM Whatman	451		1	0		99,779 %	Pass
									pH (Low/Actual/High)	5,00 <= 6,00	<= 7,00	Límpido/Incolor	
FITATO-99mTc Lot # K-20211106-020		06/11/2021	06:06	MB	Butanona	3MM Whatman	415		4	0		99,045 %	Pass
									pH (Low/Actual/High)	5,00 <= 6,00	<= 7,00	Límpido/Incolor	
MAA-99mTc Lot # K-20211106-016		06/11/2021	04:29	MB	Butanona	3MM Whatman	314		4	0		98,742 %	Pass
									pH (Low/Actual/High)	3,00 <= 5,00	<= 6,00	Turvo/Incolor	
MDP-99mTc Lot # K-20211106-005		06/11/2021	02:25	MB	Acetona	3MM Whatman	964		7	0		98,303 %	Pass
					Normal Saline	3MM Whatman	2	203	pH (Low/Actual/High)	4,00 <= 5,00	<= 7,80	Límpido/Incolor	
MDP-99mTc Lot # K-20211106-006		06/11/2021	02:44	MB	Acetona	3MM Whatman	1101		0	0		99,031 %	Pass
					Normal Saline	3MM Whatman	9	920	pH (Low/Actual/High)	4,00 <= 5,00	<= 7,80	Límpido/Incolor	
MDP-99mTc Lot # K-20211106-007		06/11/2021	02:48	MB	Acetona	3MM Whatman	641		8	0		98,223 %	Pass
					Normal Saline	3MM Whatman	3	548	pH (Low/Actual/High)	4,00 <= 5,00	<= 7,80	Límpido/Incolor	
MDP-99mTc Lot # K-20211106-009		06/11/2021	03:22	MB	Acetona	3MM Whatman	907		2	0		99,244 %	Pass
					Normal Saline	3MM Whatman	6	1114	pH (Low/Actual/High)	4,00 <= 5,00	<= 7,80	Límpido/Incolor	
MDP-99mTc Lot # K-20211106-014		06/11/2021	03:54	MB	Acetona	3MM Whatman	1039		2	0		98,951 %	Pass
					Normal Saline	3MM Whatman	10	1157	pH (Low/Actual/High)	4,00 <= 5,00	<= 7,80	Límpido/Incolor	

Product:	Manuf:	Calibration D/T	Solution	Media	Origin	Middle	Top	Solution 2	Media 2	Origin 2	Top 2	% Tag	QC By	Pass Fail
MIBI-99mTc Lot # K-20211106-002		06/11/2021 02:07	MB	Normal Saline Metanol	Sílica gel 3MM Whatman			833 18		37 1350	0	pH (Low/Actual/High)	94,431 %	Pass 5,00 <= 5,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20211106-001		06/11/2021 02:07	MB	Normal Saline Metanol	Sílica gel 3MM Whatman			1933 25		47 1912	0	pH (Low/Actual/High)	96,336 %	Pass 5,00 <= 5,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20211106-003		06/11/2021 02:14	MB	Normal Saline Metanol	Sílica gel 3MM Whatman			218 4		4 257	0	pH (Low/Actual/High)	96,666 %	Pass 5,00 <= 5,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20211106-004		06/11/2021 02:15	MB	Normal Saline Metanol	Sílica gel 3MM Whatman			211 4		5 218	0	pH (Low/Actual/High)	95,883 %	Pass 5,00 <= 5,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20211106-008		06/11/2021 03:10	MB	Normal Saline Metanol	Sílica gel 3MM Whatman			1121 15		32 1282	0	pH (Low/Actual/High)	96,068 %	Pass 5,00 <= 5,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20211106-013		06/11/2021 03:52	MB	Normal Saline Metanol	Sílica gel 3MM Whatman			1040 4		28 277	0	pH (Low/Actual/High)	95,955 %	Pass 5,00 <= 5,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20211106-015		06/11/2021 04:22	MB	Normal Saline Metanol	Sílica gel 3MM Whatman			988 34		8 1260	0	pH (Low/Actual/High)	96,569 %	Pass 5,00 <= 5,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20211106-018		06/11/2021 05:15	MB	Normal Saline Metanol	Sílica gel 3MM Whatman			1027 8		27 1608	0	pH (Low/Actual/High)	96,943 %	Pass 5,00 <= 5,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20211106-019		06/11/2021 05:20	MB	Normal Saline Metanol	Sílica gel 3MM Whatman			1845 53		21 1898	0	pH (Low/Actual/High)	96,158 %	Pass 5,00 <= 5,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20211106-021		06/11/2021 06:10	MB	Normal Saline Metanol	Sílica gel 3MM Whatman			317 3		12 250	0	pH (Low/Actual/High)	95,167 %	Pass 5,00 <= 5,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20211106-022		06/11/2021 09:28	MB	Normal Saline Metanol	Sílica gel 3MM Whatman			390 13		10 542	0	pH (Low/Actual/High)	95,158 %	Pass 5,00 <= 5,00 <= 6,00
Límpido/Incolor														
Tecnécio-99m Lot # E-20211106-001		06/11/2021 02:00	MB	Butanona	3MM Whatman			1		502	0	pH (Low/Actual/High)	99,801 %	Pass 5,00 <= 6,00 <= 6,00
Límpido e Incolor														
Tecnécio-99m Lot # E-20211106-002		06/11/2021 02:10	MB	Butanona	3MM Whatman			14		531	0	pH (Low/Actual/High)	97,431 %	Pass 5,00 <= 5,00 <= 6,00
Límpido e Incolor														

Product:	Manuf:	Calibration D/T	Solution	Media	Origin	Middle	Top	Solution 2	Media 2	Origin 2	Top 2	% Tag	QC By	Pass Fail
Tecnécio-99m Lot # E-20211106-003		06/11/2021 02:42	MB	Butanona		3MM Whatman		4		215	0	98,174 %	Pass	
												pH (Low/Actual/High)	5,00 <= 5,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20211106-004		06/11/2021 03:08	MB	Butanona		3MM Whatman		3		1734	0	99,827 %	Pass	
												pH (Low/Actual/High)	5,00 <= 5,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20211106-005		06/11/2021 03:19	MB	Butanona		3MM Whatman		17		1676	0	98,996 %	Pass	
												pH (Low/Actual/High)	5,00 <= 5,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20211106-006		06/11/2021 03:34	MB	Butanona		3MM Whatman		1		514	0	99,806 %	Pass	
												pH (Low/Actual/High)	5,00 <= 5,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20211106-007		06/11/2021 04:26	MB	Butanona		3MM Whatman		14		1176	0	98,824 %	Pass	
												pH (Low/Actual/High)	5,00 <= 5,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20211106-008		06/11/2021 04:57	MB	Butanona		3MM Whatman		2		672	0	99,703 %	Pass	
												pH (Low/Actual/High)	5,00 <= 5,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20211106-009		06/11/2021 05:12	MB	Butanona		3MM Whatman		2		206	0	99,038 %	Pass	
												pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20211106-010		06/11/2021 05:16	MB	Butanona		3MM Whatman		2		234	0	99,153 %	Pass	
												pH (Low/Actual/High)	5,00 <= 5,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20211106-011		06/11/2021 09:25	MB	Butanona		3MM Whatman		10		792	0	98,753 %	Pass	
												pH (Low/Actual/High)	5,00 <= 5,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20211106-013		06/11/2021 15:09	MP	Butanona		3MM Whatman		2		591	0	99,663 %	Pass	
												pH (Low/Actual/High)	5,00 <= 5,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20211106-012		06/11/2021 15:09	MP	Butanona		3MM Whatman		143		9580	0	98,529 %	Pass	
												pH (Low/Actual/High)	5,00 <= 5,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20211106-014		06/11/2021 15:11	MP	Butanona		3MM Whatman		22		1188	0	98,182 %	Pass	
												pH (Low/Actual/High)	5,00 <= 5,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20211106-015		06/11/2021 15:14	MP	Butanona		3MM Whatman		123		2415	0	95,154 %	Pass	
												pH (Low/Actual/High)	5,00 <= 5,00	<= 6,00
														Límpido e Incolor

Product:	Manuf:	Calibration D/T	Solution	Media	Origin	Middle	Top	Solution 2	Media 2	Origin 2	Top 2	% Tag	QC By	Pass Fail
Tecnécio-99m Lot # E-20211106-016		06/11/2021 15:18	MP	Butanona		3MM Whatman		84		8265	0		98,994 %	Pass
											pH (Low/Actual/High)	5,00 <= 5,00 <= 6,00		
												Límpido e Incolor		
Tecnécio-99m Lot # E-20211106-017		06/11/2021 15:20	MP	Butanona		3MM Whatman		97		9415	0		98,98 %	Pass
											pH (Low/Actual/High)	5,00 <= 5,00 <= 6,00		
												Límpido e Incolor		