



Kit Quality Control Report

RPH CENTRAL PHARMA

CNPJ: 19.315.658/0001-10

R: Maestro Cardim, 769 - São Paulo/SP

Farm. RT: Priscila Larcher CRF/SP-74.417

Date: From 20/12/2021 To 20/12/2021

Product:	Manuf:	Calibration	D/T	QC By	Solution	Media	Origin	Middle	Top	Background	Particle Size/ OK?	% Tag	Pass Fail
DMSA-99mTc		20/12/2021	07:36	MB	Butanona	3MM Whatman	142,3		2,6	0		97,03 %	Pass
Lot # K-20211220-023					Normal Saline	Silica gel	1,1		92,5	pH (Low/Actual/High)		2,00 <= 3,00	<= 3,00
						Límpido/Incolor							
DMSA-99mTc		20/12/2021	08:03	MB	Butanona	3MM Whatman	147,8		2,2	0		93,858 %	Pass
Lot # K-20211220-028					Normal Saline	Silica gel	4,1		83,6	pH (Low/Actual/High)		2,00 <= 3,00	<= 3,00
						Límpido/Incolor							
DMSA-99mTc		20/12/2021	10:14	GS	Butanona	3MM Whatman	87,3		0,5	0		93,951 %	Pass
Lot # K-20211220-038					Normal Saline	Silica gel	4		69	pH (Low/Actual/High)		2,00 <= 3,00	<= 3,00
						Límpido/Incolor							
DMSA-99mTc		20/12/2021	10:15	GS	Butanona	3MM Whatman	127,7		1,8	0		94,038 %	Pass
Lot # K-20211220-039					Normal Saline	Silica gel	4,7		98,1	pH (Low/Actual/High)		2,00 <= 3,00	<= 3,00
						Límpido/Incolor							
DTPA-99mTc	RPH	20/12/2021	03:17	MB	Acetona	3MM Whatman	627		3	0		97,603 %	Pass
Lot # K-20211220-009					Normal Saline	3MM Whatman	15		766	pH (Low/Actual/High)		3,80 <= 5,00	<= 7,50
						Límpido/Incolor							
DTPA-99mTc	RPH	20/12/2021	03:21	MB	Acetona	3MM Whatman	900		6	0		98,305 %	Pass
Lot # K-20211220-011					Normal Saline	3MM Whatman	6		575	pH (Low/Actual/High)		3,80 <= 5,00	<= 7,50
						Límpido/Incolor							
DTPA-99mTc	RPH	20/12/2021	07:59	MB	Acetona	3MM Whatman	567		6	0		97,131 %	Pass
Lot # K-20211220-027					Normal Saline	3MM Whatman	9		485	pH (Low/Actual/High)		3,80 <= 5,00	<= 7,50
						Límpido/Incolor							
DTPA-99mTc	RPH	20/12/2021	09:23	GS	Acetona	3MM Whatman	272		1	0		95,088 %	Pass
Lot # K-20211220-032					Normal Saline	3MM Whatman	14		294	pH (Low/Actual/High)		3,80 <= 6,00	<= 7,50
						Límpido/Incolor							
FITATO-99mTc		20/12/2021	03:19	MB	Butanona	3MM Whatman	102,6		2,6	0		97,529 %	Pass
Lot # K-20211220-010										pH (Low/Actual/High)		5,00 <= 7,00	<= 7,00
						Límpido/Incolor							
FITATO-99mTc		20/12/2021	09:25	GS	Butanona	3MM Whatman	288		1	0		99,654 %	Pass
Lot # K-20211220-033										pH (Low/Actual/High)		5,00 <= 5,00	<= 7,00
						Límpido/Incolor							
MAA-99mTc		20/12/2021	04:17	MB	Butanona	3MM Whatman	462		9	0		98,089 %	Pass
Lot # K-20211220-015										pH (Low/Actual/High)		3,00 <= 5,00	<= 6,00
						Turvo/Incolor							

Product:	Manuf:	Calibration D/T	Solution	Media	Origin	Middle	Top	Solution 2	Media 2	Origin 2	Top 2	% Tag	QC By	Pass Fail
MAA-99mTc Lot # K-20211220-025		20/12/2021 07:55	MB	Butanona		3MM Whatman		41,5		0,8	0		98,109 %	Pass
												pH (Low/Actual/High)	3,00 <= 5,00 <= 6,00	
												Turvo/Incolor		
MAA-99mTc Lot # K-20211220-026		20/12/2021 07:56	MB	Butanona		3MM Whatman		69,5		1	0		98,582 %	Pass
												pH (Low/Actual/High)	3,00 <= 5,00 <= 6,00	
												Turvo/Incolor		
MAA-99mTc Lot # K-20211220-034		20/12/2021 09:27	GS	Butanona		3MM Whatman		304		7	0		97,749 %	Pass
												pH (Low/Actual/High)	3,00 <= 5,00 <= 6,00	
												Turvo/Incolor		
MAA-99mTc Lot # K-20211220-044		20/12/2021 11:39	JS	Butanona		3MM Whatman		116		1,9	0		98,388 %	Pass
												pH (Low/Actual/High)	3,00 <= 6,00 <= 6,00	
												Turvo/Incolor		
MAA-99mTc Lot # K-20211220-045		20/12/2021 11:41	MB	Butanona		3MM Whatman		189		0,9	0		99,526 %	Pass
												pH (Low/Actual/High)	3,00 <= 5,00 <= 6,00	
												Turvo/Incolor		
MAA-99mTc Lot # K-20211220-046		20/12/2021 11:42	MB	Butanona		3MM Whatman		181,1		2,9	0		98,424 %	Pass
												pH (Low/Actual/High)	3,00 <= 5,00 <= 6,00	
												Turvo/Incolor		
MAA-99mTc Lot # K-20211220-047		20/12/2021 11:46	JS	Butanona		3MM Whatman		227		7	0		97,009 %	Pass
												pH (Low/Actual/High)	3,00 <= 5,00 <= 6,00	
												Turvo/Incolor		
MDP-99mTc Lot # K-20211220-008		20/12/2021 03:06	MB	Acetona		3MM Whatman		1055		3	0		98,519 %	Pass
				Normal Saline		3MM Whatman		15		1238		pH (Low/Actual/High)	4,00 <= 5,00 <= 7,80	
												Límpido/Incolor		
MDP-99mTc Lot # K-20211220-014		20/12/2021 03:56	MB	Acetona		3MM Whatman		1648		1	0		98,78 %	Pass
				Normal Saline		3MM Whatman		24		2046		pH (Low/Actual/High)	4,00 <= 5,00 <= 7,80	
												Límpido/Incolor		
MDP-99mTc Lot # K-20211220-016		20/12/2021 04:18	MB	Acetona		3MM Whatman		1331		2	0		99,046 %	Pass
				Normal Saline		3MM Whatman		12		1481		pH (Low/Actual/High)	4,00 <= 5,00 <= 7,80	
												Límpido/Incolor		
MDP-99mTc Lot # K-20211220-020		20/12/2021 07:10	MB	Acetona		3MM Whatman		932,2		1	0		98,603 %	Pass
				Normal Saline		3MM Whatman		12,8		979,4		pH (Low/Actual/High)	4,00 <= 6,00 <= 7,80	
												Límpido/Incolor		
MDP-99mTc Lot # K-20211220-021		20/12/2021 07:12	MB	Acetona		3MM Whatman		636		3	0		98,742 %	Pass
				Normal Saline		3MM Whatman		4		503		pH (Low/Actual/High)	4,00 <= 5,00 <= 7,80	
												Límpido/Incolor		
MDP-99mTc Lot # K-20211220-029		20/12/2021 08:11	MB	Acetona		3MM Whatman		536		1	0		98,613 %	Pass
				Normal Saline		3MM Whatman		7		576		pH (Low/Actual/High)	4,00 <= 5,00 <= 7,80	
												Límpido/Incolor		

Product:	Manuf:	Calibration D/T	Solution	Media	Origin	Middle	Top	Solution 2	Media 2	Origin 2	Top 2	% Tag	QC By	Pass Fail
MDP-99mTc Lot # K-20211220-030		20/12/2021 08:14	MB	Acetona Normal Saline	3MM Whatman 3MM Whatman			811 17		4 967	0 pH (Low/Actual/High)	97,782 %	Pass	4,00 <= 5,00 <= 7,80
Límpido/Incolor														
MDP-99mTc Lot # K-20211220-036		20/12/2021 09:33	GS	Acetona Normal Saline	3MM Whatman 3MM Whatman			532 14		10 910	0 pH (Low/Actual/High)	96,64 %	Pass	4,00 <= 6,00 <= 7,80
Límpido/Incolor														
MDP-99mTc Lot # K-20211220-040		20/12/2021 10:45	YD	Acetona Normal Saline	3MM Whatman 3MM Whatman			773 32		2 794	0 pH (Low/Actual/High)	95,868 %	Pass	4,00 <= 6,00 <= 7,80
Límpido/Incolor														
MDP-99mTc Lot # K-20211220-042		20/12/2021 11:29	JS	Acetona Normal Saline	3MM Whatman 3MM Whatman			620 3		6 363	0 pH (Low/Actual/High)	98,222 %	Pass	4,00 <= 5,00 <= 7,80
Límpido/Incolor														
MIBI-99mTc Lot # K-20211220-001		20/12/2021 01:44	MB	Normal Saline Metanol	Sílica gel 3MM Whatman			985,4 8		27,6 1683	0 pH (Low/Actual/High)	96,802 %	Pass	5,00 <= 5,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20211220-002		20/12/2021 01:47	MB	Normal Saline Metanol	Sílica gel 3MM Whatman			755,3 9		15,3 1425	0 pH (Low/Actual/High)	97,387 %	Pass	5,00 <= 5,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20211220-006		20/12/2021 01:50	MB	Normal Saline Metanol	Sílica gel 3MM Whatman			1250 31		18 1707	0 pH (Low/Actual/High)	96,797 %	Pass	5,00 <= 5,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20211220-003		20/12/2021 01:51	MB	Normal Saline Metanol	Sílica gel 3MM Whatman			1119 4		27 1820	0 pH (Low/Actual/High)	97,425 %	Pass	5,00 <= 5,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20211220-005		20/12/2021 01:53	MB	Normal Saline Metanol	Sílica gel 3MM Whatman			855,9 9		23,2 1284	0 pH (Low/Actual/High)	96,665 %	Pass	5,00 <= 5,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20211220-007		20/12/2021 01:54	MB	Normal Saline Metanol	Sílica gel 3MM Whatman			1070 8		28 1676	0 pH (Low/Actual/High)	96,975 %	Pass	5,00 <= 5,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20211220-012		20/12/2021 03:35	MB	Normal Saline Metanol	Sílica gel 3MM Whatman			1093 16		14 1128	0 pH (Low/Actual/High)	97,337 %	Pass	5,00 <= 5,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20211220-013		20/12/2021 03:43	MB	Normal Saline Metanol	Sílica gel 3MM Whatman			1982 6		29 1048	0 pH (Low/Actual/High)	97,989 %	Pass	5,00 <= 5,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20211220-017		20/12/2021 06:44	MB	Normal Saline Metanol	Sílica gel 3MM Whatman			1319 11		20 2046	0 pH (Low/Actual/High)	97,972 %	Pass	5,00 <= 5,00 <= 6,00
Límpido/Incolor														

Product:	Manuf:	Calibration D/T	Solution	Media	Origin	Middle	Top	Solution 2	Media 2	Origin 2	Top 2	% Tag	QC By	Pass Fail
MIBI-99mTc Lot # K-20211220-018		20/12/2021 06:50	MB	Normal Saline Metanol	Sílica gel 3MM Whatman			1888 3		108 1892	0	pH (Low/Actual/High)	94,431 %	Pass 5,00 <= 5,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20211220-024		20/12/2021 07:47	MB	Normal Saline Metanol	Sílica gel 3MM Whatman			1386 20		22 1628	0	pH (Low/Actual/High)	97,224 %	Pass 5,00 <= 5,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20211220-035		20/12/2021 09:31	GS	Normal Saline Metanol	Sílica gel 3MM Whatman			402 4		22 270	0	pH (Low/Actual/High)	93,351 %	Pass 5,00 <= 5,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20211220-037		20/12/2021 10:07	GS	Normal Saline Metanol	Sílica gel 3MM Whatman			355 1		7 443	0	pH (Low/Actual/High)	97,841 %	Pass 5,00 <= 6,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20211220-041		20/12/2021 11:13	JS	Normal Saline Metanol	Sílica gel 3MM Whatman			183 6,2		3,6 181,3	0	pH (Low/Actual/High)	94,764 %	Pass 5,00 <= 5,00 <= 6,00
Límpido/Incolor														
PIRO-99mTc Lot # K-20211220-022		20/12/2021 07:13	MB	Butanona Normal Saline	3MM Whatman 3MM Whatman			404,5 3,2		1 329,9	0	pH (Low/Actual/High)	98,793 %	Pass 5,00 <= 6,00 <= 7,00
Límpido/Incolor														
PSMA-68Ga Lot # K-20211220-019		20/12/2021 07:00	MB	Acetato de amônio 1M:Metanol	Sílica gel			0,7		29,15	0	pH (Low/Actual/High)	97,655 %	Pass 4,00 <= 5,00 <= 8,00
Límpido e Incolor														
PSMA-68Ga Lot # K-20211220-048		20/12/2021 12:43	YD	Acetato de amônio 1M:Metanol	Sílica gel			0,4		27,8	0	pH (Low/Actual/High)	98,582 %	Pass 4,00 <= 6,00 <= 8,00
Límpido e Incolor														
Tecnécio-99m Lot # E-20211220-001		20/12/2021 01:16	BB	Butanona	3MM Whatman			40		6591	0	pH (Low/Actual/High)	99,397 %	Pass 5,00 <= 6,00 <= 6,00
Límpido e Incolor														
Tecnécio-99m Lot # E-20211220-002		20/12/2021 01:19	BB	Butanona	3MM Whatman			6		4858	0	pH (Low/Actual/High)	99,877 %	Pass 5,00 <= 6,00 <= 6,00
Límpido e Incolor														
Tecnécio-99m Lot # E-20211220-003		20/12/2021 01:20	BB	Butanona	3MM Whatman			51		8455	0	pH (Low/Actual/High)	99,40 %	Pass 5,00 <= 6,00 <= 6,00
Límpido e Incolor														
Tecnécio-99m Lot # E-20211220-004		20/12/2021 01:23	BB	Butanona	3MM Whatman			34		10470	0	pH (Low/Actual/High)	99,676 %	Pass 5,00 <= 6,00 <= 6,00
Límpido e Incolor														
Tecnécio-99m Lot # E-20211220-005		20/12/2021 01:30	BB	Butanona	3MM Whatman			21		6654	0	pH (Low/Actual/High)	99,685 %	Pass 5,00 <= 6,00 <= 6,00
Límpido e Incolor														

Product:	Manuf:	Calibration D/T	Solution	Media	Origin	Middle	Top	Solution 2	Media 2	Origin 2	Top 2	% Tag	QC By	Pass Fail
Tecnécio-99m Lot # E-20211220-006		20/12/2021 01:33	BB	Butanona		3MM Whatman		9		7708	0		99,883 %	Pass
												pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20211220-007		20/12/2021 01:39	MB	Butanona		3MM Whatman		7		5925	0		99,882 %	Pass
												pH (Low/Actual/High)	5,00 <= 5,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20211220-008		20/12/2021 01:41	MB	Butanona		3MM Whatman		26		5228	0		99,505 %	Pass
												pH (Low/Actual/High)	5,00 <= 5,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20211220-009		20/12/2021 01:49	MB	Butanona		3MM Whatman		1,8		207,6	0		99,14 %	Pass
												pH (Low/Actual/High)	5,00 <= 5,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20211220-010		20/12/2021 03:01	MB	Butanona		3MM Whatman		3		1378	0		99,783 %	Pass
												pH (Low/Actual/High)	5,00 <= 5,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20211220-011		20/12/2021 03:02	MB	Butanona		3MM Whatman		3		1093	0		99,726 %	Pass
												pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20211220-012		20/12/2021 03:13	MB	Butanona		3MM Whatman		11,7		751,1	0		98,466 %	Pass
												pH (Low/Actual/High)	5,00 <= 5,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20211220-013		20/12/2021 03:32	MB	Butanona		3MM Whatman		2		433	0		99,54 %	Pass
												pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20211220-014		20/12/2021 03:38	MB	Butanona		3MM Whatman		2		205	0		99,034 %	Pass
												pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20211220-015		20/12/2021 03:55	MB	Butanona		3MM Whatman		5		1150	0		99,567 %	Pass
												pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20211220-016		20/12/2021 04:12	MB	Butanona		3MM Whatman		12		788	0		98,50 %	Pass
												pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20211220-017		20/12/2021 06:38	MB	Butanona		3MM Whatman		43		2566	0		98,352 %	Pass
												pH (Low/Actual/High)	5,00 <= 5,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20211220-018		20/12/2021 06:39	MB	Butanona		3MM Whatman		4,1		458,3	0		99,113 %	Pass
												pH (Low/Actual/High)	5,00 <= 5,00	<= 6,00
														Límpido e Incolor

Product:	Manuf:	Calibration D/T	Solution	Media	Origin	Middle	Top	Solution 2	Media 2	Origin 2	Top 2	% Tag	QC By	Pass Fail
Tecnécio-99m Lot # E-20211220-019		20/12/2021 06:47	MB	Butanona		3MM Whatman		6		1707	0		99,65 %	Pass
												pH (Low/Actual/High)	5,00 <= 5,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20211220-021		20/12/2021 07:07	MB	Butanona		3MM Whatman		1		687,5	0		99,855 %	Pass
												pH (Low/Actual/High)	5,00 <= 5,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20211220-022		20/12/2021 07:45	MB	Butanona		3MM Whatman		3		419,1	0		99,289 %	Pass
												pH (Low/Actual/High)	5,00 <= 5,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20211220-023		20/12/2021 08:08	MB	Butanona		3MM Whatman		2		1577	0		99,873 %	Pass
												pH (Low/Actual/High)	5,00 <= 5,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20211220-024		20/12/2021 09:03	GS	Butanona		3MM Whatman		4		965	0		99,587 %	Pass
												pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20211220-025		20/12/2021 09:07	GS	Butanona		3MM Whatman		22		1189	0		98,183 %	Pass
												pH (Low/Actual/High)	5,00 <= 5,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20211220-026		20/12/2021 09:28	GS	Butanona		3MM Whatman		1		355	0		99,719 %	Pass
												pH (Low/Actual/High)	5,00 <= 5,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20211220-027		20/12/2021 09:37	GS	Butanona		3MM Whatman		5		440	0		98,876 %	Pass
												pH (Low/Actual/High)	5,00 <= 5,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20211220-028		20/12/2021 09:58	GS	Butanona		3MM Whatman		3		1129	0		99,735 %	Pass
												pH (Low/Actual/High)	5,00 <= 5,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20211220-029		20/12/2021 10:24	GS	Butanona		3MM Whatman		38		3117	0		98,796 %	Pass
												pH (Low/Actual/High)	5,00 <= 5,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20211220-030		20/12/2021 10:35	GS	Butanona		3MM Whatman		4		1947	0		99,795 %	Pass
												pH (Low/Actual/High)	5,00 <= 5,00	<= 6,00
														Límpido e Incolor