



Kit Quality Control Report

RPH CENTRAL PHARMA

CNPJ: 19.315.658/0001-10
 R: Maestro Cardim, 769 - São Paulo/SP
 Farm. RT: Priscila Larcher CRF/SP-74.417

Date: From 15/01/2022 To 15/01/2022

Product:	Manuf:	Calibration	D/T	QC By	Solution	Media	Origin	Middle	Top	Background	Particle Size/ OK?	% Tag	Pass Fail
DTPA-99mTc Lot # K-20220115-013	RPH	15/01/2022	04:04	MB	Acetona	3MM Whatman	576		1	0		99,287 %	Pass
					Normal Saline	3MM Whatman	3	553	pH (Low/Actual/High)	3,80 <= 6,00 <= 7,50			
Límpido/Incolor													
FITATO-99mTc Lot # K-20220115-014		15/01/2022	04:06	MB	Butanona	3MM Whatman	638		6	0		99,068 %	Pass
									pH (Low/Actual/High)	5,00 <= 5,00 <= 7,00			
Límpido/Incolor													
MDP-99mTc Lot # K-20220115-007		15/01/2022	02:52	MB	Acetona	3MM Whatman	760		5	0		98,834 %	Pass
					Normal Saline	3MM Whatman	4	777	pH (Low/Actual/High)	4,00 <= 6,00 <= 7,80			
Límpido/Incolor													
MDP-99mTc Lot # K-20220115-008		15/01/2022	02:54	MB	Acetona	3MM Whatman	903		2	0		98,737 %	Pass
					Normal Saline	3MM Whatman	9	855	pH (Low/Actual/High)	4,00 <= 6,00 <= 7,80			
Límpido/Incolor													
MDP-99mTc Lot # K-20220115-009		15/01/2022	02:56	MB	Acetona	3MM Whatman	909		2	0		99,369 %	Pass
					Normal Saline	3MM Whatman	4	967	pH (Low/Actual/High)	4,00 <= 6,00 <= 7,80			
Límpido/Incolor													
MDP-99mTc Lot # K-20220115-010		15/01/2022	02:58	MB	Acetona	3MM Whatman	854		2	0		96,588 %	Pass
					Normal Saline	3MM Whatman	28	853	pH (Low/Actual/High)	4,00 <= 6,00 <= 7,80			
Límpido/Incolor													
MIBI-99mTc Lot # K-20220115-001		15/01/2022	02:30	WG	Normal Saline	Sílica gel	1637		8	0		97,535 %	Pass
					Metanol	3MM Whatman	45	2229	pH (Low/Actual/High)	5,00 <= 6,00 <= 6,00			
Límpido/Incolor													
MIBI-99mTc Lot # K-20220115-002		15/01/2022	02:31	WG	Normal Saline	Sílica gel	1616		12	0		97,294 %	Pass
					Metanol	3MM Whatman	33	1643	pH (Low/Actual/High)	5,00 <= 6,00 <= 6,00			
Límpido/Incolor													
MIBI-99mTc Lot # K-20220115-005		15/01/2022	02:36	MB	Normal Saline	Sílica gel	1443		26	0		95,167 %	Pass
					Metanol	3MM Whatman	68	2152	pH (Low/Actual/High)	5,00 <= 5,00 <= 6,00			
Límpido/Incolor													
MIBI-99mTc Lot # K-20220115-003		15/01/2022	02:38	WG	Normal Saline	Sílica gel	1643		5	0		99,215 %	Pass
					Metanol	3MM Whatman	11	2272	pH (Low/Actual/High)	5,00 <= 6,00 <= 6,00			
Límpido/Incolor													
MIBI-99mTc Lot # K-20220115-004		15/01/2022	02:39	WG	Normal Saline	Sílica gel	1795		6	0		98,057 %	Pass
					Metanol	3MM Whatman	25	1528	pH (Low/Actual/High)	5,00 <= 6,00 <= 6,00			
Límpido/Incolor													

Product:	Manuf:	Calibration D/T	Solution	Media	Origin	Middle	Top	Solution 2	Media 2	Origin 2	Top 2	% Tag	QC By	Pass Fail
MIBI-99mTc Lot # K-20220115-006	15/01/2022	02:40	MB	Normal Saline Metanol	Sílica gel 3MM Whatman			1428 3		15 1700	0	pH (Low/Actual/High)	98,784 %	Pass 5,00 <= 5,00 <= 6,00
														Límpido/Incolor
MIBI-99mTc Lot # K-20220115-011	15/01/2022	03:02	MB	Normal Saline Metanol	Sílica gel 3MM Whatman			1091 2		24 1461	0	pH (Low/Actual/High)	97,711 %	Pass 5,00 <= 5,00 <= 6,00
														Límpido/Incolor
MIBI-99mTc Lot # K-20220115-012	15/01/2022	03:03	MB	Normal Saline Metanol	Sílica gel 3MM Whatman			1321 19		16 2009	0	pH (Low/Actual/High)	97,866 %	Pass 5,00 <= 5,00 <= 6,00
														Límpido/Incolor
Tecnécio-99m Lot # E-20220115-001	15/01/2022	02:27	WG	Butanona	3MM Whatman			12		1733	0	pH (Low/Actual/High)	99,312 %	Pass 5,00 <= 6,00 <= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20220115-002	15/01/2022	02:34	WG	Butanona	3MM Whatman			23		1473	0	pH (Low/Actual/High)	98,463 %	Pass 5,00 <= 6,00 <= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20220115-003	15/01/2022	02:35	WG	Butanona	3MM Whatman			12		2987	0	pH (Low/Actual/High)	99,60 %	Pass 5,00 <= 6,00 <= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20220115-004	15/01/2022	02:49	WG	Butanona	3MM Whatman			23		1675	0	pH (Low/Actual/High)	98,645 %	Pass 5,00 <= 6,00 <= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20220115-005	15/01/2022	03:00	MB	Butanona	3MM Whatman			10		4022	0	pH (Low/Actual/High)	99,752 %	Pass 5,00 <= 5,00 <= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20220115-006	15/01/2022	04:02	MB	Butanona	3MM Whatman			3		898	0	pH (Low/Actual/High)	99,667 %	Pass 5,00 <= 5,00 <= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20220115-007	15/01/2022	04:13	MB	Butanona	3MM Whatman			26		1257	0	pH (Low/Actual/High)	97,973 %	Pass 5,00 <= 5,00 <= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20220115-008	15/01/2022	04:49	MB	Butanona	3MM Whatman			8		6659	0	pH (Low/Actual/High)	99,88 %	Pass 5,00 <= 6,00 <= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20220115-009	15/01/2022	04:53	MB	Butanona	3MM Whatman			18		5734	0	pH (Low/Actual/High)	99,687 %	Pass 5,00 <= 6,00 <= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20220115-010	15/01/2022	04:56	MB	Butanona	3MM Whatman			26		5871	0	pH (Low/Actual/High)	99,559 %	Pass 5,00 <= 6,00 <= 6,00
														Límpido e Incolor

Product:	Manuf:	Calibration D/T	Solution	Media	Origin	Middle	Top	Solution 2	Media 2	Origin 2	Top 2	% Tag	QC By	Pass Fail
Tecnécio-99m Lot # E-20220115-011		15/01/2022 05:00	MB	Butanona		3MM Whatman		11		5845	0			99,812 % Pass
												pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20220115-012		15/01/2022 05:04	MB	Butanona		3MM Whatman		2		1513	0			99,868 % Pass
												pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20220115-013		15/01/2022 05:09	MB	Butanona		3MM Whatman		1		916	0			99,891 % Pass
												pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20220115-014		15/01/2022 05:13	MB	Butanona		3MM Whatman		8		1180	0			99,327 % Pass
												pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20220115-015		15/01/2022 05:16	MB	Butanona		3MM Whatman		13		1028	0			98,751 % Pass
												pH (Low/Actual/High)	5,00 <= 5,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20220115-016		15/01/2022 05:20	MB	Butanona		3MM Whatman		10		1128	0			99,121 % Pass
												pH (Low/Actual/High)	5,00 <= 5,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20220115-017		15/01/2022 05:24	MB	Butanona		3MM Whatman		19		1159	0			98,387 % Pass
												pH (Low/Actual/High)	5,00 <= 5,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20220115-018		15/01/2022 05:28	MB	Butanona		3MM Whatman		5		1320	0			99,623 % Pass
												pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
														Límpido e Incolor