



# Kit Quality Control Report

## RPH CENTRAL PHARMA

CNPJ: 19.315.658/0001-10

R: Maestro Cardim, 769 - São Paulo/SP

Farm. RT: Priscila Larcher CRF/SP-74.417

Date: From 20/01/2022 To 20/01/2022

Product:	Manuf:	Calibration	D/T	QC By	Solution	Media	Origin	Middle	Top	Background	Particle Size/ OK?	% Tag	Pass Fail
DMSA-99mTc Lot # K-20220120-010		20/01/2022	04:01	MB	Butanona Normal Saline	3MM Whatman Sílica gel	172 6		2 177	0		95,572 %	Pass
Límpido/Incolor													
DMSA-99mTc Lot # K-20220120-030		20/01/2022	07:36	FN	Butanona Normal Saline	3MM Whatman Sílica gel	146,4 10,6		2,5 146,9	0		91,591 %	Pass
Límpido/Incolor													
DOTATOC-68Ga Lot # K-20220120-021		20/01/2022	06:56	FN	Citrato de Sódio 0,1M	Sílica gel	44,2		0,5	0		98,881 %	Pass
Límpido e Incolor													
DOTATOC-68Ga Lot # K-20220120-057		20/01/2022	12:51	LG	Citrato de Sódio 0,1M	Sílica gel	2		0,1	0		95,238 %	Pass
Límpido e Incolor													
DTPA-99mTc Lot # K-20220120-004	RPH	20/01/2022	03:09	MB	Acetona Normal Saline	3MM Whatman 3MM Whatman	332 17		1 316	0		94,595 %	Pass
Límpido/Incolor													
DTPA-99mTc Lot # K-20220120-025	RPH	20/01/2022	07:09	FN	Acetona Normal Saline	3MM Whatman 3MM Whatman	586 12		10 397	0		95,388 %	Pass
Límpido/Incolor													
DTPA-99mTc Lot # K-20220120-046	RPH	20/01/2022	10:23	WG	Acetona Normal Saline	3MM Whatman 3MM Whatman	603 2		1 845	0		99,598 %	Pass
Límpido/Incolor													
FITATO-99mTc Lot # K-20220120-005		20/01/2022	03:11	MB	Butanona	3MM Whatman	342		2	0		99,419 %	Pass
Límpido/Incolor													
FITATO-99mTc Lot # K-20220120-024		20/01/2022	07:07	FN	Butanona	3MM Whatman	475		1	0		99,79 %	Pass
Límpido/Incolor													
MAA-99mTc Lot # K-20220120-011		20/01/2022	04:03	MB	Butanona	3MM Whatman	200		1	0		99,502 %	Pass
Turvo/Incolor													
MAA-99mTc Lot # K-20220120-012		20/01/2022	04:04	MB	Butanona	3MM Whatman	166,5		1,2	0		99,284 %	Pass
Turvo/Incolor													

Product:	Manuf:	Calibration D/T	Solution	Media	Origin	Middle	Top	Solution 2	Media 2	Origin 2	Top 2	% Tag	QC By	Pass Fail
MAA-99mTc Lot # K-20220120-013		20/01/2022 04:06	MB	Butanona		3MM Whatman		433		5	0	98,858 %	Pass	
												pH (Low/Actual/High)	3,00 <= 5,00	<= 6,00
												Turvo/Incolor		
MAA-99mTc Lot # K-20220120-022		20/01/2022 06:59	FN	Butanona		3MM Whatman		57,4		0,3	0	99,48 %	Pass	
												pH (Low/Actual/High)	3,00 <= 6,00	<= 6,00
												Turvo/Incolor		
MAA-99mTc Lot # K-20220120-023		20/01/2022 07:00	FN	Butanona		3MM Whatman		231		1	0	99,569 %	Pass	
												pH (Low/Actual/High)	3,00 <= 5,00	<= 6,00
												Turvo/Incolor		
MAA-99mTc Lot # K-20220120-026		20/01/2022 07:12	FN	Butanona		3MM Whatman		147,5		2,4	0	98,399 %	Pass	
												pH (Low/Actual/High)	3,00 <= 5,00	<= 6,00
												Turvo/Incolor		
MAA-99mTc Lot # K-20220120-049		20/01/2022 10:56	PM	Butanona		3MM Whatman		254		2	0	99,219 %	Pass	
												pH (Low/Actual/High)	3,00 <= 6,00	<= 6,00
												Turvo/Incolor		
MAA-99mTc Lot # K-20220120-051		20/01/2022 11:16	PM	Butanona		3MM Whatman		384		3	0	99,225 %	Pass	
												pH (Low/Actual/High)	3,00 <= 6,00	<= 6,00
												Turvo/Incolor		
MAA-99mTc Lot # K-20220120-053		20/01/2022 11:52	VS	Butanona		3MM Whatman		1098		8	0	99,277 %	Pass	
												pH (Low/Actual/High)	3,00 <= 5,00	<= 6,00
												Turvo/Incolor		
MDP-99mTc Lot # K-20220120-006		20/01/2022 03:25	MB	Acetona		3MM Whatman		1093		11	0	97,415 %	Pass	
				Normal Saline		3MM Whatman		17		1053		pH (Low/Actual/High)	4,00 <= 5,00	<= 7,80
												Límpido/Incolor		
MDP-99mTc Lot # K-20220120-007		20/01/2022 03:27	MB	Acetona		3MM Whatman		1083		2	0	97,414 %	Pass	
				Normal Saline		3MM Whatman		25		1016		pH (Low/Actual/High)	4,00 <= 6,00	<= 7,80
												Límpido/Incolor		
MDP-99mTc Lot # K-20220120-008		20/01/2022 03:31	MB	Acetona		3MM Whatman		1078		1	0	99,11 %	Pass	
				Normal Saline		3MM Whatman		9		1120		pH (Low/Actual/High)	4,00 <= 6,00	<= 7,80
												Límpido/Incolor		
MDP-99mTc Lot # K-20220120-015		20/01/2022 04:32	MB	Acetona		3MM Whatman		797		2	0	99,253 %	Pass	
				Normal Saline		3MM Whatman		6		1201		pH (Low/Actual/High)	4,00 <= 6,00	<= 7,80
												Límpido/Incolor		
MDP-99mTc Lot # K-20220120-016		20/01/2022 04:34	MB	Acetona		3MM Whatman		1476		5	0	98,51 %	Pass	
				Normal Saline		3MM Whatman		15		1287		pH (Low/Actual/High)	4,00 <= 6,00	<= 7,80
												Límpido/Incolor		
MDP-99mTc Lot # K-20220120-018		20/01/2022 04:41	MB	Acetona		3MM Whatman		935		6	0	98,286 %	Pass	
				Normal Saline		3MM Whatman		10		919		pH (Low/Actual/High)	4,00 <= 5,00	<= 7,80
												Límpido/Incolor		

Product:	Manuf:	Calibration D/T	Solution	Media	Origin	Middle	Top	Solution 2	Media 2	Origin 2	Top 2	% Tag	QC By	Pass Fail
MDP-99mTc Lot # K-20220120-034		20/01/2022 08:02 FN	Acetona Normal Saline		3MM Whatman 3MM Whatman			309 2		1 425	0 pH (Low/Actual/High)	99,209 %		Pass <= 7,80
Límpido/Incolor														
MDP-99mTc Lot # K-20220120-035		20/01/2022 08:05 FN	Acetona Normal Saline		3MM Whatman 3MM Whatman			453 2		2 463	0 pH (Low/Actual/High)	99,13 %		Pass <= 7,80
Límpido/Incolor														
MDP-99mTc Lot # K-20220120-036		20/01/2022 08:18 FN	Acetona Normal Saline		3MM Whatman 3MM Whatman			1014 3		2 1008	0 pH (Low/Actual/High)	99,506 %		Pass <= 7,80
Límpido/Incolor														
MDP-99mTc Lot # K-20220120-037		20/01/2022 08:30 FN	Acetona Normal Saline		3MM Whatman 3MM Whatman			652 3		5 442	0 pH (Low/Actual/High)	98,565 %		Pass <= 7,80
Límpido/Incolor														
MDP-99mTc Lot # K-20220120-038		20/01/2022 08:33 FN	Acetona Normal Saline		3MM Whatman 3MM Whatman			458 2		3 478	0 pH (Low/Actual/High)	98,933 %		Pass <= 7,80
Límpido/Incolor														
MDP-99mTc Lot # K-20220120-048		20/01/2022 10:34 WG	Acetona Normal Saline		3MM Whatman 3MM Whatman			405 7		2 427	0 pH (Low/Actual/High)	97,896 %		Pass <= 7,80
Límpido/Incolor														
MDP-99mTc Lot # K-20220120-050		20/01/2022 11:11 PM	Acetona Normal Saline		3MM Whatman 3MM Whatman			1116 20		1 1116	0 pH (Low/Actual/High)	98,15 %		Pass <= 7,80
Límpido/Incolor														
MDP-99mTc Lot # K-20220120-052		20/01/2022 11:19 PM	Acetona Normal Saline		3MM Whatman 3MM Whatman			1151 18		1 1273	0 pH (Low/Actual/High)	98,519 %		Pass <= 7,80
Límpido/Incolor														
MIBI-99mTc Lot # K-20220120-001		20/01/2022 02:44 MB	Normal Saline Metanol		Sílica gel 3MM Whatman			1208 7		15 1760	0 pH (Low/Actual/High)	98,377 %		Pass <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20220120-002		20/01/2022 02:46 MB	Normal Saline Metanol		Sílica gel 3MM Whatman			1268 7		15 1457	0 pH (Low/Actual/High)	98,353 %		Pass <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20220120-003		20/01/2022 03:05 MB	Normal Saline Metanol		Sílica gel 3MM Whatman			1132 33		9 1519	0 pH (Low/Actual/High)	97,085 %		Pass <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20220120-009		20/01/2022 03:54 MB	Normal Saline Metanol		Sílica gel 3MM Whatman			1208 8		12 1398	0 pH (Low/Actual/High)	98,447 %		Pass <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20220120-014		20/01/2022 04:18 MB	Normal Saline Metanol		Sílica gel 3MM Whatman			1624 27		11 1487	0 pH (Low/Actual/High)	97,544 %		Pass <= 6,00
Límpido/Incolor														

Product:	Manuf:	Calibration D/T	Solution	Media	Origin	Middle	Top	Solution 2	Media 2	Origin 2	Top 2	% Tag	QC By	Pass Fail
MIBI-99mTc Lot # K-20220120-027		20/01/2022 07:15 FN	Normal Saline Metanol		Silica gel 3MM Whatman			837 12		9 769	0	pH (Low/Actual/High)	97,40 %	Pass 5,00 <= 6,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20220120-028		20/01/2022 07:19 FN	Normal Saline Metanol		Silica gel 3MM Whatman			755 8		10 1319	0	pH (Low/Actual/High)	98,09 %	Pass 5,00 <= 6,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20220120-029		20/01/2022 07:22 FN	Normal Saline Metanol		Silica gel 3MM Whatman			220 2		4 341	0	pH (Low/Actual/High)	97,631 %	Pass 5,00 <= 6,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20220120-031		20/01/2022 07:47 FN	Normal Saline Metanol		Silica gel 3MM Whatman			725 1		9 747	0	pH (Low/Actual/High)	98,64 %	Pass 5,00 <= 5,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20220120-032		20/01/2022 07:49 FN	Normal Saline Metanol		Silica gel 3MM Whatman			665 4		12 759	0	pH (Low/Actual/High)	97,703 %	Pass 5,00 <= 6,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20220120-044		20/01/2022 10:18 WG	Normal Saline Metanol		Silica gel 3MM Whatman			594 10		4 958	0	pH (Low/Actual/High)	98,298 %	Pass 5,00 <= 6,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20220120-045		20/01/2022 10:21 WG	Normal Saline Metanol		Silica gel 3MM Whatman			1258 5		15 765	0	pH (Low/Actual/High)	98,172 %	Pass 5,00 <= 6,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20220120-047		20/01/2022 10:30 PM	Normal Saline Metanol		Silica gel 3MM Whatman			701 21		4 914	0	pH (Low/Actual/High)	97,187 %	Pass 5,00 <= 5,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20220120-054		20/01/2022 12:24 LG	Normal Saline Metanol		Silica gel 3MM Whatman			1134 2		5 1132	0	pH (Low/Actual/High)	99,385 %	Pass 5,00 <= 6,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20220120-055		20/01/2022 12:32 LG	Normal Saline Metanol		Silica gel 3MM Whatman			1072 3		11 1648	0	pH (Low/Actual/High)	98,803 %	Pass 5,00 <= 6,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20220120-056		20/01/2022 12:40 LG	Normal Saline Metanol		Silica gel 3MM Whatman			341 4		7 569	0	pH (Low/Actual/High)	97,29 %	Pass 5,00 <= 6,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20220120-058		20/01/2022 13:07 LG	Normal Saline Metanol		Silica gel 3MM Whatman			239 1		7 249	0	pH (Low/Actual/High)	96,754 %	Pass 5,00 <= 6,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20220120-059		20/01/2022 13:18 LG	Normal Saline Metanol		Silica gel 3MM Whatman			139 1		4 486	0	pH (Low/Actual/High)	96,997 %	Pass 5,00 <= 6,00 <= 6,00
Límpido/Incolor														

Product:	Manuf:	Calibration D/T	Solution	Media	Origin	Middle	Top	Solution 2	Media 2	Origin 2	Top 2	% Tag	QC By	Pass Fail
MIBI-99mTc Lot # K-20220120-060	20/01/2022	14:30	WG	Normal Saline Metanol	Sílica gel 3MM Whatman			148 2		0,3 239	0	pH (Low/Actual/High)	98,968 %	Pass 5,00 <= 6,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20220120-061	20/01/2022	14:33	WG	Normal Saline Metanol	Sílica gel 3MM Whatman			287 3		6 469	0	pH (Low/Actual/High)	97,317 %	Pass 5,00 <= 6,00 <= 6,00
Límpido/Incolor														
PSMA-68Ga Lot # K-20220120-039	20/01/2022	09:40	WG	Acetato de amônio 1M:Metanol	Sílica gel			0,5		20,2	0	pH (Low/Actual/High)	97,585 %	Pass 4,00 <= 7,00 <= 8,00
Límpido e Incolor														
Tecnécio-99m Lot # E-20220120-001	20/01/2022	02:19	MB	Butanona	3MM Whatman			2		278	0	pH (Low/Actual/High)	99,286 %	Pass 5,00 <= 5,00 <= 6,00
Límpido e Incolor														
Tecnécio-99m Lot # E-20220120-002	20/01/2022	02:39	MB	Butanona	3MM Whatman			0,5		38,5	0	pH (Low/Actual/High)	98,718 %	Pass 5,00 <= 5,00 <= 6,00
Límpido e Incolor														
Tecnécio-99m Lot # E-20220120-003	20/01/2022	02:55	MB	Butanona	3MM Whatman			3		431	0	pH (Low/Actual/High)	99,309 %	Pass 5,00 <= 6,00 <= 6,00
Límpido e Incolor														
Tecnécio-99m Lot # E-20220120-004	20/01/2022	03:02	MB	Butanona	3MM Whatman			1		472	0	pH (Low/Actual/High)	99,789 %	Pass 5,00 <= 5,00 <= 6,00
Límpido e Incolor														
Tecnécio-99m Lot # E-20220120-005	20/01/2022	03:06	MB	Butanona	3MM Whatman			1		44,7	0	pH (Low/Actual/High)	97,812 %	Pass 5,00 <= 5,00 <= 6,00
Límpido e Incolor														
Tecnécio-99m Lot # E-20220120-006	20/01/2022	03:22	MB	Butanona	3MM Whatman			1		407	0	pH (Low/Actual/High)	99,755 %	Pass 5,00 <= 5,00 <= 6,00
Límpido e Incolor														
Tecnécio-99m Lot # E-20220120-007	20/01/2022	03:22	MB	Butanona	3MM Whatman			4		2352	0	pH (Low/Actual/High)	99,83 %	Pass 5,00 <= 6,00 <= 6,00
Límpido e Incolor														
Tecnécio-99m Lot # E-20220120-008	20/01/2022	03:51	MB	Butanona	3MM Whatman			2		241	0	pH (Low/Actual/High)	99,177 %	Pass 5,00 <= 5,00 <= 6,00
Límpido e Incolor														
Tecnécio-99m Lot # E-20220120-009	20/01/2022	03:59	MB	Butanona	3MM Whatman			1		373	0	pH (Low/Actual/High)	99,733 %	Pass 5,00 <= 5,00 <= 6,00
Límpido e Incolor														
Tecnécio-99m Lot # E-20220120-010	20/01/2022	04:16	MB	Butanona	3MM Whatman			3		625	0	pH (Low/Actual/High)	99,522 %	Pass 5,00 <= 5,00 <= 6,00
Límpido e Incolor														

Product:	Manuf:	Calibration D/T	Solution	Media	Origin	Middle	Top	Solution 2	Media 2	Origin 2	Top 2	% Tag	QC By	Pass Fail
Tecnécio-99m Lot # E-20220120-011		20/01/2022 04:30	MB	Butanona		3MM Whatman		2		157	0	98,742 %	Pass	
											pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00	
														Límpido e Incolor
Tecnécio-99m Lot # E-20220120-012		20/01/2022 04:33	MB	Butanona		3MM Whatman		3		411	0	99,275 %	Pass	
											pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00	
														Límpido e Incolor
Tecnécio-99m Lot # E-20220120-013		20/01/2022 06:45	FN	Butanona		3MM Whatman		3		749	0	99,601 %	Pass	
											pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00	
														Límpido e Incolor
Tecnécio-99m Lot # E-20220120-015		20/01/2022 07:01	FN	Butanona		3MM Whatman		5		410	0	98,795 %	Pass	
											pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00	
														Límpido e Incolor
Tecnécio-99m Lot # E-20220120-016		20/01/2022 07:03	FN	Butanona		3MM Whatman		6		558	0	98,936 %	Pass	
											pH (Low/Actual/High)	5,00 <= 5,00	<= 6,00	
														Límpido e Incolor
Tecnécio-99m Lot # E-20220120-017		20/01/2022 07:13	FN	Butanona		3MM Whatman		2		306	0	99,351 %	Pass	
											pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00	
														Límpido e Incolor
Tecnécio-99m Lot # E-20220120-018		20/01/2022 07:17	FN	Butanona		3MM Whatman		3		231	0	98,718 %	Pass	
											pH (Low/Actual/High)	5,00 <= 5,00	<= 6,00	
														Límpido e Incolor
Tecnécio-99m Lot # E-20220120-019		20/01/2022 07:20	FN	Butanona		3MM Whatman		2		250	0	99,206 %	Pass	
											pH (Low/Actual/High)	5,00 <= 5,00	<= 6,00	
														Límpido e Incolor
Tecnécio-99m Lot # E-20220120-021		20/01/2022 07:47	FN	Butanona		3MM Whatman		2		280	0	99,291 %	Pass	
											pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00	
														Límpido e Incolor
Tecnécio-99m Lot # E-20220120-023		20/01/2022 07:58	FN	Butanona		3MM Whatman		2		396	0	99,497 %	Pass	
											pH (Low/Actual/High)	5,00 <= 5,00	<= 6,00	
														Límpido e Incolor
Tecnécio-99m Lot # E-20220120-022		20/01/2022 07:58	FN	Butanona		3MM Whatman		2		1731	0	99,885 %	Pass	
											pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00	
														Límpido e Incolor
Tecnécio-99m Lot # E-20220120-024		20/01/2022 08:15	FN	Butanona		3MM Whatman		3		227	0	98,696 %	Pass	
											pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00	
														Límpido e Incolor
Tecnécio-99m Lot # E-20220120-020		20/01/2022 08:20	FN	Butanona		3MM Whatman		4		318	0	98,758 %	Pass	
											pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00	
														Límpido e Incolor

Product:	Manuf:	Calibration D/T	Solution	Media	Origin	Middle	Top	Solution 2	Media 2	Origin 2	Top 2	% Tag	QC By	Pass Fail
Tecnécio-99m Lot # E-20220120-025		20/01/2022 08:28	FN	Butanona		3MM Whatman		4		493	0		99,195 %	Pass
												pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20220120-026		20/01/2022 08:31	FN	Butanona		3MM Whatman		0,6		151,5	0		99,606 %	Pass
												pH (Low/Actual/High)	5,00 <= 5,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20220120-028		20/01/2022 09:54	WG	Butanona		3MM Whatman		3		248	0		98,805 %	Pass
												pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20220120-029		20/01/2022 10:14	WG	Butanona		3MM Whatman		2,6		167,1	0		98,468 %	Pass
												pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20220120-030		20/01/2022 10:18	WG	Butanona		3MM Whatman		1,3		153,1	0		99,158 %	Pass
												pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20220120-031		20/01/2022 10:21	WG	Butanona		3MM Whatman		2		374	0		99,468 %	Pass
												pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20220120-032		20/01/2022 10:28	WG	Butanona		3MM Whatman		1,1		164,6	0		99,336 %	Pass
												pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20220120-033		20/01/2022 10:32	WG	Butanona		3MM Whatman		2,1		70,6	0		97,111 %	Pass
												pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20220120-034		20/01/2022 10:53	VS	Butanona		3MM Whatman		7		1518	0		99,541 %	Pass
												pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20220120-035		20/01/2022 11:06	PM	Butanona		3MM Whatman		2		352	0		99,435 %	Pass
												pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20220120-036		20/01/2022 11:08	PM	Butanona		3MM Whatman		12		669	0		98,238 %	Pass
												pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20220120-037		20/01/2022 11:49	PM	Butanona		3MM Whatman		10		665	0		98,519 %	Pass
												pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20220120-038		20/01/2022 12:20	WG	Butanona		3MM Whatman		1		236	0		99,578 %	Pass
												pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
														Límpido e Incolor

Product:	Manuf:	Calibration D/T	Solution	Media	Origin	Middle	Top	Solution 2	Media 2	Origin 2	Top 2	% Tag	QC By	Pass Fail
Tecnécio-99m Lot # E-20220120-039		20/01/2022 12:21	WG	Butanona		3MM Whatman		2		158	0	98,75 %	Pass	
												pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20220120-040		20/01/2022 12:30	WG	Butanona		3MM Whatman		6		587	0	98,988 %	Pass	
												pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20220120-041		20/01/2022 12:33	WG	Butanona		3MM Whatman		3		258	0	98,851 %	Pass	
												pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20220120-043		20/01/2022 13:05	LG	Butanona		3MM Whatman		3		786	0	99,62 %	Pass	
												pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20220120-044		20/01/2022 13:11	LG	Butanona		3MM Whatman		12		523	0	97,757 %	Pass	
												pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20220120-045		20/01/2022 14:20	VS	Butanona		3MM Whatman		7		832	0	99,166 %	Pass	
												pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20220120-046		20/01/2022 14:27	VS	Butanona		3MM Whatman		6		752	0	99,208 %	Pass	
												pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
														Límpido e Incolor
TRODAT-99mTc Lot # K-20220120-019		20/01/2022 06:51	FN	Normal Saline		Sílica gel		117,2		3,1	0	95,676 %	Pass	
				Butanona		3MM Whatman		4		225		pH (Low/Actual/High)	6,50 <= 7,00	<= 8,50
														Límpido/Incolor
TRODAT-99mTc Lot # K-20220120-020		20/01/2022 06:52	FN	Normal Saline		Sílica gel		135,5		3	0	94,26 %	Pass	
				Butanona		3MM Whatman		5,7		153,8		pH (Low/Actual/High)	6,50 <= 7,00	<= 8,50
														Límpido/Incolor
TRODAT-99mTc Lot # K-20220120-040		20/01/2022 09:59	WG	Normal Saline		Sílica gel		174,5		2	0	96,522 %	Pass	
				Butanona		3MM Whatman		4,5		187,4		pH (Low/Actual/High)	6,50 <= 7,00	<= 8,50
														Límpido/Incolor
TRODAT-99mTc Lot # K-20220120-041		20/01/2022 10:00	MP	Normal Saline		Sílica gel		56,8		2,2	0	93,553 %	Pass	
				Butanona		3MM Whatman		2,4		85,9		pH (Low/Actual/High)	6,50 <= 7,00	<= 8,50
														Límpido/Incolor
TRODAT-99mTc Lot # K-20220120-042		20/01/2022 10:04	WG	Normal Saline		Sílica gel		169,8		2	0	97,372 %	Pass	
				Butanona		3MM Whatman		3		202		pH (Low/Actual/High)	6,50 <= 7,00	<= 8,50
														Límpido/Incolor
TRODAT-99mTc Lot # K-20220120-043		20/01/2022 10:05	WG	Normal Saline		Sílica gel		133,5		2,1	0	98,022 %	Pass	
				Butanona		3MM Whatman		0,8		185,4		pH (Low/Actual/High)	6,50 <= 7,00	<= 8,50
														Límpido/Incolor