



Kit Quality Control Report

RPH CENTRAL PHARMA

CNPJ: 19.315.658/0001-10
 R: Maestro Cardim, 769 - São Paulo/SP
 Farm. RT: Priscila Larcher CRF/SP-74.417

Date: From 27/01/2022 To 27/01/2022

Product:	Manuf:	Calibration	D/T	QC By	Solution	Media	Origin	Middle	Top	Background	Particle Size/ OK?	% Tag	Pass Fail
DMSA-99mTc		27/01/2022	07:49	MB	Butanona	3MM Whatman	143,1		0,3	0		98,372 %	Pass
Lot # K-20220127-025					Normal Saline	Silica gel		1,9	132	pH (Low/Actual/High)		2,00 <= 3,00	<= 3,00
						Límpido/Incolor							
DMSA-99mTc		27/01/2022	07:50	MB	Butanona	3MM Whatman	86		1,4	0		97,118 %	Pass
Lot # K-20220127-026					Normal Saline	Silica gel		1,4	108	pH (Low/Actual/High)		2,00 <= 3,00	<= 3,00
						Límpido/Incolor							
DMSA-99mTc		27/01/2022	10:39	VS	Butanona	3MM Whatman	301		11	0		95,051 %	Pass
Lot # K-20220127-045					Normal Saline	Silica gel		4	277	pH (Low/Actual/High)		2,00 <= 3,00	<= 3,00
						Límpido/Incolor							
DMSA-99mTc		27/01/2022	11:16	VS	Butanona	3MM Whatman	645		25	0		95,718 %	Pass
Lot # K-20220127-049					Normal Saline	Silica gel		3	542	pH (Low/Actual/High)		2,00 <= 3,00	<= 3,00
						Límpido/Incolor							
DOTATOC-68Ga		27/01/2022	06:59	BB	Citrato de Sódio 0,1M	Silica gel	13,2		0,3	0		97,778 %	Pass
Lot # K-20220127-018										pH (Low/Actual/High)		4,00 <= 5,00	<= 8,00
						Límpido e Incolor							
DOTATOC-68Ga		27/01/2022	08:01	BB	Citrato de Sódio 0,1M	Silica gel	12,4		0,2	0		98,413 %	Pass
Lot # K-20220127-029										pH (Low/Actual/High)		4,00 <= 5,00	<= 8,00
						Límpido e Incolor							
DOTATOC-68Ga		27/01/2022	12:55	YD	Citrato de Sódio 0,1M	Silica gel	21		0,3	0		98,592 %	Pass
Lot # K-20220127-058										pH (Low/Actual/High)		4,00 <= 5,00	<= 8,00
						Límpido e Incolor							
DTPA-99mTc	RPH	27/01/2022	03:30	MB	Acetona	3MM Whatman	332		1	0		96,214 %	Pass
Lot # K-20220127-005					Normal Saline	3MM Whatman		13	360	pH (Low/Actual/High)		3,80 <= 5,00	<= 7,50
						Límpido/Incolor							
DTPA-99mTc	RPH	27/01/2022	04:39	MB	Acetona	3MM Whatman	410		1	0		97,742 %	Pass
Lot # K-20220127-014					Normal Saline	3MM Whatman		8	389	pH (Low/Actual/High)		3,80 <= 6,00	<= 7,50
						Límpido/Incolor							
DTPA-99mTc	RPH	27/01/2022	08:20	MB	Acetona	3MM Whatman	178		2	0		96,909 %	Pass
Lot # K-20220127-032					Normal Saline	3MM Whatman		4	198	pH (Low/Actual/High)		3,80 <= 5,00	<= 7,50
						Límpido/Incolor							
DTPA-99mTc	RPH	27/01/2022	12:04	VS	Acetona	3MM Whatman	903		7	0		98,399 %	Pass
Lot # K-20220127-052					Normal Saline	3MM Whatman		6	715	pH (Low/Actual/High)		3,80 <= 6,00	<= 7,50
						Límpido/Incolor							

Product:	Manuf:	Calibration D/T	Solution	Media	Origin	Middle	Top	Solution 2	Media 2	Origin 2	Top 2	% Tag	QC By	Pass Fail
FITATO-99mTc Lot # K-20220127-004		27/01/2022 03:13	MB	Butanona		3MM Whatman		389		1	0	99,744 %	Pass	
											pH (Low/Actual/High)	5,00 <= 6,00	<= 7,00	
											Límpido/Incolor			
FITATO-99mTc Lot # K-20220127-027		27/01/2022 07:53	MB	Butanona		3MM Whatman		165		2	0	98,802 %	Pass	
											pH (Low/Actual/High)	5,00 <= 6,00	<= 7,00	
											Límpido/Incolor			
FITATO-99mTc Lot # K-20220127-051		27/01/2022 11:57	VS	Butanona		3MM Whatman		816		6	0	99,27 %	Pass	
											pH (Low/Actual/High)	5,00 <= 6,00	<= 7,00	
											Límpido/Incolor			
MAA-99mTc Lot # K-20220127-013		27/01/2022 04:26	MB	Butanona		3MM Whatman		91,9	0,6		0	99,351 %	Pass	
											pH (Low/Actual/High)	3,00 <= 5,00	<= 6,00	
											Turvo/Incolor			
MAA-99mTc Lot # K-20220127-024		27/01/2022 07:27	MB	Butanona		3MM Whatman		140		2	0	98,592 %	Pass	
											pH (Low/Actual/High)	3,00 <= 4,00	<= 6,00	
											Turvo/Incolor			
MAA-99mTc Lot # K-20220127-055		27/01/2022 12:40	VS	Butanona		3MM Whatman		782		2	0	99,745 %	Pass	
											pH (Low/Actual/High)	3,00 <= 5,00	<= 6,00	
											Turvo/Incolor			
MAA-99mTc Lot # K-20220127-056		27/01/2022 12:47	VS	Butanona		3MM Whatman		341		4	0	98,841 %	Pass	
											pH (Low/Actual/High)	3,00 <= 5,00	<= 6,00	
											Turvo/Incolor			
MAA-99mTc Lot # K-20220127-057		27/01/2022 12:50	VS	Butanona		3MM Whatman		231		2	0	99,142 %	Pass	
											pH (Low/Actual/High)	3,00 <= 5,00	<= 6,00	
											Turvo/Incolor			
MDP-99mTc Lot # K-20220127-007		27/01/2022 03:49	MB	Acetona		3MM Whatman		1010		2	0	98,801 %	Pass	
				Normal Saline		3MM Whatman		9		890	pH (Low/Actual/High)	4,00 <= 6,00	<= 7,80	
											Límpido/Incolor			
MDP-99mTc Lot # K-20220127-008		27/01/2022 03:51	MB	Acetona		3MM Whatman		308		1	0	97,767 %	Pass	
				Normal Saline		3MM Whatman		11		565	pH (Low/Actual/High)	4,00 <= 5,00	<= 7,80	
											Límpido/Incolor			
MDP-99mTc Lot # K-20220127-009		27/01/2022 03:55	MB	Acetona		3MM Whatman		683		1	0	98,656 %	Pass	
				Normal Saline		3MM Whatman		8		660	pH (Low/Actual/High)	4,00 <= 6,00	<= 7,80	
											Límpido/Incolor			
MDP-99mTc Lot # K-20220127-011		27/01/2022 04:17	MB	Acetona		3MM Whatman		1091		3	0	97,881 %	Pass	
				Normal Saline		3MM Whatman		19		1011	pH (Low/Actual/High)	4,00 <= 5,00	<= 7,80	
											Límpido/Incolor			
MDP-99mTc Lot # K-20220127-012		27/01/2022 04:22	MB	Acetona		3MM Whatman		1311		2	0	99,544 %	Pass	
				Normal Saline		3MM Whatman		3		985	pH (Low/Actual/High)	4,00 <= 5,00	<= 7,80	
											Límpido/Incolor			

Product:	Manuf:	Calibration D/T	Solution	Media	Origin	Middle	Top	Solution 2	Media 2	Origin 2	Top 2	% Tag	QC By	Pass Fail
MDP-99mTc Lot # K-20220127-016		27/01/2022 04:54	MB	Acetona Normal Saline	3MM Whatman 3MM Whatman			346 1		1 381	0	pH (Low/Actual/High)	99,45 %	Pass 4,00 <= 6,00 <= 7,80
Límpido/Incolor														
MDP-99mTc Lot # K-20220127-017		27/01/2022 04:56	MB	Acetona Normal Saline	3MM Whatman 3MM Whatman			576 1		1 491	0	pH (Low/Actual/High)	99,623 %	Pass 4,00 <= 5,00 <= 7,80
Límpido/Incolor														
MDP-99mTc Lot # K-20220127-033		27/01/2022 08:23	MB	Acetona Normal Saline	3MM Whatman 3MM Whatman			857 4		3 506	0	pH (Low/Actual/High)	98,867 %	Pass 4,00 <= 6,00 <= 7,80
Límpido/Incolor														
MDP-99mTc Lot # K-20220127-034		27/01/2022 08:32	MB	Acetona Normal Saline	3MM Whatman 3MM Whatman			646 13		1 569	0	pH (Low/Actual/High)	97,612 %	Pass 4,00 <= 5,00 <= 7,80
Límpido/Incolor														
MDP-99mTc Lot # K-20220127-046		27/01/2022 10:43	VS	Acetona Normal Saline	3MM Whatman 3MM Whatman			836 6		5 673	0	pH (Low/Actual/High)	98,522 %	Pass 4,00 <= 6,00 <= 7,80
Límpido/Incolor														
MDP-99mTc Lot # K-20220127-062		27/01/2022 13:59	WG	Acetona Normal Saline	3MM Whatman 3MM Whatman			173 8		0,9 269	0	pH (Low/Actual/High)	96,594 %	Pass 4,00 <= 6,00 <= 7,80
Límpido/Incolor														
MIBI-99mTc Lot # K-20220127-001		27/01/2022 02:44	MB	Normal Saline Metanol	Sílica gel 3MM Whatman			1017 5		15 1571	0	pH (Low/Actual/High)	98,229 %	Pass 5,00 <= 5,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20220127-002		27/01/2022 02:45	MB	Normal Saline Metanol	Sílica gel 3MM Whatman			1015 2		15 1586	0	pH (Low/Actual/High)	98,418 %	Pass 5,00 <= 5,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20220127-003		27/01/2022 03:05	MB	Normal Saline Metanol	Sílica gel 3MM Whatman			1087 5		9 2008	0	pH (Low/Actual/High)	98,93 %	Pass 5,00 <= 5,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20220127-006		27/01/2022 03:38	MB	Normal Saline Metanol	Sílica gel 3MM Whatman			1301 11		8 1799	0	pH (Low/Actual/High)	98,781 %	Pass 5,00 <= 5,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20220127-010		27/01/2022 04:14	MB	Normal Saline Metanol	Sílica gel 3MM Whatman			973 7		9 1371	0	pH (Low/Actual/High)	98,576 %	Pass 5,00 <= 6,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20220127-015		27/01/2022 04:41	MB	Normal Saline Metanol	Sílica gel 3MM Whatman			1185 7		10 1357	0	pH (Low/Actual/High)	98,65 %	Pass 5,00 <= 5,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20220127-019		27/01/2022 07:14	MB	Normal Saline Metanol	Sílica gel 3MM Whatman			425 1		5 667	0	pH (Low/Actual/High)	98,688 %	Pass 5,00 <= 5,00 <= 6,00
Límpido/Incolor														

Product:	Manuf:	Calibration D/T	Solution	Media	Origin	Middle	Top	Solution 2	Media 2	Origin 2	Top 2	% Tag	QC By	Pass Fail
MIBI-99mTc Lot # K-20220127-023		27/01/2022 07:26	MB	Normal Saline Metanol	Silica gel 3MM Whatman			415 1		14 375	0	pH (Low/Actual/High)	96,471 %	Pass 5,00 <= 5,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20220127-030		27/01/2022 08:09	MB	Normal Saline Metanol	Silica gel 3MM Whatman			275 10		5 311	0	pH (Low/Actual/High)	95,099 %	Pass 5,00 <= 6,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20220127-031		27/01/2022 08:16	MB	Normal Saline Metanol	Silica gel 3MM Whatman			426 4		4 638	0	pH (Low/Actual/High)	98,447 %	Pass 5,00 <= 5,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20220127-035		27/01/2022 08:38	MB	Normal Saline Metanol	Silica gel 3MM Whatman			239 3		3 365	0	pH (Low/Actual/High)	97,945 %	Pass 5,00 <= 5,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20220127-036		27/01/2022 10:02	VS	Normal Saline Metanol	Silica gel 3MM Whatman			1455 2		5 231	0	pH (Low/Actual/High)	98,799 %	Pass 5,00 <= 6,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20220127-042		27/01/2022 10:33	VS	Normal Saline Metanol	Silica gel 3MM Whatman			478 2		12 1304	0	pH (Low/Actual/High)	97,398 %	Pass 5,00 <= 6,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20220127-048		27/01/2022 11:03	VS	Normal Saline Metanol	Silica gel 3MM Whatman			634 6		2 432	0	pH (Low/Actual/High)	98,316 %	Pass 5,00 <= 6,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20220127-050		27/01/2022 11:18	WG	Normal Saline Metanol	Silica gel 3MM Whatman			379 5		12 298	0	pH (Low/Actual/High)	95,281 %	Pass 5,00 <= 6,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20220127-053		27/01/2022 12:26	WG	Normal Saline Metanol	Silica gel 3MM Whatman			439 7		1 248	0	pH (Low/Actual/High)	97,028 %	Pass 5,00 <= 6,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20220127-054		27/01/2022 12:33	WG	Normal Saline Metanol	Silica gel 3MM Whatman			187 2		8 211	0	pH (Low/Actual/High)	94,958 %	Pass 5,00 <= 6,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20220127-059		27/01/2022 12:56	WG	Normal Saline Metanol	Silica gel 3MM Whatman			274 9		6 199	0	pH (Low/Actual/High)	93,53 %	Pass 5,00 <= 6,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20220127-060		27/01/2022 13:27	WG	Normal Saline Metanol	Silica gel 3MM Whatman			203 2		6 187	0	pH (Low/Actual/High)	96,071 %	Pass 5,00 <= 6,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20220127-061		27/01/2022 13:34	WG	Normal Saline Metanol	Silica gel 3MM Whatman			174 5		8 302	0	pH (Low/Actual/High)	93,976 %	Pass 5,00 <= 6,00 <= 6,00
Límpido/Incolor														

Product:	Manuf:	Calibration D/T	Solution	Media	Origin	Middle	Top	Solution 2	Media 2	Origin 2	Top 2	% Tag	QC By	Pass Fail
PIRO-99mTc Lot # K-20220127-028	27/01/2022	07:57	MB	Butanona Normal Saline	3MM	Whatman		428 3		1 460	0	pH (Low/Actual/High)	99,119 %	Pass 5,00 <= 6,00 <= 7,00
														Límpido/Incolor
PSMA-68Ga Lot # K-20220127-037	27/01/2022	10:05	VS	Acetato de amônio 1M:Metanol	Sílica gel			3		284	0	pH (Low/Actual/High)	98,955 %	Pass 4,00 <= 6,00 <= 8,00
														Límpido e Incolor
PSMA-68Ga Lot # K-20220127-047	27/01/2022	10:53	VS	Acetato de amônio 1M:Metanol	Sílica gel			4		269	0	pH (Low/Actual/High)	98,535 %	Pass 4,00 <= 6,00 <= 8,00
														Límpido e Incolor
PSMA-68Ga Lot # K-20220127-063	27/01/2022	15:42	VS	Acetato de amônio 1M:Metanol	Sílica gel			6		393	0	pH (Low/Actual/High)	98,496 %	Pass 4,00 <= 6,00 <= 8,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20220127-001	27/01/2022	02:38	MB	Butanona	3MM	Whatman		3		347	0	pH (Low/Actual/High)	99,143 %	Pass 5,00 <= 5,00 <= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20220127-002	27/01/2022	02:39	MB	Butanona	3MM	Whatman		2		148	0	pH (Low/Actual/High)	98,667 %	Pass 5,00 <= 6,00 <= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20220127-003	27/01/2022	03:03	MB	Butanona	3MM	Whatman		6		381	0	pH (Low/Actual/High)	98,45 %	Pass 5,00 <= 5,00 <= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20220127-004	27/01/2022	03:07	MB	Butanona	3MM	Whatman		1		320	0	pH (Low/Actual/High)	99,688 %	Pass 5,00 <= 6,00 <= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20220127-005	27/01/2022	03:11	MB	Butanona	3MM	Whatman		4		392	0	pH (Low/Actual/High)	98,99 %	Pass 5,00 <= 6,00 <= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20220127-006	27/01/2022	03:28	MB	Butanona	3MM	Whatman		3		355	0	pH (Low/Actual/High)	99,162 %	Pass 5,00 <= 6,00 <= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20220127-007	27/01/2022	03:36	MB	Butanona	3MM	Whatman		1		394	0	pH (Low/Actual/High)	99,747 %	Pass 5,00 <= 5,00 <= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20220127-008	27/01/2022	03:46	MB	Butanona	3MM	Whatman		2		237	0	pH (Low/Actual/High)	99,163 %	Pass 5,00 <= 6,00 <= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20220127-009	27/01/2022	03:50	MB	Butanona	3MM	Whatman		1		201	0	pH (Low/Actual/High)	99,505 %	Pass 5,00 <= 6,00 <= 6,00
														Límpido e Incolor

Product:	Manuf:	Calibration D/T	Solution	Media	Origin	Middle	Top	Solution 2	Media 2	Origin 2	Top 2	% Tag	QC By	Pass Fail
Tecnécio-99m Lot # E-20220127-010		27/01/2022 04:10	MB	Butanona		3MM Whatman		4		336	0		98,824 %	Pass
												pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20220127-011		27/01/2022 04:10	MB	Butanona		3MM Whatman		1		263	0		99,621 %	Pass
												pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20220127-012		27/01/2022 04:15	MB	Butanona		3MM Whatman		7		453	0		98,478 %	Pass
												pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20220127-013		27/01/2022 04:24	MB	Butanona		3MM Whatman		2		108	0		98,182 %	Pass
												pH (Low/Actual/High)	5,00 <= 5,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20220127-014		27/01/2022 04:36	MB	Butanona		3MM Whatman		5		729	0		99,319 %	Pass
												pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20220127-015		27/01/2022 04:52	MB	Butanona		3MM Whatman		1		75,8	0		98,698 %	Pass
												pH (Low/Actual/High)	5,00 <= 5,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20220127-016		27/01/2022 04:54	MB	Butanona		3MM Whatman		1		138	0		99,281 %	Pass
												pH (Low/Actual/High)	5,00 <= 5,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20220127-018		27/01/2022 06:59	MB	Butanona		3MM Whatman		1,2		99,8	0		98,812 %	Pass
												pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20220127-019		27/01/2022 07:11	MB	Butanona		3MM Whatman		7		676	0		98,975 %	Pass
												pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20220127-020		27/01/2022 07:23	MB	Butanona		3MM Whatman		1		985	0		99,899 %	Pass
												pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20220127-021		27/01/2022 07:36	MB	Butanona		3MM Whatman		2		124	0		98,413 %	Pass
												pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20220127-022		27/01/2022 07:45	MB	Butanona		3MM Whatman		3		105	0		97,222 %	Pass
												pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20220127-023		27/01/2022 07:54	MB	Butanona		3MM Whatman		5		154	0		96,855 %	Pass
												pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
														Límpido e Incolor

Product:	Manuf:	Calibration D/T	Solution	Media	Origin	Middle	Top	Solution 2	Media 2	Origin 2	Top 2	% Tag	QC By	Pass Fail
Tecnécio-99m Lot # E-20220127-024		27/01/2022 07:54	MB	Butanona		3MM Whatman		2		390	0	99,49 %	Pass	
											pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00	
														Límpido e Incolor
Tecnécio-99m Lot # E-20220127-026		27/01/2022 08:14	MB	Butanona		3MM Whatman		2		313	0	99,365 %	Pass	
											pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00	
														Límpido e Incolor
Tecnécio-99m Lot # E-20220127-027		27/01/2022 08:17	MB	Butanona		3MM Whatman		3		272	0	98,909 %	Pass	
											pH (Low/Actual/High)	5,00 <= 5,00	<= 6,00	
														Límpido e Incolor
Tecnécio-99m Lot # E-20220127-028		27/01/2022 08:21	MB	Butanona		3MM Whatman		1		232	0	99,571 %	Pass	
											pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00	
														Límpido e Incolor
Tecnécio-99m Lot # E-20220127-029		27/01/2022 08:30	MB	Butanona		3MM Whatman		2		277	0	99,283 %	Pass	
											pH (Low/Actual/High)	5,00 <= 5,00	<= 6,00	
														Límpido e Incolor
Tecnécio-99m Lot # E-20220127-030		27/01/2022 08:33	MB	Butanona		3MM Whatman		1		86,1	0	98,852 %	Pass	
											pH (Low/Actual/High)	5,00 <= 5,00	<= 6,00	
														Límpido e Incolor
Tecnécio-99m Lot # E-20220127-031		27/01/2022 08:35	MB	Butanona		3MM Whatman		3		284	0	98,955 %	Pass	
											pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00	
														Límpido e Incolor
Tecnécio-99m Lot # E-20220127-032		27/01/2022 10:00	VS	Butanona		3MM Whatman		1		209	0	99,524 %	Pass	
											pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00	
														Límpido e Incolor
Tecnécio-99m Lot # E-20220127-034		27/01/2022 10:18	VS	Butanona		3MM Whatman		12		823	0	98,563 %	Pass	
											pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00	
														Límpido e Incolor
Tecnécio-99m Lot # E-20220127-035		27/01/2022 10:18	VS	Butanona		3MM Whatman		12		835	0	98,583 %	Pass	
											pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00	
														Límpido e Incolor
Tecnécio-99m Lot # E-20220127-036		27/01/2022 10:26	VS	Butanona		3MM Whatman		9		385	0	97,716 %	Pass	
											pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00	
														Límpido e Incolor
Tecnécio-99m Lot # E-20220127-037		27/01/2022 10:41	VS	Butanona		3MM Whatman		13		1017	0	98,738 %	Pass	
											pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00	
														Límpido e Incolor
Tecnécio-99m Lot # E-20220127-039		27/01/2022 11:01	VS	Butanona		3MM Whatman		8		523	0	98,493 %	Pass	
											pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00	
														Límpido e Incolor

Product:	Manuf:	Calibration D/T	Solution	Media	Origin	Middle	Top	Solution 2	Media 2	Origin 2	Top 2	% Tag	QC By	Pass Fail
Tecnécio-99m Lot # E-20220127-040	27/01/2022	11:12	VS	Butanona		3MM Whatman		5		238	0	97,942 %	Pass	
							Límpido e Incolor				pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00	
Tecnécio-99m Lot # E-20220127-041	27/01/2022	11:13	VS	Butanona		3MM Whatman		12		288	0	96,00 %	Pass	
							Límpido e Incolor				pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00	
Tecnécio-99m Lot # E-20220127-042	27/01/2022	11:52	WG	Butanona		3MM Whatman		2		198	0	99,00 %	Pass	
							Límpido e Incolor				pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00	
Tecnécio-99m Lot # E-20220127-043	27/01/2022	12:19	WG	Butanona		3MM Whatman		8		239	0	96,761 %	Pass	
							Límpido e Incolor				pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00	
Tecnécio-99m Lot # E-20220127-044	27/01/2022	12:30	WG	Butanona		3MM Whatman		5		347	0	98,58 %	Pass	
							Límpido e Incolor				pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00	
Tecnécio-99m Lot # E-20220127-045	27/01/2022	12:37	WG	Butanona		3MM Whatman		0,3		174	0	99,828 %	Pass	
							Límpido e Incolor				pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00	
Tecnécio-99m Lot # E-20220127-047	27/01/2022	13:16	WG	Butanona		3MM Whatman		3		198	0	98,507 %	Pass	
							Límpido e Incolor				pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00	
Tecnécio-99m Lot # E-20220127-048	27/01/2022	13:24	WG	Butanona		3MM Whatman		2		403	0	99,506 %	Pass	
							Límpido e Incolor				pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00	
Tecnécio-99m Lot # E-20220127-049	27/01/2022	13:29	WG	Butanona		3MM Whatman		2		146	0	98,649 %	Pass	
							Límpido e Incolor				pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00	
Tecnécio-99m Lot # E-20220127-050	27/01/2022	13:37	WG	Butanona		3MM Whatman		5		179	0	97,283 %	Pass	
							Límpido e Incolor				pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00	
Tecnécio-99m Lot # E-20220127-051	27/01/2022	13:53	WG	Butanona		3MM Whatman		12		308	0	96,25 %	Pass	
							Límpido e Incolor				pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00	
TRODAT-99mTc Lot # K-20220127-020	27/01/2022	07:16	MB	Normal Saline		Sílica gel		261		7	0	96,759 %	Pass	
				Butanona		3MM Whatman		2		316	pH (Low/Actual/High)	6,50 <= 7,00	<= 8,50	
							Límpido/Incolor							
TRODAT-99mTc Lot # K-20220127-021	27/01/2022	07:16	MB	Normal Saline		Sílica gel		269		11	0	93,381 %	Pass	
				Butanona		3MM Whatman		6		217	pH (Low/Actual/High)	6,50 <= 7,00	<= 8,50	
							Límpido/Incolor							

Product:	Manuf:	Calibration D/T	Solution	Media	Origin	Middle	Top	Solution 2	Media 2	Origin 2	Top 2	% Tag	QC By	Pass Fail
TRODAT-99mTc Lot # K-20220127-022	27/01/2022	07:17	MB	Normal Saline	Silica gel			286		8	0	95,006 %	Pass	
				Butanona	3MM Whatman			7		301	pH (Low/Actual/High)	6,50 <= 7,00 <= 8,50		
											Límpido/Incolor			
TRODAT-99mTc Lot # K-20220127-038	27/01/2022	10:29	VS	Normal Saline	Silica gel			352		2	0	98,71 %	Pass	
				Butanona	3MM Whatman			3		411	pH (Low/Actual/High)	6,50 <= 7,00 <= 8,50		
											Límpido/Incolor			
TRODAT-99mTc Lot # K-20220127-039	27/01/2022	10:30	VS	Normal Saline	Silica gel			596		2	0	99,421 %	Pass	
				Butanona	3MM Whatman			1		408	pH (Low/Actual/High)	6,50 <= 7,00 <= 8,50		
											Límpido/Incolor			
TRODAT-99mTc Lot # K-20220127-040	27/01/2022	10:31	VS	Normal Saline	Silica gel			304		2	0	98,473 %	Pass	
				Butanona	3MM Whatman			2		227	pH (Low/Actual/High)	6,50 <= 7,00 <= 8,50		
											Límpido/Incolor			
TRODAT-99mTc Lot # K-20220127-041	27/01/2022	10:31	VS	Normal Saline	Silica gel			297		2	0	98,756 %	Pass	
				Butanona	3MM Whatman			3		519	pH (Low/Actual/High)	6,50 <= 7,00 <= 8,50		
											Límpido/Incolor			
TRODAT-99mTc Lot # K-20220127-043	27/01/2022	10:35	VS	Normal Saline	Silica gel			196,6		1,3	0	98,722 %	Pass	
				Butanona	3MM Whatman			3		480	pH (Low/Actual/High)	6,50 <= 7,00 <= 8,50		
											Límpido/Incolor			
TRODAT-99mTc Lot # K-20220127-044	27/01/2022	10:35	VS	Normal Saline	Silica gel			186,3		2,1	0	97,979 %	Pass	
				Butanona	3MM Whatman			3		328	pH (Low/Actual/High)	6,50 <= 7,00 <= 8,50		
											Límpido/Incolor			