



Kit Quality Control Report

RPH CENTRAL PHARMA

CNPJ: 19.315.658/0001-10

R: Maestro Cardim, 769 - São Paulo/SP

Farm. RT: Priscila Larcher CRF/SP-74.417

Date: From 28/01/2022 To 28/01/2022

Product:	Manuf:	Calibration	D/T	QC By	Solution	Media	Origin	Middle	Top	Background	Particle Size/ OK?	% Tag	Pass Fail
DMSA-99mTc Lot # K-20220128-017		28/01/2022	07:30	MB	Butanona	3MM Whatman	151,3		1,3	0	pH (Low/Actual/High)	98,058 %	Pass
					Normal Saline	Silica gel	1,9		172,4	2,00 <= 3,00 <= 3,00			
Límpido/Incolor													
DMSA-99mTc Lot # K-20220128-020		28/01/2022	07:46	MB	Butanona	3MM Whatman	147,4		1	0	pH (Low/Actual/High)	97,79 %	Pass
					Normal Saline	Silica gel	2		128,2	2,00 <= 3,00 <= 3,00			
Límpido/Incolor													
DMSA-99mTc Lot # K-20220128-021		28/01/2022	07:49	MB	Butanona	3MM Whatman	180,2		1,4	0	pH (Low/Actual/High)	95,692 %	Pass
					Normal Saline	Silica gel	5,1		139,1	2,00 <= 3,00 <= 3,00			
Límpido/Incolor													
DMSA-99mTc Lot # K-20220128-022		28/01/2022	07:54	MB	Butanona	3MM Whatman	204,4		2,7	0	pH (Low/Actual/High)	95,73 %	Pass
					Normal Saline	Silica gel	4,5		147,2	2,00 <= 3,00 <= 3,00			
Límpido/Incolor													
DMSA-99mTc Lot # K-20220128-033		28/01/2022	10:26	VS	Butanona	3MM Whatman	406,3		6,3	0	pH (Low/Actual/High)	96,259 %	Pass
					Normal Saline	Silica gel	5,2		229,7	2,00 <= 3,00 <= 3,00			
Límpido/Incolor													
DMSA-99mTc Lot # K-20220128-034		28/01/2022	10:27	VS	Butanona	3MM Whatman	273,9		2,6	0	pH (Low/Actual/High)	97,685 %	Pass
					Normal Saline	Silica gel	4,3		308,6	2,00 <= 3,00 <= 3,00			
Límpido/Incolor													
DOTATOC-68Ga Lot # K-20220128-049		28/01/2022	12:59	YD	Citrato de Sódio 0,1M	Silica gel	17,4		0,1	0	pH (Low/Actual/High)	99,429 %	Pass
										4,00 <= 5,00 <= 8,00			
Límpido e Incolor													
DTPA-99mTc Lot # K-20220128-005	RPH	28/01/2022	03:43	MB	Acetona	3MM Whatman	273		1	0	pH (Low/Actual/High)	98,556 %	Pass
					Normal Saline	3MM Whatman	3		275	3,80 <= 5,00 <= 7,50			
Límpido/Incolor													
DTPA-99mTc Lot # K-20220128-007	RPH	28/01/2022	03:58	MB	Acetona	3MM Whatman	265		1	0	pH (Low/Actual/High)	96,584 %	Pass
					Normal Saline	3MM Whatman	9		287	3,80 <= 6,00 <= 7,50			
Límpido/Incolor													
DTPA-99mTc Lot # K-20220128-023	RPH	28/01/2022	07:57	MB	Acetona	3MM Whatman	217		7	0	pH (Low/Actual/High)	95,986 %	Pass
					Normal Saline	3MM Whatman	2		223	3,80 <= 5,00 <= 7,50			
Límpido/Incolor													
DTPA-99mTc Lot # K-20220128-040	RPH	28/01/2022	11:27	VS	Acetona	3MM Whatman	367,4		3,8	0	pH (Low/Actual/High)	97,935 %	Pass
					Normal Saline	3MM Whatman	3		285	3,80 <= 6,00 <= 7,50			
Límpido/Incolor													

Product:	Manuf:	Calibration D/T	Solution	Media	Origin	Middle	Top	Solution 2	Media 2	Origin 2	Top 2	% Tag	QC By	Pass Fail
FITATO-99mTc Lot # K-20220128-004		28/01/2022 03:25	MB	Butanona		3MM Whatman		312		1	0	99,681 %	Pass	
											pH (Low/Actual/High)	5,00 <= 6,00	<= 7,00	
											Límpido/Incolor			
FITATO-99mTc Lot # K-20220128-028		28/01/2022 08:19	MB	Butanona		3MM Whatman		222		2	0	99,107 %	Pass	
											pH (Low/Actual/High)	5,00 <= 6,00	<= 7,00	
											Límpido/Incolor			
FITATO-99mTc Lot # K-20220128-037		28/01/2022 10:53	VS	Butanona		3MM Whatman		180		9	0	95,238 %	Pass	
											pH (Low/Actual/High)	5,00 <= 6,00	<= 7,00	
											Límpido/Incolor			
MAA-99mTc Lot # K-20220128-010		28/01/2022 04:29	MB	Butanona		3MM Whatman		193		1	0	99,485 %	Pass	
											pH (Low/Actual/High)	3,00 <= 5,00	<= 6,00	
											Turvo/Incolor			
MAA-99mTc Lot # K-20220128-027		28/01/2022 08:18	MB	Butanona		3MM Whatman		100		3	0	97,087 %	Pass	
											pH (Low/Actual/High)	3,00 <= 5,00	<= 6,00	
											Turvo/Incolor			
MAA-99mTc Lot # K-20220128-035		28/01/2022 10:33	VS	Butanona		3MM Whatman		580,2		3,2	0	99,451 %	Pass	
											pH (Low/Actual/High)	3,00 <= 5,00	<= 6,00	
											Turvo/Incolor			
MAA-99mTc Lot # K-20220128-039		28/01/2022 11:24	VS	Butanona		3MM Whatman		539		4	0	99,263 %	Pass	
											pH (Low/Actual/High)	3,00 <= 6,00	<= 6,00	
											Turvo/Incolor			
MAA-99mTc Lot # K-20220128-043		28/01/2022 12:00	MB	Butanona		3MM Whatman		105		1,1	0	98,963 %	Pass	
											pH (Low/Actual/High)	3,00 <= 5,00	<= 6,00	
											Turvo/Incolor			
MAA-99mTc Lot # K-20220128-044		28/01/2022 12:31	YD	Butanona		3MM Whatman		130,4		1,8	0	98,638 %	Pass	
											pH (Low/Actual/High)	3,00 <= 5,00	<= 6,00	
											Turvo/Incolor			
MAA-99mTc Lot # K-20220128-045		28/01/2022 12:35	YD	Butanona		3MM Whatman		115,5		1,4	0	98,802 %	Pass	
											pH (Low/Actual/High)	3,00 <= 5,00	<= 6,00	
											Turvo/Incolor			
MDP-99mTc Lot # K-20220128-008		28/01/2022 04:04	MB	Acetona		3MM Whatman		1185		3	0	99,16 %	Pass	
				Normal Saline		3MM Whatman		6		1015	pH (Low/Actual/High)	4,00 <= 6,00	<= 7,80	
											Límpido/Incolor			
MDP-99mTc Lot # K-20220128-011		28/01/2022 04:32	MB	Acetona		3MM Whatman		220		1	0	97,635 %	Pass	
				Normal Saline		3MM Whatman		7		359	pH (Low/Actual/High)	4,00 <= 6,00	<= 7,80	
											Límpido/Incolor			
MDP-99mTc Lot # K-20220128-029		28/01/2022 08:22	MB	Acetona		3MM Whatman		639		1	0	97,63 %	Pass	
				Normal Saline		3MM Whatman		12		530	pH (Low/Actual/High)	4,00 <= 6,00	<= 7,80	
											Límpido/Incolor			

Product:	Manuf:	Calibration D/T	Solution	Media	Origin	Middle	Top	Solution 2	Media 2	Origin 2	Top 2	% Tag	QC By	Pass Fail
MDP-99mTc Lot # K-20220128-031		28/01/2022 08:36	MB	Acetona Normal Saline	3MM Whatman 3MM Whatman			548 2		2 367	0 pH (Low/Actual/High)	99,094 %		Pass 4,00 <= 5,00 <= 7,80
Límpido/Incolor														
MDP-99mTc Lot # K-20220128-038		28/01/2022 11:00	VS	Acetona Normal Saline	3MM Whatman 3MM Whatman			659 2		12 801	0 pH (Low/Actual/High)	97,963 %		Pass 4,00 <= 6,00 <= 7,80
Límpido/Incolor														
MDP-99mTc Lot # K-20220128-041		28/01/2022 11:44	MB	Acetona Normal Saline	3MM Whatman 3MM Whatman			180 1		2 153	0 pH (Low/Actual/High)	98,252 %		Pass 4,00 <= 5,00 <= 7,80
Límpido/Incolor														
MDP-99mTc Lot # K-20220128-050		28/01/2022 13:43	VS	Acetona Normal Saline	3MM Whatman 3MM Whatman			1192 12		8 1779	0 pH (Low/Actual/High)	98,663 %		Pass 4,00 <= 6,00 <= 7,80
Límpido/Incolor														
MIBI-99mTc Lot # K-20220128-001		28/01/2022 03:03	MB	Normal Saline Metanol	Sílica gel 3MM Whatman			94,3 3		1,8 256	0 pH (Low/Actual/High)	96,969 %		Pass 5,00 <= 5,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20220128-002		28/01/2022 03:06	MB	Normal Saline Metanol	Sílica gel 3MM Whatman			630 4		7 968	0 pH (Low/Actual/High)	98,49 %		Pass 5,00 <= 5,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20220128-003		28/01/2022 03:22	MB	Normal Saline Metanol	Sílica gel 3MM Whatman			842 21		11 989	0 pH (Low/Actual/High)	96,631 %		Pass 5,00 <= 5,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20220128-006		28/01/2022 03:53	MB	Normal Saline Metanol	Sílica gel 3MM Whatman			271 5		5 613	0 pH (Low/Actual/High)	97,379 %		Pass 5,00 <= 5,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20220128-009		28/01/2022 04:10	MB	Normal Saline Metanol	Sílica gel 3MM Whatman			632 3		4 752	0 pH (Low/Actual/High)	98,974 %		Pass 5,00 <= 6,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20220128-012		28/01/2022 04:40	MB	Normal Saline Metanol	Sílica gel 3MM Whatman			534 6		5 579	0 pH (Low/Actual/High)	98,047 %		Pass 5,00 <= 5,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20220128-013		28/01/2022 04:45	MB	Normal Saline Metanol	Sílica gel 3MM Whatman			367 5		2 390	0 pH (Low/Actual/High)	98,192 %		Pass 5,00 <= 5,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20220128-015		28/01/2022 07:14	MB	Normal Saline Metanol	Sílica gel 3MM Whatman			86 1		1,8 162	0 pH (Low/Actual/High)	97,336 %		Pass 5,00 <= 5,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20220128-016		28/01/2022 07:16	MB	Normal Saline Metanol	Sílica gel 3MM Whatman			186 3		3 195	0 pH (Low/Actual/High)	96,898 %		Pass 5,00 <= 5,00 <= 6,00
Límpido/Incolor														

Product:	Manuf:	Calibration D/T	Solution	Media	Origin	Middle	Top	Solution 2	Media 2	Origin 2	Top 2	% Tag	QC By	Pass Fail
MIBI-99mTc Lot # K-20220128-018	28/01/2022	07:33	MB	Normal Saline Metanol	Sílica gel 3MM Whatman			114 4		2 188	0 pH (Low/Actual/High)	96,193 %	Pass	5,00 <= 5,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20220128-019	28/01/2022	07:36	MB	Normal Saline Metanol	Sílica gel 3MM Whatman			259 1		3 347	0 pH (Low/Actual/High)	98,568 %	Pass	5,00 <= 5,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20220128-025	28/01/2022	08:05	MB	Normal Saline Metanol	Sílica gel 3MM Whatman			358 4		2 565	0 pH (Low/Actual/High)	98,741 %	Pass	5,00 <= 5,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20220128-030	28/01/2022	08:25	MB	Normal Saline Metanol	Sílica gel 3MM Whatman			254 5		3 258	0 pH (Low/Actual/High)	96,932 %	Pass	5,00 <= 5,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20220128-036	28/01/2022	10:43	VS	Normal Saline Metanol	Sílica gel 3MM Whatman			1056 12		15 1516	0 pH (Low/Actual/High)	97,814 %	Pass	5,00 <= 6,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20220128-046	28/01/2022	12:36	YD	Normal Saline Metanol	Sílica gel 3MM Whatman			375 3		3 461	0 pH (Low/Actual/High)	98,56 %	Pass	5,00 <= 6,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20220128-048	28/01/2022	12:50	YD	Normal Saline Metanol	Sílica gel 3MM Whatman			431 1		7 479	0 pH (Low/Actual/High)	98,193 %	Pass	5,00 <= 6,00 <= 6,00
Límpido/Incolor														
PIRO-99mTc Lot # K-20220128-024	28/01/2022	08:02	MB	Butanona Normal Saline	3MM Whatman 3MM Whatman			301 2		1 297	0 pH (Low/Actual/High)	99,00 %	Pass	5,00 <= 6,00 <= 7,00
Límpido/Incolor														
PIRO-99mTc Lot # K-20220128-026	28/01/2022	08:08	MB	Butanona Normal Saline	3MM Whatman 3MM Whatman			265 2		3 252	0 pH (Low/Actual/High)	98,093 %	Pass	5,00 <= 5,00 <= 7,00
Límpido/Incolor														
PIRO-99mTc Lot # K-20220128-042	28/01/2022	11:56	VS	Butanona Normal Saline	3MM Whatman 3MM Whatman			809 2		10 247	0 pH (Low/Actual/High)	97,976 %	Pass	5,00 <= 6,00 <= 7,00
Límpido/Incolor														
PSMA-68Ga Lot # K-20220128-014	28/01/2022	06:49	MB	Acetato de amônio 1M:Metanol	Sílica gel			0,5		23,5	0 pH (Low/Actual/High)	97,917 %	Pass	4,00 <= 6,00 <= 8,00
Límpido e Incolor														
PSMA-68Ga Lot # K-20220128-032	28/01/2022	09:44	MB	Acetato de amônio 1M:Metanol	Sílica gel			0,3		25,8	0 pH (Low/Actual/High)	98,851 %	Pass	4,00 <= 6,00 <= 8,00
Límpido e Incolor														
Tecnécio-99m Lot # E-20220128-001	28/01/2022	02:56	MB	Butanona	3MM Whatman			3		243	0 pH (Low/Actual/High)	98,78 %	Pass	5,00 <= 6,00 <= 6,00
Límpido e Incolor														

Product:	Manuf:	Calibration D/T	Solution	Media	Origin	Middle	Top	Solution 2	Media 2	Origin 2	Top 2	% Tag	QC By	Pass Fail
Tecnécio-99m Lot # E-20220128-002		28/01/2022 02:59	MB	Butanona		3MM Whatman		1		93,2	0		98,938 %	Pass
												pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20220128-003		28/01/2022 03:19	MB	Butanona		3MM Whatman		1		199	0		99,50 %	Pass
												pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20220128-004		28/01/2022 03:23	MB	Butanona		3MM Whatman		1		308	0		99,676 %	Pass
												pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20220128-005		28/01/2022 03:29	MB	Butanona		3MM Whatman		1		59,9	0		98,358 %	Pass
												pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20220128-006		28/01/2022 03:42	MB	Butanona		3MM Whatman		1		53,4	0		98,162 %	Pass
												pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20220128-007		28/01/2022 03:51	MB	Butanona		3MM Whatman		6		197	0		97,044 %	Pass
												pH (Low/Actual/High)	5,00 <= 5,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20220128-008		28/01/2022 03:57	MB	Butanona		3MM Whatman		1		87,1	0		98,865 %	Pass
												pH (Low/Actual/High)	5,00 <= 5,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20220128-009		28/01/2022 04:01	MB	Butanona		3MM Whatman		4		182	0		97,849 %	Pass
												pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20220128-010		28/01/2022 04:06	MB	Butanona		3MM Whatman		27		676	0		96,159 %	Pass
												pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20220128-011		28/01/2022 04:27	MB	Butanona		3MM Whatman		4		435	0		99,089 %	Pass
												pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20220128-012		28/01/2022 04:38	MB	Butanona		3MM Whatman		2		305	0		99,349 %	Pass
												pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20220128-013		28/01/2022 04:42	MB	Butanona		3MM Whatman		3		385	0		99,227 %	Pass
												pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20220128-015		28/01/2022 07:10	MB	Butanona		3MM Whatman		3		172	0		98,286 %	Pass
												pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
														Límpido e Incolor

Product:	Manuf:	Calibration D/T	Solution	Media	Origin	Middle	Top	Solution 2	Media 2	Origin 2	Top 2	% Tag	QC By	Pass Fail
Tecnécio-99m Lot # E-20220128-016		28/01/2022 07:11	MB	Butanona		3MM Whatman		7		335	0		97,953 %	Pass
												pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20220128-017		28/01/2022 07:22	MB	Butanona		3MM Whatman		5		235	0		97,917 %	Pass
												pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20220128-018		28/01/2022 07:29	BB	Butanona		3MM Whatman		1,4		99,5	0		98,612 %	Pass
												pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20220128-019		28/01/2022 07:29	MB	Butanona		3MM Whatman		5		214	0		97,717 %	Pass
												pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20220128-020		28/01/2022 07:44	MB	Butanona		3MM Whatman		1		79	0		98,75 %	Pass
												pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20220128-021		28/01/2022 07:51	MB	Butanona		3MM Whatman		1		49	0		98,00 %	Pass
												pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20220128-022		28/01/2022 07:56	MB	Butanona		3MM Whatman		1		45	0		97,826 %	Pass
												pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20220128-023		28/01/2022 08:00	MB	Butanona		3MM Whatman		2		181	0		98,907 %	Pass
												pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20220128-024		28/01/2022 08:07	MB	Butanona		3MM Whatman		4		85	0		95,506 %	Pass
												pH (Low/Actual/High)	5,00 <= 5,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20220128-025		28/01/2022 08:20	MB	Butanona		3MM Whatman		1		398	0		99,749 %	Pass
												pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20220128-026		28/01/2022 08:20	MB	Butanona		3MM Whatman		4		391	0		98,987 %	Pass
												pH (Low/Actual/High)	5,00 <= 5,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20220128-027		28/01/2022 08:34	MB	Butanona		3MM Whatman		17		680	0		97,561 %	Pass
												pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20220128-029		28/01/2022 10:22	VS	Butanona		3MM Whatman		5		313,3	0		98,429 %	Pass
												pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
														Límpido e Incolor

Product:	Manuf:	Calibration D/T	Solution	Media	Origin	Middle	Top	Solution 2	Media 2	Origin 2	Top 2	% Tag	QC By	Pass Fail
Tecnécio-99m Lot # E-20220128-030		28/01/2022 10:39 VS	Butanona		3MM	Whatman		12		648	0		98,182 %	Pass
											pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00	
														Límpido e Incolor
Tecnécio-99m Lot # E-20220128-031		28/01/2022 10:56 VS	Butanona		3MM	Whatman		3		452	0		99,341 %	Pass
											pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00	
														Límpido e Incolor
Tecnécio-99m Lot # E-20220128-032		28/01/2022 11:22 VS	Butanona		3MM	Whatman		7		325	0		97,892 %	Pass
											pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00	
														Límpido e Incolor
Tecnécio-99m Lot # E-20220128-033		28/01/2022 11:35 MB	Butanona		3MM	Whatman		2		164	0		98,795 %	Pass
											pH (Low/Actual/High)	5,00 <= 5,00	<= 6,00	
														Límpido e Incolor
Tecnécio-99m Lot # E-20220128-034		28/01/2022 12:18 VS	Butanona		3MM	Whatman		14		563	0		97,574 %	Pass
											pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00	
														Límpido e Incolor
Tecnécio-99m Lot # E-20220128-035		28/01/2022 12:24 VS	Butanona		3MM	Whatman		12		698	0		98,31 %	Pass
											pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00	
														Límpido e Incolor
Tecnécio-99m Lot # E-20220128-036		28/01/2022 12:43 VS	Butanona		3MM	Whatman		6		796	0		99,252 %	Pass
											pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00	
														Límpido e Incolor
Tecnécio-99m Lot # E-20220128-038		28/01/2022 13:28 VS	Butanona		3MM	Whatman		3,7		189,2	0		98,082 %	Pass
											pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00	
														Límpido e Incolor