



Kit Quality Control Report

RPH CENTRAL PHARMA

CNPJ: 19.315.658/0001-10

R: Maestro Cardim, 769 - São Paulo/SP

Farm. RT: Priscila Larcher CRF/SP-74.417

Date: From 01/02/2022 To 01/02/2022

Product:	Manuf:	Calibration	D/T	QC By	Solution	Media	Origin	Middle	Top	Background	Particle Size/ OK?	% Tag	Pass Fail
DMSA-99mTc Lot # K-20220201-009		01/02/2022	04:02	MB	Butanona Normal Saline	3MM Whatman Sílica gel	154,1 4		1,5 97,2	0	pH (Low/Actual/High)	95,083 %	Pass 2,00 <= 3,00 <= 3,00
						Límpido/Incolor							
DMSA-99mTc Lot # K-20220201-024		01/02/2022	07:31	PM	Butanona Normal Saline	3MM Whatman Sílica gel	253 6		5 214	0	pH (Low/Actual/High)	95,335 %	Pass 2,00 <= 3,00 <= 3,00
						Límpido/Incolor							
DMSA-99mTc Lot # K-20220201-025		01/02/2022	07:32	PM	Butanona Normal Saline	3MM Whatman Sílica gel	212 10		2 195	0	pH (Low/Actual/High)	94,187 %	Pass 2,00 <= 3,00 <= 3,00
						Límpido/Incolor							
DOTATOC-68Ga Lot # K-20220201-018		01/02/2022	06:58	PM	Citrato de Sódio 0,1M	Sílica gel	60,1		1	0	pH (Low/Actual/High)	98,363 %	Pass 4,00 <= 6,00 <= 8,00
						Límpido e Incolor							
DOTATOC-68Ga Lot # K-20220201-034		01/02/2022	09:57	MB	Citrato de Sódio 0,1M	Sílica gel	106,1		2	0	pH (Low/Actual/High)	98,15 %	Pass 4,00 <= 6,00 <= 8,00
						Límpido e Incolor							
DOTATOC-68Ga Lot # K-20220201-052		01/02/2022	12:56	YD	Citrato de Sódio 0,1M	Sílica gel	12,35		0,2	0	pH (Low/Actual/High)	98,406 %	Pass 4,00 <= 5,00 <= 8,00
						Límpido e Incolor							
DTPA-99mTc Lot # K-20220201-008	RPH	01/02/2022	03:38	MB	Acetona Normal Saline	3MM Whatman 3MM Whatman	569 15		9 662	0	pH (Low/Actual/High)	96,227 %	Pass 3,80 <= 5,00 <= 7,50
						Límpido/Incolor							
DTPA-99mTc Lot # K-20220201-026	RPH	01/02/2022	07:36	PM	Acetona Normal Saline	3MM Whatman 3MM Whatman	745 2		7 745	0	pH (Low/Actual/High)	98,801 %	Pass 3,80 <= 6,00 <= 7,50
						Límpido/Incolor							
DTPA-99mTc Lot # K-20220201-041	RPH	01/02/2022	10:36	PM	Acetona Normal Saline	3MM Whatman 3MM Whatman	494 6		3 394	0	pH (Low/Actual/High)	97,896 %	Pass 3,80 <= 6,00 <= 7,50
						Límpido/Incolor							
FITATO-99mTc Lot # K-20220201-007		01/02/2022	03:37	MB	Butanona	3MM Whatman	454		4	0	pH (Low/Actual/High)	99,127 %	Pass 5,00 <= 6,00 <= 7,00
						Límpido/Incolor							
FITATO-99mTc Lot # K-20220201-042		01/02/2022	10:52	PM	Butanona	3MM Whatman	753		3	0	pH (Low/Actual/High)	99,603 %	Pass 5,00 <= 7,00 <= 7,00
						Límpido/Incolor							

Product:	Manuf:	Calibration D/T	Solution	Media	Origin	Middle	Top	Solution 2	Media 2	Origin 2	Top 2	% Tag	QC By	Pass Fail
MAA-99mTc Lot # K-20220201-012		01/02/2022 04:13	MB	Butanona		3MM Whatman		217		2	0	99,087 %	Pass	
							Turvo/Incolor				pH (Low/Actual/High)	3,00 <= 5,00	<= 6,00	
MAA-99mTc Lot # K-20220201-023		01/02/2022 07:30	PM	Butanona		3MM Whatman		347		3	0	99,143 %	Pass	
							Turvo/Incolor				pH (Low/Actual/High)	3,00 <= 5,00	<= 6,00	
MAA-99mTc Lot # K-20220201-032		01/02/2022 08:43	PM	Butanona		3MM Whatman		305		6	0	98,071 %	Pass	
							Turvo/Incolor				pH (Low/Actual/High)	3,00 <= 6,00	<= 6,00	
MAA-99mTc Lot # K-20220201-047		01/02/2022 12:15	YD	Butanona		3MM Whatman		153		1	0	99,351 %	Pass	
							Turvo/Incolor				pH (Low/Actual/High)	3,00 <= 5,00	<= 6,00	
MDP-99mTc Lot # K-20220201-003		01/02/2022 03:14	MB	Acetona		3MM Whatman		603		1	0	97,975 %	Pass	
				Normal Saline		3MM Whatman		9		475	pH (Low/Actual/High)	4,00 <= 6,00	<= 7,80	
							Límpido/Incolor							
MDP-99mTc Lot # K-20220201-004		01/02/2022 03:19	MB	Acetona		3MM Whatman		647		2	0	96,585 %	Pass	
				Normal Saline		3MM Whatman		16		499	pH (Low/Actual/High)	4,00 <= 5,00	<= 7,80	
							Límpido/Incolor							
MDP-99mTc Lot # K-20220201-015		01/02/2022 04:32	MB	Acetona		3MM Whatman		825		2	0	98,92 %	Pass	
				Normal Saline		3MM Whatman		6		710	pH (Low/Actual/High)	4,00 <= 5,00	<= 7,80	
							Límpido/Incolor							
MDP-99mTc Lot # K-20220201-016		01/02/2022 04:34	MB	Acetona		3MM Whatman		1176		1	0	98,549 %	Pass	
				Normal Saline		3MM Whatman		14		1011	pH (Low/Actual/High)	4,00 <= 5,00	<= 7,80	
							Límpido/Incolor							
MDP-99mTc Lot # K-20220201-017		01/02/2022 06:37	PM	Acetona		3MM Whatman		1251		1	0	99,615 %	Pass	
				Normal Saline		3MM Whatman		4		1305	pH (Low/Actual/High)	4,00 <= 6,00	<= 7,80	
							Límpido/Incolor							
MDP-99mTc Lot # K-20220201-021		01/02/2022 07:11	PM	Acetona		3MM Whatman		490		3	0	98,885 %	Pass	
				Normal Saline		3MM Whatman		4		786	pH (Low/Actual/High)	4,00 <= 6,00	<= 7,80	
							Límpido/Incolor							
MDP-99mTc Lot # K-20220201-030		01/02/2022 08:20	PM	Acetona		3MM Whatman		240		10	0	95,241 %	Pass	
				Normal Saline		3MM Whatman		3		392	pH (Low/Actual/High)	4,00 <= 6,00	<= 7,80	
							Límpido/Incolor							
MDP-99mTc Lot # K-20220201-031		01/02/2022 08:25	PM	Acetona		3MM Whatman		597		6	0	97,194 %	Pass	
				Normal Saline		3MM Whatman		14		759	pH (Low/Actual/High)	4,00 <= 6,00	<= 7,80	
							Límpido/Incolor							
MDP-99mTc Lot # K-20220201-043		01/02/2022 10:53	PM	Acetona		3MM Whatman		983		11	0	98,074 %	Pass	
				Normal Saline		3MM Whatman		11		1331	pH (Low/Actual/High)	4,00 <= 6,00	<= 7,80	
							Límpido/Incolor							

Product:	Manuf:	Calibration D/T	Solution	Media	Origin	Middle	Top	Solution 2	Media 2	Origin 2	Top 2	% Tag	QC By	Pass Fail
MDP-99mTc Lot # K-20220201-044		01/02/2022 10:57 PM	Acetona Normal Saline		3MM Whatman 3MM Whatman			1094 32		2 1800	0 pH (Low/Actual/High)		98,071 %	Pass 4,00 <= 6,00 <= 7,80
Límpido/Incolor														
MDP-99mTc Lot # K-20220201-050		01/02/2022 12:27 VS	Acetona Normal Saline		3MM Whatman 3MM Whatman			427 3		8 231	0 pH (Low/Actual/High)		96,879 %	Pass 4,00 <= 6,00 <= 7,80
Límpido/Incolor														
MIBI-99mTc Lot # K-20220201-001		01/02/2022 02:59 MB	Normal Saline Metanol		Sílica gel 3MM Whatman			1885 52		40 1952	0 pH (Low/Actual/High)		95,327 %	Pass 5,00 <= 5,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20220201-002		01/02/2022 03:01 MB	Normal Saline Metanol		Sílica gel 3MM Whatman			1521 14		34 1905	0 pH (Low/Actual/High)		97,084 %	Pass 5,00 <= 5,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20220201-005		01/02/2022 03:24 MB	Normal Saline Metanol		Sílica gel 3MM Whatman			1683 5		24 1961	0 pH (Low/Actual/High)		98,34 %	Pass 5,00 <= 5,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20220201-006		01/02/2022 03:24 MB	Normal Saline Metanol		Sílica gel 3MM Whatman			1627 49		4 2221	0 pH (Low/Actual/High)		97,596 %	Pass 5,00 <= 5,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20220201-010		01/02/2022 04:07 MB	Normal Saline Metanol		Sílica gel 3MM Whatman			1702 33		23 1603	0 pH (Low/Actual/High)		96,65 %	Pass 5,00 <= 5,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20220201-011		01/02/2022 04:09 MB	Normal Saline Metanol		Sílica gel 3MM Whatman			1065 3		19 348	0 pH (Low/Actual/High)		97,393 %	Pass 5,00 <= 5,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20220201-013		01/02/2022 04:22 MB	Normal Saline Metanol		Sílica gel 3MM Whatman			1232 1		18 1266	0 pH (Low/Actual/High)		98,481 %	Pass 5,00 <= 5,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20220201-014		01/02/2022 04:23 MB	Normal Saline Metanol		Sílica gel 3MM Whatman			1085 2		11 1315	0 pH (Low/Actual/High)		98,844 %	Pass 5,00 <= 5,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20220201-019		01/02/2022 07:06 PM	Normal Saline Metanol		Sílica gel 3MM Whatman			1677 5		10 1189	0 pH (Low/Actual/High)		98,988 %	Pass 5,00 <= 5,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20220201-020		01/02/2022 07:08 PM	Normal Saline Metanol		Sílica gel 3MM Whatman			1263 7		8 1313	0 pH (Low/Actual/High)		98,84 %	Pass 5,00 <= 5,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20220201-027		01/02/2022 07:40 PM	Normal Saline Metanol		Sílica gel 3MM Whatman			1112 6		5 1347	0 pH (Low/Actual/High)		99,109 %	Pass 5,00 <= 5,00 <= 6,00
Límpido/Incolor														

Product:	Manuf:	Calibration D/T	Solution	Media	Origin	Middle	Top	Solution 2	Media 2	Origin 2	Top 2	% Tag	QC By	Pass Fail
MIBI-99mTc Lot # K-20220201-028		01/02/2022 08:02 PM	Normal Saline Metanol		Sílica gel 3MM Whatman			1263 24		36 1094	0	pH (Low/Actual/High)	95,082 %	Pass 5,00 <= 5,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20220201-029		01/02/2022 08:05 PM	Normal Saline Metanol		Sílica gel 3MM Whatman			1237 19		3 1473	0	pH (Low/Actual/High)	98,485 %	Pass 5,00 <= 5,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20220201-033		01/02/2022 08:45 PM	Normal Saline Metanol		Sílica gel 3MM Whatman			1171 32		8 1200	0	pH (Low/Actual/High)	96,724 %	Pass 5,00 <= 5,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20220201-040		01/02/2022 10:33 PM	Normal Saline Metanol		Sílica gel 3MM Whatman			1527 40		12 1903	0	pH (Low/Actual/High)	97,162 %	Pass 5,00 <= 5,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20220201-045		01/02/2022 11:58 YD	Normal Saline Metanol		Sílica gel 3MM Whatman			1169 77		9 1466	0	pH (Low/Actual/High)	94,246 %	Pass 5,00 <= 6,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20220201-046		01/02/2022 12:01 YD	Normal Saline Metanol		Sílica gel 3MM Whatman			1006 36		16 1874	0	pH (Low/Actual/High)	96,55 %	Pass 5,00 <= 6,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20220201-048		01/02/2022 12:17 VS	Normal Saline Metanol		Sílica gel 3MM Whatman			624 9		13 523	0	pH (Low/Actual/High)	96,267 %	Pass 5,00 <= 6,00 <= 6,00
Límpido/Incolor														
MIBI-99mTc Lot # K-20220201-051		01/02/2022 12:49 VS	Normal Saline Metanol		Sílica gel 3MM Whatman			231 6		4 418	0	pH (Low/Actual/High)	96,883 %	Pass 5,00 <= 6,00 <= 6,00
Límpido/Incolor														
PIRO-99mTc Lot # K-20220201-022		01/02/2022 07:27 PM	Butanona Normal Saline		3MM Whatman 3MM Whatman			607 5		7 585	0	pH (Low/Actual/High)	98,012 %	Pass 5,00 <= 6,00 <= 7,00
Límpido/Incolor														
Tecnécio-99m Lot # E-20220201-001		01/02/2022 02:52 MB	Butanona		3MM Whatman			2		115	0	pH (Low/Actual/High)	98,291 %	Pass 5,00 <= 6,00 <= 6,00
Límpido e Incolor														
Tecnécio-99m Lot # E-20220201-002		01/02/2022 02:54 MB	Butanona		3MM Whatman			1		98,2	0	pH (Low/Actual/High)	98,992 %	Pass 5,00 <= 6,00 <= 6,00
Límpido e Incolor														
Tecnécio-99m Lot # E-20220201-003		01/02/2022 03:12 MB	Butanona		3MM Whatman			2		110	0	pH (Low/Actual/High)	98,214 %	Pass 5,00 <= 6,00 <= 6,00
Límpido e Incolor														
Tecnécio-99m Lot # E-20220201-004		01/02/2022 03:16 MB	Butanona		3MM Whatman			3		119	0	pH (Low/Actual/High)	97,541 %	Pass 5,00 <= 6,00 <= 6,00
Límpido e Incolor														

Product:	Manuf:	Calibration D/T	Solution	Media	Origin	Middle	Top	Solution 2	Media 2	Origin 2	Top 2	% Tag	QC By	Pass Fail
Tecnécio-99m Lot # E-20220201-005		01/02/2022 03:22	MB	Butanona		3MM Whatman		6		1121	0		99,468 %	Pass
												pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20220201-006		01/02/2022 03:23	MB	Butanona		3MM Whatman		21		2071	0		98,996 %	Pass
												pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20220201-007		01/02/2022 03:35	MB	Butanona		3MM Whatman		17		739	0		97,751 %	Pass
												pH (Low/Actual/High)	5,00 <= 5,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20220201-008		01/02/2022 04:01	MB	Butanona		3MM Whatman		1		332	0		99,70 %	Pass
												pH (Low/Actual/High)	5,00 <= 5,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20220201-009		01/02/2022 04:04	MB	Butanona		3MM Whatman		2		207	0		99,043 %	Pass
												pH (Low/Actual/High)	5,00 <= 5,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20220201-010		01/02/2022 04:04	MB	Butanona		3MM Whatman		4		763	0		99,478 %	Pass
												pH (Low/Actual/High)	5,00 <= 5,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20220201-011		01/02/2022 04:18	MB	Butanona		3MM Whatman		10		955	0		98,964 %	Pass
												pH (Low/Actual/High)	5,00 <= 5,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20220201-012		01/02/2022 04:18	MB	Butanona		3MM Whatman		6		1090	0		99,453 %	Pass
												pH (Low/Actual/High)	5,00 <= 5,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20220201-013		01/02/2022 04:31	MB	Butanona		3MM Whatman		1		428	0		99,767 %	Pass
												pH (Low/Actual/High)	5,00 <= 5,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20220201-014		01/02/2022 06:51	PM	Butanona		3MM Whatman		14		416	0		96,744 %	Pass
												pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20220201-017		01/02/2022 07:04	PM	Butanona		3MM Whatman		5		238	0		97,942 %	Pass
												pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20220201-016		01/02/2022 07:04	PM	Butanona		3MM Whatman		6		376	0		98,429 %	Pass
												pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
														Límpido e Incolor
Tecnécio-99m Lot # E-20220201-018		01/02/2022 07:20	PM	Butanona		3MM Whatman		9		950	0		99,062 %	Pass
												pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
														Límpido e Incolor

Product:	Manuf:	Calibration D/T	Solution	Media	Origin	Middle	Top	Solution 2	Media 2	Origin 2	Top 2	% Tag	QC By	Pass Fail
Tecnécio-99m Lot # E-20220201-019		01/02/2022 07:20 PM	Butanona		3MM	Whatman		28		818	0		96,69 %	Pass
							Límpido e Incolor					pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
Tecnécio-99m Lot # E-20220201-020		01/02/2022 07:37 PM	Butanona		3MM	Whatman		40		1273	0		96,954 %	Pass
							Límpido e Incolor					pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
Tecnécio-99m Lot # E-20220201-021		01/02/2022 07:59 PM	Butanona		3MM	Whatman		4		618	0		99,357 %	Pass
							Límpido e Incolor					pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
Tecnécio-99m Lot # E-20220201-022		01/02/2022 07:59 PM	Butanona		3MM	Whatman		10		834	0		98,815 %	Pass
							Límpido e Incolor					pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
Tecnécio-99m Lot # E-20220201-023		01/02/2022 08:16 PM	Butanona		3MM	Whatman		4		213	0		98,157 %	Pass
							Límpido e Incolor					pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
Tecnécio-99m Lot # E-20220201-024		01/02/2022 08:17 PM	Butanona		3MM	Whatman		13		1180	0		98,91 %	Pass
							Límpido e Incolor					pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
Tecnécio-99m Lot # E-20220201-025		01/02/2022 08:22 PM	Butanona		3MM	Whatman		7		192	0		96,482 %	Pass
							Límpido e Incolor					pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
Tecnécio-99m Lot # E-20220201-026		01/02/2022 08:36 PM	Butanona		3MM	Whatman		6		284	0		97,931 %	Pass
							Límpido e Incolor					pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
Tecnécio-99m Lot # E-20220201-028		01/02/2022 10:16 PM	Butanona		3MM	Whatman		6		500	0		98,814 %	Pass
							Límpido e Incolor					pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
Tecnécio-99m Lot # E-20220201-029		01/02/2022 10:16 PM	Butanona		3MM	Whatman		6		1692	0		99,647 %	Pass
							Límpido e Incolor					pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
Tecnécio-99m Lot # E-20220201-030		01/02/2022 10:25 PM	Butanona		3MM	Whatman		3		262	0		98,868 %	Pass
							Límpido e Incolor					pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
Tecnécio-99m Lot # E-20220201-031		01/02/2022 10:47 PM	Butanona		3MM	Whatman		7		208	0		96,744 %	Pass
							Límpido e Incolor					pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00
Tecnécio-99m Lot # E-20220201-032		01/02/2022 10:48 PM	Butanona		3MM	Whatman		10		387	0		97,481 %	Pass
							Límpido e Incolor					pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00

Product:	Manuf:	Calibration D/T	Solution	Media	Origin	Middle	Top	Solution 2	Media 2	Origin 2	Top 2	% Tag	QC By	Pass Fail
Tecnécio-99m Lot # E-20220201-033		01/02/2022 11:53 PM	Butanona		3MM Whatman			16		381	0	95,97 %	Pass	
							Límpido e Incolor				pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00	
Tecnécio-99m Lot # E-20220201-034		01/02/2022 11:54 PM	Butanona		3MM Whatman			20		402	0	95,261 %	Pass	
							Límpido e Incolor				pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00	
Tecnécio-99m Lot # E-20220201-035		01/02/2022 12:06 YD	Butanona		3MM Whatman			8		658	0	98,799 %	Pass	
							Límpido e Incolor				pH (Low/Actual/High)	5,00 <= 5,00	<= 6,00	
Tecnécio-99m Lot # E-20220201-036		01/02/2022 12:21 YD	Butanona		3MM Whatman			4		151	0	97,419 %	Pass	
							Límpido e Incolor				pH (Low/Actual/High)	5,00 <= 5,00	<= 6,00	
Tecnécio-99m Lot # E-20220201-037		01/02/2022 12:47 VS	Butanona		3MM Whatman			14		527	0	97,412 %	Pass	
							Límpido e Incolor				pH (Low/Actual/High)	5,00 <= 6,00	<= 6,00	
TRODAT-99mTc Lot # K-20220201-035		01/02/2022 10:26 PM	Normal Saline		Sílica gel			170,5		0,8	0	98,755 %	Pass	
			Butanona		3MM Whatman			1,7		216,7	pH (Low/Actual/High)	6,50 <= 7,00	<= 8,50	
							Límpido/Incolor							
TRODAT-99mTc Lot # K-20220201-036		01/02/2022 10:27 PM	Normal Saline		Sílica gel			95,2		3,2	0	96,065 %	Pass	
			Butanona		3MM Whatman			1,5		218	pH (Low/Actual/High)	6,50 <= 7,00	<= 8,50	
							Límpido/Incolor							
TRODAT-99mTc Lot # K-20220201-037		01/02/2022 10:29 PM	Normal Saline		Sílica gel			150		2,2	0	97,131 %	Pass	
			Butanona		3MM Whatman			3,4		235,5	pH (Low/Actual/High)	6,50 <= 7,00	<= 8,50	
							Límpido/Incolor							
TRODAT-99mTc Lot # K-20220201-038		01/02/2022 10:30 PM	Normal Saline		Sílica gel			154,9		1,4	0	97,663 %	Pass	
			Butanona		3MM Whatman			2,4		164,1	pH (Low/Actual/High)	6,50 <= 7,00	<= 8,50	
							Límpido/Incolor							
TRODAT-99mTc Lot # K-20220201-039		01/02/2022 10:31 PM	Normal Saline		Sílica gel			181,1		1	0	97,813 %	Pass	
			Butanona		3MM Whatman			4		240,2	pH (Low/Actual/High)	6,50 <= 7,00	<= 8,50	
							Límpido/Incolor							